



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121824

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	08/31/2022	09:10
2	N4444-01		1	12	100.00	>212.0	>212.0	08/31/2022	10:45
3	N4446-07		1	13	100.00	>212.0	>212.0	08/31/2022	11:15
4	N4447-04		1	15	100.00	>212.0	>212.0	08/31/2022	11:45
5	N4447-05		1	12	100.00	>212.0	>212.0	08/31/2022	12:15
6	N4452-01		1	15	100.00	>212.0	>212.0	08/31/2022	13:00
7	N4452-01DUP		1	15	100.00	>212.0	>212.0	08/31/2022	13:30
8	N4455-01		1	12	100.00	>212.0	>212.0	08/31/2022	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB121824

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FLASH POINT-WATER-C

WorkList ID : 162817

Department : Wet-Chemistry

Date : 08-31-2022 09:41:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/02/2022	Water	N4444-01	Flash Point	Cool 4 deg C	PSEG03	M11	WH-1	08/30/2022	1010B
09/12/2022	Water	N4446-07	Flash Point	Cool 4 deg C	BART02	M11	HDW-WATER-LIQUID	08/29/2022	1010B
09/02/2022	Water	N4447-04	Flash Point	Cool 4 deg C	PSEG03	M11	AUD-1656-WATER	08/30/2022	1010B
09/02/2022	Water	N4447-05	Flash Point	Cool 4 deg C	PSEG03	M11	AUD-1660	08/30/2022	1010B
09/12/2022	Water	N4452-01	Flash Point	Cool 4 deg C	TECT01	M11	MW-2-(22)	08/30/2022	1010B
09/06/2022	Water	N4455-01	Flash Point	Cool 4 deg C	PARS15	M11	KTS2-WC-L-MH-05	08/29/2022	1010B

Date/Time

8/31/22 9:45

Raw Sample Received by:

SI (wet-chem)

Raw Sample Relinquished by:

GN (SM)

Date/Time

8/31/22 14:30

Raw Sample Received by:

GN (SM)

Raw Sample Relinquished by:

SI (wet-chem)