

Analysis Method: SM2510 BANALYST: IwonaParameter: ConductanceSUPERVISOR REVIEW BY: jigneshRun Number: LB121835Meter ID: WC CONDUCTANCE METER-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP98799

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB121835BL		1	22.7	0.78	0.82	08/30/2022	11:20
2	LB121835BS	1413	1	22.9	1366.00	1423.08	08/30/2022	11:23
3	N4429-01		1	23.2	3.41	3.53	08/30/2022	11:26
4	N4429-01DUP		1	23.3	3.40	3.51	08/30/2022	11:30

Note : Conductivity result is corrected to 25°C

Result = <u> </u> (Reading) 1 + 0.0191 (Temperature - 25)

WORKLIST(Hardcopy Internal Chain)

LB121835

WorkList Name : CONDUCTANCE-083022 WorkList ID : 162761 Department : Wet-Chemistry Date : 08-30-2022 09:17:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/12/2022	Water	N4429-01	Conductance	Cool 4 deg C	STER06	M11	POOL-WATER	08/24/2022	SM2510 B

Date/Time 8/30/22 11:00
 Raw Sample Received by: 121507
 Raw Sample Relinquished by: 201607

Date/Time 8/30/22 11:00
 Raw Sample Received by: 121507
 Raw Sample Relinquished by: 201607