

LB121897


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

35754



A= 0.0000e+000

B= 2.8099e-004

C= 3.7871e-002

Rho= 0.9996772

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% 1)
0.0	0.000	0.017	0.017	-73	0.000	-73					
0.2	0.200	0.213	0.013	625	0.0 %	625					
2.5	2.500	2.600	0.100	9117	0.0 %	9117					
5.0	5.000	4.815	-0.185	17000	0.0 %	17000					
7.5	7.500	7.471	-0.029	26452	0.0 %	26452					
10.0	10.000	10.084	0.084	35754	0.0 %	35754					

100
-4
6
-4
0
1

LB121837 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Sep 2022 12:21:51

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-73	-	0.0000	7470A	PPB	01 Sep 2022 12:22:49	S	Std
0.2 - 1	502	625	-	0.2000	7470A	PPB	01 Sep 2022 12:25:06	S	Std
2.5 - 1	52.5	9117	-	2.5000	7470A	PPB	01 Sep 2022 12:27:23	S	Std
5.0 - 1	55	17000	-	5.0000	7470A	PPB	01 Sep 2022 12:29:40	S	Std
7.5 - 1	57.5	26452	-	7.5000	7470A	PPB	01 Sep 2022 12:31:57	S	Std
10.0 - 1	510	35754	-	10.0000	7470A	PPB	01 Sep 2022 12:34:13	S	Std
ICV15 - 1	ICV15	13894	3.9419	-7470A	PPB	01 Sep 2022 12:36:55	U	SMPL	
ICB15 - 1	ICB15	-98	0.0103	-7470A	PPB	01 Sep 2022 12:39:10	U	SMPL	
CCV39 - 1	CCV39	16939	4.7975	-7470A	PPB	01 Sep 2022 12:41:25	U	SMPL	
CCB39 - 1	CCB39	-79	0.0157	-7470A	PPB	01 Sep 2022 12:43:41	U	SMPL	
CRA - 1	CRA	643	0.2185	-7470A	PPB	01 Sep 2022 12:45:56	U	SMPL	
HighStd - 1	HighStd	35797	10.0964	-7470A	PPB	01 Sep 2022 12:48:11	U	SMPL	
ChkStd - 1	ChkStd	27120	7.6583	-7470A	PPB	01 Sep 2022 12:50:26	U	SMPL	
PB147421BL - 1	PBW	-125	0.0027	-7470A	PPB	01 Sep 2022 12:52:44	U	SMPL	
PB147421BS - 1	LCSW	15544	4.4055	-7470A	PPB	01 Sep 2022 12:55:01	U	SMPL	
N4375-02 - 1	WC-1	-6	0.0362	-7470A	PPB	01 Sep 2022 12:57:17	U	SMPL	
N4375-02DUP - 1	WC-1DUP	48	0.0514	-7470A	PPB	01 Sep 2022 12:59:34	U	SMPL	
N4375-02MS - 1	WC-1MS	15326	4.3443	-7470A	PPB	01 Sep 2022 13:01:51	U	SMPL	
N4375-02MSD - 1	WC-1MSD	15218	4.3139	-7470A	PPB	01 Sep 2022 13:04:08	U	SMPL	
N4375-05 - 1	WC-2	-83	0.0145	-7470A	PPB	01 Sep 2022 13:06:23	U	SMPL	
CCV40 - 1	CCV40	19310	5.4637	-7470A	PPB	01 Sep 2022 13:08:39	U	SMPL	
CCB40 - 1	CCB40	-127	0.0022	-7470A	PPB	01 Sep 2022 13:10:55	U	SMPL	
N4375-08 - 1	WC-3	-33	0.0286	-7470A	PPB	01 Sep 2022 13:13:11	U	SMPL	
N4375-11 - 1	WC-4	-20	0.0323	-7470A	PPB	01 Sep 2022 13:15:27	U	SMPL	
N4403-01 - 1	149TH-STREET-DRUM-5-PPE	-27	0.0303	-7470A	PPB	01 Sep 2022 13:17:42	U	SMPL	
N4403-02 - 1	149TH-STREET-DRUM-6-POLY	-31	0.0292	-7470A	PPB	01 Sep 2022 13:19:58	U	SMPL	
PB147311TB - 1	PB147311TB	-6	0.0362	-7470A	PPB	01 Sep 2022 13:22:14	U	SMPL	
N4375-02LX5 - 1		0	0.0379	-7470A	PPB	01 Sep 2022 13:24:30	U	SMPL	
N4375-02A - 1		14932	4.2336	-7470A	PPB	01 Sep 2022 13:26:47	U	SMPL	
CCV40-02-1	CCV41	18350	5.1940	-7470A	PPB	01 Sep 2022 15:06:51	U	SMPL	
CCB41 - 1	CCB41	-94	0.0115	-7470A	PPB	01 Sep 2022 15:09:08	U	SMPL	

3/11/22