



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/2/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 09/01/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 09/02/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121840

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4466-01	SAP2-T-5-(10-17)	1	0.91	8.85	9.76	8.51	85.9	
N4466-02	SAP2-T-5-(17-23)	2	0.9	8.82	9.72	8.15	82.2	
N4466-03	SAP2-T-5-(15-35)	3	0.93	9.04	9.97	8.61	85.0	
N4473-01	TOPSOIL-STORMWATER-PES T	4	0.93	8.84	9.77	7.92	79.1	
N4479-01	SB-8	13	0.91	8.82	9.73	9.71	99.8	
N4479-02	SB-9	14	0.91	9.01	9.92	9.78	98.4	
N4479-03	SB-10	15	0.92	8.60	9.52	9.5	99.8	
N4479-04	SB-11	16	0.93	8.75	9.68	9.65	99.7	
N4479-05	SB-12	17	0.93	8.97	9.9	9.85	99.4	
N4479-06	SB-13	18	0.91	8.64	9.55	9.51	99.5	
N4479-07	SB-14	19	0.93	8.66	9.59	9.55	99.5	
N4479-08	34128	20	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4479-09	34129	21	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4479-10	34130	22	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4485-01	AP-PIPE-1	5	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4485-02	AP-PIPE-2	6	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4485-03	AP-PIPE-3	7	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4485-04	CONCRETE-COMP-1	8	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N4485-05	CONCRETE-COMP-2	9	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N4485-06	AP-PIPE-2-SOIL	10	0.91	9.06	9.97	9.05	89.8	
N4486-01	PC-1	11	0.95	8.74	9.69	6.64	65.1	
N4487-01	H-1	12	0.91	8.78	9.69	9.15	93.8	
N4488-01	OR-3-090122	23	0.94	8.61	9.55	8.92	92.7	
N4488-02	OR-3-090122-EPH-2	24	0.93	8.78	9.71	9.17	93.8	
N4489-01	ORA-73-74-S	25	0.91	8.81	9.72	6.87	67.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$