

LB121851


7470A

INSTRUMENT ID: CV4

Linear

μ Abs.:

44249



A= 0.0000e+000

B= 2.2377e-004

C= 7.5908e-003

Rho= 0.9996363

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/8/11
0.0	0.000	0.018	0.018	47	0.000	47					1
0.2	0.200	0.213	0.013	920	0.0 %	920					6
2.5	2.500	2.475	-0.025	11025	0.0 %	11025					-1
5.0	5.000	4.890	-0.110	21819	0.0 %	21819					-2
7.5	7.500	7.695	0.195	34353	0.0 %	34353					3
10.0	10.000	9.909	-0.091	44249	0.0 %	44249					-1

LB121851 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Sep 2022 07:32:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7470A	PPB	02 Sep 2022 07:37:11	S	Std
0.2 - 1	50.2	920	-	0.2000	7470A	PPB	02 Sep 2022 07:39:28	S	Std
2.5 - 1	52.5	11025	-	2.5000	7470A	PPB	02 Sep 2022 07:41:45	S	Std
5.0 - 1	55	21819	-	5.0000	7470A	PPB	02 Sep 2022 07:44:02	S	Std
7.5 - 1	57.5	34353	-	7.5000	7470A	PPB	02 Sep 2022 07:46:18	S	Std
10.0 - 1	59	44249	-	10.0000	7470A	PPB	02 Sep 2022 07:48:35	S	Std
ICV16 - 1	ICV16	17334	3.8864	-	7470A	PPB	02 Sep 2022 07:51:20	U	SMPL
ICB16 - 1	ICB16	-151	-0.0262	-	7470A	PPB	02 Sep 2022 07:53:35	U	SMPL
CCV42 - 1	CCV42	21302	4.7743	-	7470A	PPB	02 Sep 2022 07:55:51	U	SMPL
CCB42 - 1	CCB42	-144	-0.0246	-	7470A	PPB	02 Sep 2022 07:58:07	U	SMPL
CRA - 1	CRA	891	0.2070	-	7470A	PPB	02 Sep 2022 08:00:23	U	SMPL
HighStd - 1	HighStd	43406	9.7205	-	7470A	PPB	02 Sep 2022 08:02:39	U	SMPL
ChkStd - 1	ChkStd	29921	6.7030	-	7470A	PPB	02 Sep 2022 08:04:54	U	SMPL
PB147386BL - 1	PBW	-179	-0.0325	-	7470A	PPB	02 Sep 2022 08:07:13	U	SMPL
PB147386BS - 1	LCSW	17035	3.8195	-	7470A	PPB	02 Sep 2022 08:09:31	U	SMPL
N4438-01 - 1	KTS2-WC-S-TR-01-02-03	-33	0.0002	-	7470A	PPB	02 Sep 2022 08:11:47	U	SMPL
N4438-02 - 1	KTS2-WC-S-MH-01-0-20	266	0.0671	-	7470A	PPB	02 Sep 2022 08:14:03	U	SMPL
N4438-03 - 1	KTS2-WC-S-MH-02-0-20	41	0.0168	-	7470A	PPB	02 Sep 2022 08:16:20	U	SMPL
N4438-03DUP - 1	KTS2-WC-S-MH-02-0-20DUP	59	0.0208	-	7470A	PPB	02 Sep 2022 08:18:37	U	SMPL
N4438-03MS - 1	KTS2-WC-S-MH-02-0-20MS	17189	3.8639	-	7470A	PPB	02 Sep 2022 08:20:53	U	SMPL
CCV43 - 1	CCV43	21512	4.8213	-	7470A	PPB	02 Sep 2022 08:23:09	U	SMPL
CCB43 - 1	CCB43	-157	-0.0275	-	7470A	PPB	02 Sep 2022 08:25:25	U	SMPL
N4438-03MSD - 1	KTS2-WC-S-MH-02-0-20MSD	15193	3.4073	-	7470A	PPB	02 Sep 2022 08:27:42	U	SMPL
PB147387BL - 1	PBW	-125	-0.0204	-	7470A	PPB	02 Sep 2022 08:29:57	U	SMPL
PB147387BS - 1	LCSW	17950	4.0242	-	7470A	PPB	02 Sep 2022 08:32:14	U	SMPL
N4427-01 - 1	D-171	-139	-0.0235	-	7470A	PPB	02 Sep 2022 08:34:29	U	SMPL
N4427-01DUP - 1	D-171DUP	-73	-0.0087	-	7470A	PPB	02 Sep 2022 08:36:46	U	SMPL
N4427-01MS - 1	D-171MS	15807	3.5447	-	7470A	PPB	02 Sep 2022 08:39:02	U	SMPL
N4427-01MSD - 1	D-171MSD	14825	3.3250	-	7470A	PPB	02 Sep 2022 08:41:19	U	SMPL
PB147388BL - 1	PBW	-185	-0.0338	-	7470A	PPB	02 Sep 2022 08:43:37	U	SMPL
PB147388BS - 1	LCSW	15694	3.5194	-	7470A	PPB	02 Sep 2022 08:45:55	U	SMPL
N4444-01 - 1	WH-1	580	0.1374	-	7470A	PPB	02 Sep 2022 08:48:11	U	SMPL
CCV44 - 1	CCV44	21799	4.8855	-	7470A	PPB	02 Sep 2022 08:50:27	U	SMPL
CCB44 - 1	CCB44	-151	-0.0262	-	7470A	PPB	02 Sep 2022 08:52:42	U	SMPL
N4446-01 - 1	WELL-01	-161	-0.0284	-	7470A	PPB	02 Sep 2022 08:54:59	U	SMPL
N4446-02 - 1	WELL-02	-138	-0.0233	-	7470A	PPB	02 Sep 2022 08:57:14	U	SMPL
N4446-03 - 1	WELL-03	-141	-0.0240	-	7470A	PPB	02 Sep 2022 08:59:30	U	SMPL
N4446-04 - 1	MW-04	-143	-0.0244	-	7470A	PPB	02 Sep 2022 09:01:45	U	SMPL
N4446-05 - 1	MW-05	-143	-0.0244	-	7470A	PPB	02 Sep 2022 09:04:01	U	SMPL
N4446-05DUP - 1	MW-05DUP	-135	-0.0226	-	7470A	PPB	02 Sep 2022 09:06:17	U	SMPL
N4446-06 - 1	MW-06	-136	-0.0228	-	7470A	PPB	02 Sep 2022 09:08:34	U	SMPL
N4446-08 - 1	DUP-Y	-153	-0.0266	-	7470A	PPB	02 Sep 2022 09:10:50	U	SMPL
N4446-09 - 1	N4446-05MS	14802	3.3198	-	7470A	PPB	02 Sep 2022 09:13:08	U	SMPL
N4446-10 - 1	N4446-05MSD	17561	3.9372	-	7470A	PPB	02 Sep 2022 09:15:24	U	SMPL
CCV45 - 1	CCV45	21859	4.8989	-	7470A	PPB	02 Sep 2022 09:17:40	U	SMPL
CCB45 - 1	CCB45	-157	-0.0275	-	7470A	PPB	02 Sep 2022 09:19:56	U	SMPL
N4446-11 - 1	EQUIP-BLANK	-163	-0.0289	-	7470A	PPB	02 Sep 2022 09:22:13	U	SMPL
N4447-04 - 1	AUD-1656-WATER	369	0.0902	-	7470A	PPB	02 Sep 2022 09:24:29	U	SMPL
N4447-05 - 1	AUD-1660	10033	2.2527	-	7470A	PPB	02 Sep 2022 09:26:44	U	SMPL
N4455-01 - 1	KTS2-WC-L-MH-02	45431	10.1736	-	7470A	PPB	02 Sep 2022 09:29:00	U	SMPL
N4456-01 - 1	KTS2-GW-MH-02-5-20	36391	8.1507	-	7470A	PPB	02 Sep 2022 09:31:19	U	SMPL
PB147367TB - 1	PB147367TB	-236	-0.0452	-	7470A	PPB	02 Sep 2022 09:33:38	U	SMPL
PB147368TB - 1	PB147368TB	-73	-0.0087	-	7470A	PPB	02 Sep 2022 09:35:56	U	SMPL
CCV46 - 1	CCV46	21392	4.7944	-	7470A	PPB	02 Sep 2022 09:38:13	U	SMPL
CCB46 - 1	CCB46	-297	-0.0589	-	7470A	PPB	02 Sep 2022 09:40:30	U	SMPL
N4438-03LX5 - 1		-79	-0.0101	-	7470A	PPB	02 Sep 2022 09:42:46	U	SMPL
N4438-03A - 1		17626	3.9517	-	7470A	PPB	02 Sep 2022 09:45:02	U	SMPL
N4427-01LX5 - 1		-158	-0.0278	-	7470A	PPB	02 Sep 2022 09:47:18	U	SMPL
N4427-01A - 1		17469	3.9166	-	7470A	PPB	02 Sep 2022 09:49:34	U	SMPL
N4446-05LX5 - 1		-143	-0.0244	-	7470A	PPB	02 Sep 2022 09:51:50	U	SMPL
N4446-05A - 1		16279	3.6503	-	7470A	PPB	02 Sep 2022 09:54:06	U	SMPL
CCV47 - 1	CCV47	21977	4.9253	-	7470A	PPB	02 Sep 2022 09:56:22	U	SMPL
CCB47 - 1	CCB47	-156	-0.0273	-	7470A	PPB	02 Sep 2022 09:58:38	U	SMPL