



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121854

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	09/02/2022	09:05
2	N4478-01		1	13	100.00	>212.0	>212.0	09/02/2022	09:30
3	N4478-01DUP		1	14	100.00	>212.0	>212.0	09/02/2022	10:05
4	N4489-06		1	15	100.00	>212.0	>212.0	09/02/2022	10:35
5	N4489-07		1	13	100.00	>212.0	>212.0	09/02/2022	11:05
6	N4489-08		1	14	100.00	>212.0	>212.0	09/02/2022	11:35
7	N4489-09		1	16	100.00	>212.0	>212.0	09/02/2022	12:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 12/1854

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FlashPoint-090222

WorkList ID : 162875

Department : Wet-Chemistry

Date : 09-02-2022 08:55:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/06/2022	Water	N4478-01	Flash Point	Cool 4 deg C	PSEG03	M11	FRAC-TANK-SV28486	09/01/2022	1010B
09/06/2022	Water	N4489-06	Flash Point	Cool 4 deg C	PSEG03	M11	DRIP-COMP	09/01/2022	1010B
09/06/2022	Water	N4489-07	Flash Point	Cool 4 deg C	PSEG03	M11	ORA-0067	09/01/2022	1010B
09/06/2022	Water	N4489-08	Flash Point	Cool 4 deg C	PSEG03	M11	ORA-0075	09/01/2022	1010B
09/06/2022	Water	N4489-09	Flash Point	Cool 4 deg C	PSEG03	M11	ORA-73-74-W	09/01/2022	1010B

Date/Time 9/2/22 9:00
Raw Sample Received by: ST (Wetchem)
Raw Sample Relinquished by: Ap (Ext.) (SM)

Date/Time 9/2/22 12:30
Raw Sample Received by: ST (Ext.) (SM)
Raw Sample Relinquished by: ST (Wetchem)