

LB 121867


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41650



A= 0.0000e+000

B= 2.4107e-004 *slope*C= 2.5938e-002 *y-intercept*

Rho= 0.9999077

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.011	0.011	-62	0.000	-62					
0.05	0.050					185	0				
0.20	0.200	0.230	0.030	845	0.0 %	845					
2.50	2.500	2.520	0.020	10347	0.0 %	10347					
5.00	5.000	4.912	-0.088	20270	0.0 %	20270					
7.50	7.500	7.460	-0.040	30839	0.0 %	30839					
10.0	10.000	10.066	0.066	41650	0.0 %	41650					

LB121867

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-62	-		0 SFAM01.1	PPB	9/2/2022 13:46	Std
	0.2 S01	845	-		0.2 SFAM01.1	PPB	9/2/2022 13:48	Std
	2.5 S02	10347	-		2.5 SFAM01.1	PPB	9/2/2022 13:50	Std
	5 S03	20270	-		5 SFAM01.1	PPB	9/2/2022 13:53	Std
	7.5 S04	30839	-		7.5 SFAM01.1	PPB	9/2/2022 13:55	Std
	10 S05	41650	-		10 SFAM01.1	PPB	9/2/2022 13:57	Std
ICV017	ICV017	16719	4.0564 -		SFAM01.1	PPB	9/2/2022 14:00	SMPL
ICB017	ICB017	-144	-0.0088 -		SFAM01.1	PPB	9/2/2022 14:02	SMPL
CCV048	CCV048	20281	4.915 -		SFAM01.1	PPB	9/2/2022 14:05	SMPL
CCB048	CCB048	-143	-0.0085 -		SFAM01.1	PPB	9/2/2022 14:07	SMPL
PB147459BL	PBW459	97	0.0493 -		SFAM01.1	PPB	9/2/2022 14:09	SMPL
N4321-01	MC0AX2	1222	0.3205 -		SFAM01.1	PPB	9/2/2022 14:11	SMPL
N4321-02	MC0AX3	726	0.201 -		SFAM01.1	PPB	9/2/2022 14:14	SMPL
N4321-03	MC0AX6	392	0.1204 -		SFAM01.1	PPB	9/2/2022 14:16	SMPL
N4321-04	MC0AX7	337	0.1072 -		SFAM01.1	PPB	9/2/2022 14:18	SMPL
N4321-05	MC0AX8	-70	0.0091 -		SFAM01.1	PPB	9/2/2022 14:20	SMPL
N4321-06	MC0AX9	-69	0.0093 -		SFAM01.1	PPB	9/2/2022 14:23	SMPL
N4321-09	MC0AX9D	-54	0.0129 -		SFAM01.1	PPB	9/2/2022 14:25	SMPL
N4321-10	MC0AX9S	4055	1.0035 -		SFAM01.1	PPB	9/2/2022 14:27	SMPL
N4321-11	MC0AX4	-79	0.0069 -		SFAM01.1	PPB	9/2/2022 14:30	SMPL
N4321-12	MC0AX5	-50	0.0139 -		SFAM01.1	PPB	9/2/2022 14:32	SMPL
CCV049	CCV049	20113	4.8745 -		SFAM01.1	PPB	9/2/2022 14:34	SMPL
CCB049	CCB049	-156	-0.0117 -		SFAM01.1	PPB	9/2/2022 14:36	SMPL