

LB 121880


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39135



A= 0.0000e+000

B= 2.5346e-004

C= 3.4179e-002

Rho= 0.9997976

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	401)
0.0	0.000	0.060	0.060	102	0.000	102					1
0.2	0.200	0.214	0.014	709	0.0 %	709					7
2.5	2.500	2.391	-0.109	9299	0.0 %	9299					-4
5.0	5.000	4.964	-0.036	19449	0.0 %	19449					-1
7.5	7.500	7.618	0.118	29921	0.0 %	29921					2
10.0	10.000	9.953	-0.047	39135	0.0 %	39135					0

LB121880 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Sep 2022 08:47:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	30	102	-	0.0007470A	PPB	06 Sep 2022 08:52:54	S	Std
0.2 - 1	50.2	709	-	0.20007470A	PPB	06 Sep 2022 08:55:11	S	Std
2.5 - 1	52.5	9299	-	2.50007470A	PPB	06 Sep 2022 08:57:28	S	Std
5.0 - 1	55	19449	-	5.00007470A	PPB	06 Sep 2022 08:59:45	S	Std
7.5 - 1	57.5	29921	-	7.50007470A	PPB	06 Sep 2022 09:07:07	S	Std
10.0 - 1	510	39135	-	10.00007470A	PPB	06 Sep 2022 09:11:56	S	Std
ICV18 - 1	ICV18	15607	3.9899	-7470A	PPB	06 Sep 2022 09:17:39	U	SMPL
ICB18 - 1	ICB18	15	0.0380	-7470A	PPB	06 Sep 2022 09:19:53	U	SMPL
CCV50 - 1	CCV50	19893	5.0763	-7470A	PPB	06 Sep 2022 09:22:09	U	SMPL
CCB50 - 1	CCB50	10	0.0367	-7470A	PPB	06 Sep 2022 09:24:24	U	SMPL
CRA - 1	CRA	759	0.2266	-7470A	PPB	06 Sep 2022 09:26:40	U	SMPL
HighStd - 1	HighStd	38364	9.7579	-7470A	PPB	06 Sep 2022 09:28:56	U	SMPL
ChkStd - 1	ChkStd	27318	6.9582	-7470A	PPB	06 Sep 2022 09:31:11	U	SMPL
PB147451BL - 1	PBW	-39	0.0243	-7470A	PPB	06 Sep 2022 09:33:28	U	SMPL
PB147451BS - 1	LCSW	15711	4.0163	-7470A	PPB	06 Sep 2022 09:35:45	U	SMPL
N4295-05 - 1	XRDS-RECYCLED-SAND-SP-A-COMP	-90	0.0114	-7470A	PPB	06 Sep 2022 09:38:01	U	SMPL
N4295-05DUP - 1	XRDS-RECYCLED-SAND-SP-A-COMP	-15	0.0304	-7470A	PPB	06 Sep 2022 09:40:18	U	SMPL
N4295-05MS - 1	XRDS-RECYCLED-SAND-SP-A-COMPMS	15648	4.0003	-7470A	PPB	06 Sep 2022 09:42:35	U	SMPL
N4295-05MSD - 1	XRDS-RECYCLED-SAND-SP-A-COMPMSD	16293	4.1638	-7470A	PPB	06 Sep 2022 09:44:52	U	SMPL
PB147452BL - 1	PBW	66	0.0514	-7470A	PPB	06 Sep 2022 09:47:07	U	SMPL
CCV51 - 1	CCV51	19802	5.0532	-7470A	PPB	06 Sep 2022 09:49:23	U	SMPL
CCB51 - 1	CCB51	-1	0.0339	-7470A	PPB	06 Sep 2022 09:51:38	U	SMPL
PB147452BS - 1	LCSW	15455	3.9514	-7470A	PPB	06 Sep 2022 09:53:54	U	SMPL
N4472-01 - 1	SS-1	-86	0.0124	-7470A	PPB	06 Sep 2022 09:56:09	U	SMPL
N4472-02 - 1	SS-2	-22	0.0286	-7470A	PPB	06 Sep 2022 09:58:25	U	SMPL
N4472-03 - 1	SS-3	-63	0.0182	-7470A	PPB	06 Sep 2022 10:00:41	U	SMPL
N4472-04 - 1	SS-4	-62	0.0185	-7470A	PPB	06 Sep 2022 10:02:57	U	SMPL
N4472-05 - 1	SS-5	-58	0.0195	-7470A	PPB	06 Sep 2022 10:05:13	U	SMPL
N4472-06 - 1	SS-6	-42	0.0235	-7470A	PPB	06 Sep 2022 10:07:30	U	SMPL
N4472-07 - 1	SS-7	-34	0.0256	-7470A	PPB	06 Sep 2022 10:09:47	U	SMPL
N4472-07DUP - 1	SS-7DUP	-35	0.0253	-7470A	PPB	06 Sep 2022 10:12:04	U	SMPL
N4472-07MS - 1	SS-7MS	17841	4.5562	-7470A	PPB	06 Sep 2022 10:14:20	U	SMPL
CCV52 - 1	CCV52	19961	5.0935	-7470A	PPB	06 Sep 2022 10:16:35	U	SMPL
CCB52 - 1	CCB52	6	0.0357	-7470A	PPB	06 Sep 2022 10:18:52	U	SMPL
N4472-07MSD - 1	SS-7MSD	16376	4.1848	-7470A	PPB	06 Sep 2022 10:21:08	U	SMPL
PB147453BL - 1	PBW	66	0.0509	-7470A	PPB	06 Sep 2022 10:23:24	U	SMPL
PB147453BS - 1	LCSW	15655	4.0021	-7470A	PPB	06 Sep 2022 10:25:40	U	SMPL
N4479-01 - 1	SB-8	71	0.0522	-7470A	PPB	06 Sep 2022 10:27:56	U	SMPL
N4479-02 - 1	SB-9	322	0.1158	-7470A	PPB	06 Sep 2022 10:30:11	U	SMPL
N4479-04 - 1	SB-11	145	0.0709	-7470A	PPB	06 Sep 2022 10:32:28	U	SMPL
N4479-05 - 1	SB-12	144	0.0707	-7470A	PPB	06 Sep 2022 10:34:44	U	SMPL
N4479-06 - 1	SB-13	102	0.0600	-7470A	PPB	06 Sep 2022 10:37:01	U	SMPL
N4479-07 - 1	SB-14	137	0.0689	-7470A	PPB	06 Sep 2022 10:39:18	U	SMPL
N4479-07DUP - 1	SB-14DUP	121	0.0648	-7470A	PPB	06 Sep 2022 10:41:33	U	SMPL
CCV53 - 1	CCV53	20205	5.1553	-7470A	PPB	06 Sep 2022 10:43:49	U	SMPL
CCB53 - 1	CCB53	13	0.0375	-7470A	PPB	06 Sep 2022 10:46:05	U	SMPL
N4479-07MS - 1	SB-14MS	16108	4.1169	-7470A	PPB	06 Sep 2022 10:48:21	U	SMPL
N4479-07MSD - 1	SB-14MSD	15757	4.0279	-7470A	PPB	06 Sep 2022 10:50:36	U	SMPL
PB147454BL - 1	PBW	37	0.0436	-7470A	PPB	06 Sep 2022 10:52:52	U	SMPL
PB147454BS - 1	LCSW	15455	3.9514	-7470A	PPB	06 Sep 2022 10:55:08	U	SMPL
N4478-01 - 1	FRAC-TANK-SV28486	13558	3.4706	-7470A	PPB	06 Sep 2022 10:57:24	U	SMPL
N4478-01DUP - 1	FRAC-TANK-SV28486DUP	12196	3.1254	-7470A	PPB	06 Sep 2022 10:59:40	U	SMPL
N4478-01MS - 1	FRAC-TANK-SV28486MS	28603	7.2839	-7470A	PPB	06 Sep 2022 11:01:57	U	SMPL
N4478-01MSD - 1	FRAC-TANK-SV28486MSD	30326	7.7206	-7470A	PPB	06 Sep 2022 11:04:15	U	SMPL
N4489-07 - 1	ORA-0067	514	0.1645	-7470A	PPB	06 Sep 2022 11:12:29	U	SMPL
N4489-08 - 1	ORA-0075	-146	-0.0028	-7470A	PPB	06 Sep 2022 11:14:44	U	SMPL
CCV54 - 1	CCV54	21564	5.4998	-7470A	PPB	06 Sep 2022 11:17:02	U	SMPL
CCB54 - 1	CCB54	40	0.0443	-7470A	PPB	06 Sep 2022 11:19:17	U	SMPL
N4295-05LX5 - 1		-119	0.0040	-7470A	PPB	06 Sep 2022 11:21:34	U	SMPL
N4295-05A - 1		16756	4.2811	-7470A	PPB	06 Sep 2022 11:23:49	U	SMPL
N4472-07LX5 - 1		-161	-0.0066	-7470A	PPB	06 Sep 2022 11:26:05	U	SMPL
N4472-07A - 1		17879	4.5658	-7470A	PPB	06 Sep 2022 11:28:20	U	SMPL
N4479-07LX5 - 1		-124	0.0028	-7470A	PPB	06 Sep 2022 11:36:12	U	SMPL
N4479-07A - 1		17237	4.4031	-7470A	PPB	06 Sep 2022 11:38:28	U	SMPL
N4478-01LX5 - 1		2617	0.6975	-7470A	PPB	06 Sep 2022 11:40:44	U	SMPL
N4478-01A - 1		30117	7.6676	-7470A	PPB	06 Sep 2022 11:43:01	U	SMPL
CCV55 - 1	CCV55	20692	5.2788	-7470A	PPB	06 Sep 2022 11:45:18	U	SMPL
CCB55 - 1	CCB55	23	0.0400	-7470A	PPB	06 Sep 2022 11:47:35	U	SMPL
PB147451TB - 1	PB147399TB	-49	0.0218	-7470A	PPB	06 Sep 2022 11:49:51	U	SMPL
PB147452TB - 1	PB147426TB	-16	0.0301	-7470A	PPB	06 Sep 2022 11:52:08	U	SMPL
PB147427TB - 1	PB147427TB	108	0.0616	-7470A	PPB	06 Sep 2022 11:54:24	U	SMPL
N4455-01DLX2 - 1		21578	5.5033	-7470A	PPB	06 Sep 2022 11:56:41	U	SMPL
CCV56 - 1	CCV56	20630	5.2631	-7470A	PPB	06 Sep 2022 11:58:57	U	SMPL
CCB56 - 1	CCB56	-122	0.0033	-7470A	PPB	06 Sep 2022 12:01:13	U	SMPL