

LB12188P


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

56617



A= 0.0000e+000

B= 1.7611e-004

C= 5.6910e-002

Rho= 0.9999503

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/11
0.0	0.000	0.042	0.042	-85	0.000	-85					
0.2	0.200	0.208	0.008	860	0.0 %	860					
2.5	2.500	2.474	-0.026	13727	0.0 %	13727					
5.0	5.000	4.932	-0.068	27682	0.0 %	27682					
7.5	7.500	7.515	0.015	42351	0.0 %	42351					
10.0	10.000	10.028	0.028	56617	0.0 %	56617					

LB121888 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Sep 2022 14:14:17

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	-85	-	0.0000	7471B	PPB	06 Sep 2022 14:14:31	S	Std
0.2 - 1	S0.2	860	-	0.2000	7471B	PPB	06 Sep 2022 14:16:48	S	Std
2.5 - 1	S2.5	13727	-	2.5000	7471B	PPB	06 Sep 2022 14:19:05	S	Std
5.0 - 1	S5	27682	-	5.0000	7471B	PPB	06 Sep 2022 14:21:21	S	Std
7.5 - 1	S7.5	42351	-	7.5000	7471B	PPB	06 Sep 2022 14:23:38	S	Std
10.0 - 1	S10	56617	-	10.0000	7471B	PPB	06 Sep 2022 14:25:55	S	Std
ICV19 - 1	ICV19	21241	3.7977	-	7471B	PPB	06 Sep 2022 14:29:01	U	SMPL
ICB19 - 1	ICB19	-188	0.0238	-	7471B	PPB	06 Sep 2022 14:31:15	U	SMPL
CCV57 - 1	CCV57	27313	4.8671	-	7471B	PPB	06 Sep 2022 14:36:19	U	SMPL
CCB57 - 1	CCB57	-194	0.0227	-	7471B	PPB	06 Sep 2022 14:38:34	U	SMPL
CRA - 1	CRA	854	0.2073	-	7471B	PPB	06 Sep 2022 14:40:51	U	SMPL
HighStd - 1	HighStd	56522	10.0111	-	7471B	PPB	06 Sep 2022 14:43:06	U	SMPL
ChkStd - 1	ChkStd	38867	6.9018	-	7471B	PPB	06 Sep 2022 14:45:22	U	SMPL
PB147492BL - 1	PBS	-299	0.0043	-	7471B	PPB	06 Sep 2022 14:47:40	U	SMPL
PB147492BS - 1	LCSS	19857	3.5540	-	7471B	PPB	06 Sep 2022 14:49:58	U	SMPL
N4466-01 - 1	SAP2-T-5-(10-17)	211	0.0941	-	7471B	PPB	06 Sep 2022 14:52:15	U	SMPL
N4466-02 - 1	SAP2-T-5-(17-23)	111	0.0765	-	7471B	PPB	06 Sep 2022 14:54:32	U	SMPL
N4466-03 - 1	SAP2-T-5-(15-35)	79	0.0708	-	7471B	PPB	06 Sep 2022 14:56:50	U	SMPL
N4485-06 - 1	AP-PIPE-2-SOIL	2910	0.5694	-	7471B	PPB	06 Sep 2022 14:59:07	U	SMPL
N4486-01 - 1	PC-1	2629	0.5199	-	7471B	PPB	06 Sep 2022 15:01:22	U	SMPL
CCV58 - 1	CCV58	27322	4.8686	-	7471B	PPB	06 Sep 2022 15:03:38	U	SMPL
CCB58 - 1	CCB58	-201	0.0215	-	7471B	PPB	06 Sep 2022 15:05:53	U	SMPL
N4487-01 - 1	H-1	298	0.1094	-	7471B	PPB	06 Sep 2022 15:08:10	U	SMPL
N4488-01 - 1	OR-3-090122	103662	18.3130	-	7471B	PPB	06 Sep 2022 15:10:25	U	SMPL
N4489-01 - 1	ORA-73-74-S	-283	0.0071	-	7471B	PPB	06 Sep 2022 15:12:40	U	SMPL
N4490-01 - 1	WC-14-A-SP	48649	8.6246	-	7471B	PPB	06 Sep 2022 15:14:58	U	SMPL
N4490-06 - 1	WC-14-B-SP	254104	44.8076	-	7471B	PPB	06 Sep 2022 15:17:14	U	SMPL
N4495-01 - 1	ETGI-345	51984	9.2119	-	7471B	PPB	06 Sep 2022 15:19:33	U	SMPL
N4498-01 - 1	CL-02-090222	2791	0.5484	-	7471B	PPB	06 Sep 2022 15:22:00	U	SMPL
N4505-01 - 1	31181	227	0.0969	-	7471B	PPB	06 Sep 2022 15:24:20	U	SMPL
N4506-01 - 1	OK-03-090222	3899	0.7436	-	7471B	PPB	06 Sep 2022 15:26:37	U	SMPL
N4506-01DUP - 1	OK-03-090222DUP	3469	0.6678	-	7471B	PPB	06 Sep 2022 15:28:53	U	SMPL
CCV59 - 1	CCV59	26923	4.7984	-	7471B	PPB	06 Sep 2022 15:31:08	U	SMPL
CCB59 - 1	CCB59	-211	0.0198	-	7471B	PPB	06 Sep 2022 15:33:24	U	SMPL
N4506-01MS - 1	OK-03-090222MS	23674	4.2262	-	7471B	PPB	06 Sep 2022 15:35:41	U	SMPL
N4506-01MSD - 1	OK-03-090222MSD	21941	3.9210	-	7471B	PPB	06 Sep 2022 15:43:01	U	SMPL
N4506-01LX5 - 1		374	0.1228	-	7471B	PPB	06 Sep 2022 15:45:18	U	SMPL
N4506-01A - 1		26159	4.6638	-	7471B	PPB	06 Sep 2022 15:47:36	U	SMPL
N4488-01DLX5 - 1		21756	3.8884	-	7471B	PPB	06 Sep 2022 15:49:51	U	SMPL
N4490-06DLX10 - 1		27026	4.8165	-	7471B	PPB	06 Sep 2022 15:52:10	U	SMPL
CCV60 - 1	CCV60	28483	5.0731	-	7471B	PPB	06 Sep 2022 15:54:27	U	SMPL
CCB60 - 1	CCB60	-272	0.0090	-	7471B	PPB	06 Sep 2022 15:56:46	U	SMPL