



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121889

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	09/06/2022	08:40
2	N4502-02		1	13	100.00	>212.0	>212.0	09/06/2022	09:05
3	N4505-03		1	14	100.00	>212.0	>212.0	09/06/2022	09:35
4	N4505-03DUP		1	12	100.00	>212.0	>212.0	09/06/2022	10:05
5	N4505-04		1	15	100.00	>212.0	>212.0	09/06/2022	10:35
6	N4505-05		1	14	100.00	>212.0	>212.0	09/06/2022	11:05
7	N4505-06		1	14	100.00	>212.0	>212.0	09/06/2022	11:35
8	N4505-07		1	17	100.00	>212.0	>212.0	09/06/2022	12:05
9	N4505-08		1	14	100.00	>212.0	>212.0	09/06/2022	12:35
10	N4505-09		1	15	100.00	>212.0	>212.0	09/06/2022	13:05
11	N4505-10		1	12	100.00	>212.0	>212.0	09/06/2022	13:35
12	N4505-11		1	12	100.00	>212.0	>212.0	09/06/2022	14:02
13	N4505-12		1	15	100.00	>212.0	>212.0	09/06/2022	14:38
14	N4505-13		1	13	100.00	>212.0	>212.0	09/06/2022	15:05
15	N4505-14		1	14	100.00	>212.0	>212.0	09/06/2022	15:35
16	N4505-15		1	12	100.00	>212.0	>212.0	09/06/2022	16:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 121889

WorkList Name : FLASH POINT-0906

WorkList ID : 162942

Department : Wet-Chemistry

Date : 09-06-2022 8:30:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/16/2022	Water	N4502-02	Flash Point	Cool 4 deg C	ICES01	M11	DEACON-WATER-COMP	09/02/2022	1010B
09/07/2022	Water	N4505-03	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-79	09/02/2022	1010B
09/07/2022	Water	N4505-04	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-107	09/02/2022	1010B
09/07/2022	Water	N4505-05	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-108	09/02/2022	1010B
09/07/2022	Water	N4505-06	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-109	09/02/2022	1010B
09/07/2022	Water	N4505-07	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-76	09/02/2022	1010B
09/07/2022	Water	N4505-08	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-77	09/02/2022	1010B
09/07/2022	Water	N4505-09	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-50	09/02/2022	1010B
09/07/2022	Water	N4505-10	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-51	09/02/2022	1010B
09/07/2022	Water	N4505-11	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-52	09/02/2022	1010B
09/07/2022	Water	N4505-12	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-49	09/02/2022	1010B
09/07/2022	Water	N4505-13	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-78	09/02/2022	1010B
09/07/2022	Water	N4505-14	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-0111	09/02/2022	1010B
09/07/2022	Water	N4505-15	Flash Point	Cool 4 deg C	PSEG03	M11	31792	09/02/2022	1010B

Date/Time 9/6/22 8:30
 Raw Sample Received by: SJ (wet chem)
 Raw Sample Relinquished by: CP (SM)

Date/Time 9/6/22 16:45
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: SJ (wet chem)