



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB121890

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/06/2022	16:00
2	CAL2	1	Water	NA	NA	20.7	7.00	09/06/2022	16:01
3	CAL3	1	Water	NA	NA	20.7	10.03	09/06/2022	16:02
4	ICV	1	Water	NA	NA	20.8	6.98	09/06/2022	16:05
5	CCV1	1	Water	NA	NA	20.8	2.01	09/06/2022	16:10
6	N4509-09	1	Solid	20.02	20	22.3	7.78	09/06/2022	16:13
7	N4509-09DUP	1	Solid	20.03	20	22.4	7.79	09/06/2022	16:14
8	N4509-10	1	Solid	20.03	20	23.0	8.79	09/06/2022	16:20
9	N4509-11	1	Solid	20.02	20	22.1	8.13	09/06/2022	16:23
10	N4509-12	1	Solid	20.02	20	22.1	8.41	09/06/2022	16:25
11	N4509-13	1	Solid	20.03	20	21.4	8.53	09/06/2022	16:30
12	N4509-14	1	Solid	20.02	20	22.5	8.89	09/06/2022	16:33
13	N4509-15	1	Solid	20.04	20	22.3	7.60	09/06/2022	16:35
14	N4509-16	1	Solid	20.02	20	22.1	8.62	09/06/2022	16:38
15	CCV2	1	Water	NA	NA	20.8	12.03	09/06/2022	16:40

WORKLIST(Hardcopy Internal Chain)

121890

WorkList Name : CORROSIONITY N4509

WorkList ID : 162953

Department : Wet-Chemistry

Date : 09-06-2022 15:49:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/02/2022	Solid	N4509-09	Corrosivity	Cool 4 deg C	TETR06	M11	S1-SO_112D-20220902	09/02/2022	9045D
09/02/2022	Solid	N4509-10	Corrosivity	Cool 4 deg C	TETR06	M11	S1-SO_113D-20220902	09/02/2022	9045D
09/02/2022	Solid	N4509-11	Corrosivity	Cool 4 deg C	TETR06	M11	S1-SO_114D-20220902	09/02/2022	9045D
09/02/2022	Solid	N4509-12	Corrosivity	Cool 4 deg C	TETR06	M11	S1-SO_115D-20220902	09/02/2022	9045D
09/02/2022	Solid	N4509-13	Corrosivity	Cool 4 deg C	TETR06	M11	S1-SO_116D-20220902	09/02/2022	9045D
09/02/2022	Solid	N4509-14	Corrosivity	Cool 4 deg C	TETR06	M11	S1-SO_PW001-20220902	09/02/2022	9045D
09/02/2022	Solid	N4509-15	Corrosivity	Cool 4 deg C	TETR06	M11	S1-SO_PW002-20220902	09/02/2022	9045D
09/02/2022	Solid	N4509-16	Corrosivity	Cool 4 deg C	TETR06	M11	S1-SO_UNK-20220902	09/02/2022	9045D

16150

Date/Time 09/06/22
 Raw Sample Received by: J.P.C.
 Raw Sample Relinquished by: J.P.C. (sm)

Date/Time 09/06/22
 Raw Sample Received by: J.P.C. (sm)
 Raw Sample Relinquished by: J.P.C.