



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB121891

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/06/2022	16:00
2	CAL2	1	Water	NA	NA	20.7	7.00	09/06/2022	16:01
3	CAL3	1	Water	NA	NA	20.7	10.03	09/06/2022	16:02
4	ICV	1	Water	NA	NA	20.8	6.98	09/06/2022	16:05
5	CCV1	1	Water	NA	NA	20.7	2.02	09/06/2022	16:43
6	N4511-02	1	Water	NA	NA	20.7	9.91	09/06/2022	16:46
7	N4511-02DUP	1	Water	NA	NA	20.8	9.92	09/06/2022	16:47
8	CCV2	1	Water	NA	NA	20.7	12.03	09/06/2022	17:00

WORKLIST(Hardcopy Internal Chain)

121891

WorkList Name : CORROSIONITY N4511 w

WorkList ID : 162954

Department : Wet-Chemistry

Date : 09-06-2022 15:51:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/02/2022	Water	N4511-02	Corrosivity	Cool 4 deg C	TETR06	M11	S1-AQ_DM-IDW-20220902	09/02/2022	9040C

Date/Time

09/06/22

15:50

Raw Sample Received by:

10/10/22

Raw Sample Relinquished by:

10/10/22

Date/Time

09/06/22

17:30

Raw Sample Received by:

10/10/22

Raw Sample Relinquished by:

10/10/22