

LB121899


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40915



A= 0.0000e+000

B= 2.4099e-004

C= -3.3494e-002

Rho= 0.9995430

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.058	-0.058	-101	0.000	-101				
0.2	0.200	0.210	0.010	1012	0.0 %	1012				
2.5	2.500	2.470	-0.030	10388	0.0 %	10388				
5.0	5.000	5.060	0.060	21135	0.0 %	21135				
7.5	7.500	7.691	0.191	32055	0.0 %	32055				
10.0	10.000	9.826	-0.174	40915	0.0 %	40915				

LB121899 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Sep 2022 10:59:24

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-101	-	0.0000	7470A	PPB	07 Sep 2022 11:42:34	S	Std
0.2 - 1	50.2	1012	-	0.2000	7470A	PPB	07 Sep 2022 11:47:30	S	Std
2.5 - 1	52.5	10388	-	2.5000	7470A	PPB	07 Sep 2022 11:49:47	S	Std
5.0 - 1	55	21135	-	5.0000	7470A	PPB	07 Sep 2022 11:52:04	S	Std
7.5 - 1	57.5	32055	-	7.5000	7470A	PPB	07 Sep 2022 11:54:20	S	Std
10.0 - 1	510	40915	-	10.0000	7470A	PPB	07 Sep 2022 12:01:47	S	Std
ICV20 - 1	ICV20	16321	3.8997	-	7470A	PPB	07 Sep 2022 12:04:36	U	SMPL
ICB20 - 1	ICB20	-211	-0.0843	-	7470A	PPB	07 Sep 2022 12:06:52	U	SMPL
CCV61 - 1	CCV61	20830	4.9863	-	7470A	PPB	07 Sep 2022 12:09:06	U	SMPL
CCB61 - 1	CCB61	24	-0.0277	-	7470A	PPB	07 Sep 2022 12:11:22	U	SMPL
CRA - 1	CRA	983	0.2034	-	7470A	PPB	07 Sep 2022 12:13:37	U	SMPL
HighStd - 1	HighStd	41543	9.9778	-	7470A	PPB	07 Sep 2022 12:15:53	U	SMPL
ChkStd - 1	ChkStd	29736	7.1325	-	7470A	PPB	07 Sep 2022 12:18:08	U	SMPL
PB147499BL - 1	PBW	-79	-0.0525	-	7470A	PPB	07 Sep 2022 12:22:57	U	SMPL
PB147499BS - 1	LCSW	15363	3.6688	-	7470A	PPB	07 Sep 2022 12:25:13	U	SMPL
N4490-02 - 1	WC-14-A-SP	420	0.0677	-	7470A	PPB	07 Sep 2022 12:27:29	U	SMPL
N4490-07 - 1	WC-14-B-SP	413	0.0660	-	7470A	PPB	07 Sep 2022 12:29:46	U	SMPL
N4495-02 - 1	ETGI-345	469	0.0795	-	7470A	PPB	07 Sep 2022 12:32:03	U	SMPL
N4500-02 - 1	SB-24-COMP	435	0.0713	-	7470A	PPB	07 Sep 2022 12:34:20	U	SMPL
N4500-04 - 1	SB-25-COMP	508	0.0889	-	7470A	PPB	07 Sep 2022 12:36:35	U	SMPL
CCV62 - 1	CCV62	20816	4.9829	-	7470A	PPB	07 Sep 2022 12:38:51	U	SMPL
CCB62 - 1	CCB62	-249	-0.0935	-	7470A	PPB	07 Sep 2022 12:41:07	U	SMPL
N4500-06 - 1	SB-14-COMP	465	0.0786	-	7470A	PPB	07 Sep 2022 12:43:22	U	SMPL
N4500-08 - 1	SB-09-COMP	491	0.0848	-	7470A	PPB	07 Sep 2022 12:45:38	U	SMPL
N4500-10 - 1	SB-01-COMP	482	0.0827	-	7470A	PPB	07 Sep 2022 12:47:53	U	SMPL
N4500-12 - 1	SB-02-COMP	480	0.0774	-	7470A	PPB	07 Sep 2022 12:50:08	U	SMPL
N4500-14 - 1	SB-03-COMP	472	0.0803	-	7470A	PPB	07 Sep 2022 12:52:25	U	SMPL
N4500-16 - 1	SB-04-COMP	477	0.0815	-	7470A	PPB	07 Sep 2022 12:54:41	U	SMPL
N4500-18 - 1	SB-05-COMP	473	0.0805	-	7470A	PPB	07 Sep 2022 12:56:57	U	SMPL
N4500-20 - 1	SB-06-COMP	459	0.0771	-	7470A	PPB	07 Sep 2022 12:59:14	U	SMPL
N4502-01 - 1	SLUDGE-DRUM-1-4-COMP	506	0.0884	-	7470A	PPB	07 Sep 2022 13:06:45	U	SMPL
N4502-01DUP - 1	SLUDGE-DRUM-1-4-COMPDUP	486	0.0836	-	7470A	PPB	07 Sep 2022 13:09:00	U	SMPL
CCV63 - 1	CCV63	20597	4.9301	-	7470A	PPB	07 Sep 2022 13:11:18	U	SMPL
CCB63 - 1	CCB63	314	0.0422	-	7470A	PPB	07 Sep 2022 13:16:06	U	SMPL
N4502-01MS - 1	SLUDGE-DRUM-1-4-COMPMS	16204	3.8715	-	7470A	PPB	07 Sep 2022 13:18:21	U	SMPL
N4502-01MSD - 1	SLUDGE-DRUM-1-4-COMPMSD	15800	3.7741	-	7470A	PPB	07 Sep 2022 13:20:36	U	SMPL
PB147500BL - 1	PBW	-144	-0.0682	-	7470A	PPB	07 Sep 2022 13:44:16	U	SMPL
PB147500BS - 1	LCSW	15613	3.7290	-	7470A	PPB	07 Sep 2022 13:46:32	U	SMPL
N4502-02 - 1	DEACON-WATER-COMP	466	0.0788	-	7470A	PPB	07 Sep 2022 13:48:49	U	SMPL
N4502-02DUP - 1	DEACON-WATER-COMPDUP	529	0.0940	-	7470A	PPB	07 Sep 2022 13:51:05	U	SMPL
N4502-02MS - 1	DEACON-WATER-COMPMS	14867	3.5493	-	7470A	PPB	07 Sep 2022 13:56:08	U	SMPL
N4502-02MSD - 1	DEACON-WATER-COMPMSD	14678	3.5037	-	7470A	PPB	07 Sep 2022 13:58:25	U	SMPL
PB147501BL - 1	PBW	-137	-0.0665	-	7470A	PPB	07 Sep 2022 14:03:21	U	SMPL
PB147501BS - 1	LCSW	15474	3.6955	-	7470A	PPB	07 Sep 2022 14:05:36	U	SMPL
CCV64 - 1	CCV64	20106	4.8118	-	7470A	PPB	07 Sep 2022 14:07:52	U	SMPL
CCB64 - 1	CCB64	146	0.0017	-	7470A	PPB	07 Sep 2022 14:10:08	U	SMPL
N4383-01 - 1	252672-1	359	0.0530	-	7470A	PPB	07 Sep 2022 14:12:24	U	SMPL
N4383-02 - 1	252672-2	370	0.0557	-	7470A	PPB	07 Sep 2022 14:14:40	U	SMPL
N4385-01 - 1	265751-1	372	0.0562	-	7470A	PPB	07 Sep 2022 14:16:56	U	SMPL
N4385-02 - 1	265751-2	423	0.0684	-	7470A	PPB	07 Sep 2022 14:19:12	U	SMPL
N4387-01 - 1	251563-1	407	0.0646	-	7470A	PPB	07 Sep 2022 14:21:28	U	SMPL
N4387-02 - 1	251563-2	419	0.0675	-	7470A	PPB	07 Sep 2022 14:23:44	U	SMPL
N4389-01 - 1	265393-3	370	0.0557	-	7470A	PPB	07 Sep 2022 14:26:01	U	SMPL
N4389-02 - 1	265393-4	269	0.0313	-	7470A	PPB	07 Sep 2022 14:28:18	U	SMPL
N4391-01 - 1	256317-3	450	0.0750	-	7470A	PPB	07 Sep 2022 14:30:35	U	SMPL
N4391-02 - 1	256317-4	405	0.0641	-	7470A	PPB	07 Sep 2022 14:32:50	U	SMPL
CCV65 - 1	CCV65	20012	4.7891	-	7470A	PPB	07 Sep 2022 14:35:06	U	SMPL
CCB65 - 1	CCB65	-263	-0.0969	-	7470A	PPB	07 Sep 2022 14:37:22	U	SMPL
N4505-09 - 1	CLI-50	1419	0.3086	-	7470A	PPB	07 Sep 2022 14:39:39	U	SMPL
N4505-10 - 1	CLI-51	-37	-0.0424	-	7470A	PPB	07 Sep 2022 14:41:55	U	SMPL
N4505-11 - 1	CLI-52	278	0.0335	-	7470A	PPB	07 Sep 2022 14:44:10	U	SMPL
N4505-12 - 1	CLI-49	3920	0.9112	-	7470A	PPB	07 Sep 2022 14:49:09	U	SMPL
N4505-13 - 1	CLI-78	12660	3.0174	-	7470A	PPB	07 Sep 2022 14:54:13	U	SMPL
N4505-14 - 1	CLI-0111	84	-0.0133	-	7470A	PPB	07 Sep 2022 14:56:29	U	SMPL
N4511-02 - 1	S1-AQ_DM-IDW-20220902	192	0.0128	-	7470A	PPB	07 Sep 2022 14:58:47	U	SMPL
N4511-02DUP - 1	S1-AQ_DM-IDW-20220902DUP	162	0.0055	-	7470A	PPB	07 Sep 2022 15:01:04	U	SMPL
N4511-02MS - 1	S1-AQ_DM-IDW-20220902MS	17487	4.1808	-	7470A	PPB	07 Sep 2022 15:03:21	U	SMPL
N4511-02MSD - 1	S1-AQ_DM-IDW-20220902MSD	16251	3.8828	-	7470A	PPB	07 Sep 2022 15:08:38	U	SMPL
CCV66 - 1	CCV66	22098	5.2918	-	7470A	PPB	07 Sep 2022 15:10:54	U	SMPL
CCB66 - 1	CCB66	165	0.0063	-	7470A	PPB	07 Sep 2022 15:17:45	U	SMPL
PB147456TB - 1	PB147456TB	496	0.0860	-	7470A	PPB	07 Sep 2022 15:20:01	U	SMPL
PB147485TB - 1	PB147485TB	-178	-0.0764	-	7470A	PPB	07 Sep 2022 15:26:44	U	SMPL
N4502-01LX5 - 1		-141	-0.0675	-	7470A	PPB	07 Sep 2022 15:28:59	U	SMPL
N4502-01A - 1		16852	4.0276	-	7470A	PPB	07 Sep 2022 15:31:16	U	SMPL
N4502-02LX5 - 1		-128	-0.0643	-	7470A	PPB	07 Sep 2022 15:33:32	U	SMPL
N4502-02A - 1		16876	4.0334	-	7470A	PPB	07 Sep 2022 15:35:48	U	SMPL
N4511-02LX5 - 1		-195	-0.0805	-	7470A	PPB	07 Sep 2022 15:38:04	U	SMPL
N4511-02A - 1		16401	3.9189	-	7470A	PPB	07 Sep 2022 15:40:21	U	SMPL
CCV67 - 1	CCV67	21537	5.1566	-	7470A	PPB	07 Sep 2022 15:42:38	U	SMPL
CCB67 - 1	CCB67	-22	-0.0388	-	7470A	PPB	07 Sep 2022 15:44:54	U	SMPL