

LB121918


SFAM01.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

35766



A= 0.0000e+000

B= 2.7477e-004 slope

C= 4.7204e-002 y-intercept

Rho= 0.9996478

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.00	0.000	-0.001	-0.001	-176	0.000	-176					
0.05	0.050					185	0				
0.20	0.200	0.217	0.017	618	0.0 %	618					
2.50	2.500	2.381	-0.119	8494	0.0 %	8494					
5.00	5.000	5.089	0.089	18350	0.0 %	18350					
7.50	7.500	7.639	0.139	27631	0.0 %	27631					
10.0	10.000	9.875	-0.125	35766	0.0 %	35766					

Handwritten calculations for %RSD:
 176 / 18350 = 0.00096
 185 / 8494 = 0.00218
 618 / 18350 = 0.00337
 8494 / 27631 = 0.03076
 18350 / 35766 = 0.05131
 27631 / 35766 = 0.07700
 35766 / 35766 = 1.00000
 The values 0.00096, 0.00218, 0.00337, 0.03076, 0.05131, 0.07700, and 1.00000 are the %RSD values for each concentration level.

LB121918

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-176	-		0 SFAM01.1	PPB	9/8/2022 10:05	Std
	0.2 S01	618	-		0.2 SFAM01.1	PPB	9/8/2022 10:07	Std
	2.5 S02	8494	-		2.5 SFAM01.1	PPB	9/8/2022 10:09	Std
	5 S03	18350	-		5 SFAM01.1	PPB	9/8/2022 10:11	Std
	7.5 S04	27631	-		7.5 SFAM01.1	PPB	9/8/2022 10:21	Std
	10 S05	35766	-		10 SFAM01.1	PPB	9/8/2022 10:29	Std
ICV021	ICV021	15657	4.3492 -		SFAM01.1	PPB	9/8/2022 10:32	SMPL
ICB021	ICB021	-162	0.0027 -		SFAM01.1	PPB	9/8/2022 10:34	SMPL
CCV068	CCV068	18431	5.1115 -		SFAM01.1	PPB	9/8/2022 10:36	SMPL
CCB068	CCB068	-268	-0.0264 -		SFAM01.1	PPB	9/8/2022 10:38	SMPL
N4480-01	MDBVF1	400	0.1571 -		SFAM01.1	PPB	9/8/2022 10:45	SMPL
N4480-02	MDBVF2	494	0.1829 -		SFAM01.1	PPB	9/8/2022 10:47	SMPL
N4480-03	MDBVF3	675	0.2327 -		SFAM01.1	PPB	9/8/2022 10:50	SMPL
N4480-04	MDBVF4	673	0.2321 -		SFAM01.1	PPB	9/8/2022 10:52	SMPL
N4480-05	MDBVF5	915	0.2986 -		SFAM01.1	PPB	9/8/2022 10:54	SMPL
N4480-06	MDBVF5D	892	0.2923 -		SFAM01.1	PPB	9/8/2022 10:56	SMPL
N4480-07	MDBVF5S	4837	1.3763 -		SFAM01.1	PPB	9/8/2022 10:59	SMPL
PB147516BL	PBW516	6	0.0489 -		SFAM01.1	PPB	9/8/2022 11:01	SMPL
PB147439TB	LEB439	89	0.0717 -		SFAM01.1	PPB	9/8/2022 11:03	SMPL
CCV069	CCV069	16702	4.6364 -		SFAM01.1	PPB	9/8/2022 11:11	SMPL
CCB069	CCB069	38	0.0576 -		SFAM01.1	PPB	9/8/2022 11:15	SMPL