



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121920

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/07/2022	16:30
2	CAL2	1	Water	NA	NA	20.7	7.00	09/07/2022	16:31
3	CAL3	1	Water	NA	NA	20.7	10.03	09/07/2022	16:32
4	ICV	1	Water	NA	NA	20.8	6.98	09/07/2022	16:33
5	CCV1	1	Water	NA	NA	20.7	2.01	09/07/2022	16:35
6	N4515-08	1	Water	NA	NA	20.6	6.33	09/07/2022	16:36
7	N4523-01	1	Water	NA	NA	22.1	9.20	09/07/2022	16:37
8	N4532-01	1	Water	NA	NA	20.5	8.51	09/07/2022	16:40
9	N4533-01	1	Water	NA	NA	22.9	6.95	09/07/2022	16:42
10	N4536-07	1	Water	NA	NA	21.6	7.43	09/07/2022	16:44
11	N4536-08	1	Water	NA	NA	22.8	7.51	09/07/2022	16:45
12	N4536-08DUP	1	Water	NA	NA	22.7	7.52	09/07/2022	16:46
13	CCV2	1	Water	NA	NA	20.8	12.03	09/07/2022	16:49

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N4533 W

WorkList ID : 162996

Department : Wet-Chemistry

Date : 09-07-2022 15:56:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/06/2022	Water	N4515-08	pH	Cool 4 deg C	MAJO01	N11	DRAIN-WATER-TANK	09/06/2022	9040C
09/07/2022	Water	N4523-01	pH	Cool 4 deg C	PSEG03	N11	NWB-1828	09/07/2022	9040C
09/07/2022	Water	N4532-01	pH	Cool 4 deg C	MAJO01	N11	DRAIN-WATER-TANK-1	09/07/2022	9040C
09/07/2022	Water	N4533-01	pH	Cool 4 deg C	PSEG03	M11	34127	09/07/2022	9040C
09/07/2022	Water	N4536-07	pH	Cool 4 deg C	PSEG03	N11	ORG-0071	09/07/2022	9040C
09/07/2022	Water	N4536-08	pH	Cool 4 deg C	PSEG03	N11	31025	09/07/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

09/07/22 16:25

2016007

CP SM

Date/Time 09/07/22 18:00

Raw Sample Received by:

Raw Sample Relinquished by:

CP SM

18(WC)