



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121923

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/08/2022	11:40
2	CAL2	1	Water	NA	NA	20.5	7.00	09/08/2022	11:41
3	CAL3	1	Water	NA	NA	20.5	10.03	09/08/2022	11:42
4	ICV	1	Water	NA	NA	20.6	6.98	09/08/2022	11:45
5	CCV1	1	Water	NA	NA	20.6	2.01	09/08/2022	11:46
6	N4540-01	1	Water	NA	NA	20.2	6.17	09/08/2022	11:50
7	N4540-01DUP	1	Water	NA	NA	20.2	6.18	09/08/2022	11:51
8	N4541-01	1	Water	NA	NA	22.8	5.05	09/08/2022	12:00
9	N4541-02	1	Water	NA	NA	22.5	4.99	09/08/2022	12:01
10	N4541-03	1	Water	NA	NA	22.8	5.28	09/08/2022	12:05
11	CCV2	1	Water	NA	NA	20.8	12.03	09/08/2022	12:10

121923

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N4540

WorkList ID : 163028

Department : Wet-Chemistry

Date : 09-08-2022 11:28:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/07/2022	Water	N4540-01	pH	Cool 4 deg C	STER12	N11	A-B-CELL-POOL-WATER-3RD-Q	09/07/2022	9040C
09/02/2022	Water	N4541-01	pH	Cool 4 deg C	STER07	N11	CELL-A	09/02/2022	9040C
09/02/2022	Water	N4541-02	pH	Cool 4 deg C	STER07	N11	CELL-B/C	09/02/2022	9040C
09/02/2022	Water	N4541-03	pH	Cool 4 deg C	STER07	N11	CELL-D	09/02/2022	9040C

Date/Time 09/08/22 11:30

Raw Sample Received by: JP LCC

Raw Sample Relinquished by: JP SM

Date/Time 09/08/22 12:45

Raw Sample Received by: CP SM

Raw Sample Relinquished by: JP LCC