



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121929

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	09/08/2022	11:30
2	N4505-02		1	15	100.00	>212.0	>212.0	09/08/2022	12:05
3	N4536-05		1	16	100.00	>212.0	>212.0	09/08/2022	12:40
4	N4536-06		1	16	100.00	>212.0	>212.0	09/08/2022	13:10
5	N4536-06DUP		1	17	100.00	>212.0	>212.0	09/08/2022	13:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB121929

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flash point-w-09082022

WorkList ID : 163027

Department : Wet-Chemistry

Date : 09-08-2022 11:23:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/07/2022	Solid	N4505-02	Flash Point	Cool 4 deg C	PSEG03	M11	CLI-78-OIL	09/02/2022	1010B
09/12/2022	Solid	N4536-05	Flash Point	Cool 4 deg C	PSEG03	N11	31025-OIL	09/07/2022	1010B
09/12/2022	Solid	N4536-06	Flash Point	Cool 4 deg C	PSEG03	N11	16750	09/07/2022	1010B

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

9/8/22

11:25

51 (wet-chem)

CP(5m)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

9/8/22

14:00

CP(5m)

51 (wet-chem)