



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121931

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	09/08/2022	11:30
2	N4511-02		1	14	100.00	>212.0	>212.0	09/08/2022	14:05
3	N4511-02DUP		1	14	100.00	>212.0	>212.0	09/08/2022	14:35
4	N4523-01		1	16	100.00	>212.0	>212.0	09/08/2022	15:10
5	N4533-01		1	16	100.00	>212.0	>212.0	09/08/2022	15:40
6	N4536-07		1	16	100.00	>212.0	>212.0	09/08/2022	16:05
7	N4536-08		1	15	100.00	>212.0	>212.0	09/08/2022	16:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 121931

WorkList Name : FLASHPOINT-09082022

WorkList ID : 163033

Department : Wet-Chemistry

Date : 09-08-2022 11:11:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/16/2022	Water	N4511-02	Flash Point	Cool 4 deg C	TETR06	M11	S1-AQ_DM-IDW-20220902	09/02/2022	1010B
09/14/2022	Water	N4523-01	Flash Point	Cool 4 deg C	PSEG03	N11	NWB-1828	09/07/2022	1010B
09/14/2022	Water	N4533-01	Flash Point	Cool 4 deg C	PSEG03	M11	34127	09/07/2022	1010B
09/12/2022	Water	N4536-07	Flash Point	Cool 4 deg C	PSEG03	N11	ORG-0071	09/07/2022	1010B
09/12/2022	Water	N4536-08	Flash Point	Cool 4 deg C	PSEG03	N11	31025	09/07/2022	1010B

Date/Time 9/8/22 11:15
 Raw Sample Received by: SJ (wetchem)
 Raw Sample Relinquished by: CP (Sri)

Date/Time 9/8/22 16:00
 Raw Sample Received by: CP (Sri)
 Raw Sample Relinquished by: SJ (wetchem)