


LB121933

7471B

INSTRUMENT ID: CV4



Linear

A= 0.0000e+000

B= 1.8236e-004

C= 2.3219e-002

Rho= 0.9995164

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.065	0.065	230	0.000	230					11
0.2	0.200	0.221	0.021	1083	0.0 %	1083					11
2.5	2.500	2.485	-0.015	13502	0.0 %	13502					11
5.0	5.000	4.973	-0.027	27143	0.0 %	27143					11
7.5	7.500	7.288	-0.212	39836	0.0 %	39836					13
10.0	10.000	10.168	0.168	55632	0.0 %	55632					2

LB121933 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Sep 2022 15:06:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	230	-	0.0000	7471B	PPB	08 Sep 2022 15:08:12	S	Std
0.2 - 1	50.2	1083	-	0.2000	7471B	PPB	08 Sep 2022 15:10:29	S	Std
2.5 - 1	52.5	13502	-	2.5000	7471B	PPB	08 Sep 2022 15:12:47	S	Std
5.0 - 1	55	27143	-	5.0000	7471B	PPB	08 Sep 2022 15:15:04	S	Std
7.5 - 1	57.5	39836	-	7.5000	7471B	PPB	08 Sep 2022 15:25:02	S	Std
10.0 - 1	510	55632	-	10.0000	7471B	PPB	08 Sep 2022 15:32:33	S	Std
ICV22 - 1	ICV22	22965	4.2111	-	7471B	PPB	08 Sep 2022 15:37:45	U	SMPL
ICB22 - 1	ICB22	62	0.0345	-	7471B	PPB	08 Sep 2022 15:40:00	U	SMPL
CCV70 - 1	CCV70	27027	4.9518	-	7471B	PPB	08 Sep 2022 15:42:17	U	SMPL
CCB70 - 1	CCB70	22	0.0272	-	7471B	PPB	08 Sep 2022 15:44:32	U	SMPL
CRA - 1	CRA	1070	0.2183	-	7471B	PPB	08 Sep 2022 15:46:49	U	SMPL
HighStd - 1	HighStd	55619	10.1658	-	7471B	PPB	08 Sep 2022 15:49:05	U	SMPL
ChkStd - 1	ChkStd	44814	8.1954	-	7471B	PPB	08 Sep 2022 15:51:20	U	SMPL
PB147543BL - 1	PBS	20	0.0269	-	7471B	PPB	08 Sep 2022 15:53:39	U	SMPL
PB147543BS - 1	LCSS	22652	4.1540	-	7471B	PPB	08 Sep 2022 15:55:57	U	SMPL
N4512-02 - 1	VNJ-214	3249	0.6157	-	7471B	PPB	08 Sep 2022 15:58:13	U	SMPL
N4512-02DUP - 1	VNJ-214DUP	2912	0.5542	-	7471B	PPB	08 Sep 2022 16:00:31	U	SMPL
N4512-02MS - 1	VNJ-214MS	23346	4.2805	-	7471B	PPB	08 Sep 2022 16:02:48	U	SMPL
N4512-02MSD - 1	VNJ-214MSD	24750	4.5366	-	7471B	PPB	08 Sep 2022 16:05:05	U	SMPL
N4513-01 - 1	72-11939	1777	0.3473	-	7471B	PPB	08 Sep 2022 16:07:22	U	SMPL
CCV71 - 1	CCV71	26954	4.9385	-	7471B	PPB	08 Sep 2022 16:09:40	U	SMPL
CCB71 - 1	CCB71	107	0.0427	-	7471B	PPB	08 Sep 2022 16:11:55	U	SMPL
N4514-01 - 1	EO-02-090622	3957	0.7448	-	7471B	PPB	08 Sep 2022 16:14:13	U	SMPL
N4518-01 - 1	SU-04-090622	3114	0.5911	-	7471B	PPB	08 Sep 2022 16:16:28	U	SMPL
N4520-01 - 1	600535069	8931	1.6519	-	7471B	PPB	08 Sep 2022 16:18:43	U	SMPL
N4520-03 - 1	600540642	7739	1.4345	-	7471B	PPB	08 Sep 2022 16:20:58	U	SMPL
N4528-01 - 1	PIT-10	19882	3.6489	-	7471B	PPB	08 Sep 2022 16:23:14	U	SMPL
N4528-03 - 1	PIT-12	38760	7.0914	-	7471B	PPB	08 Sep 2022 16:25:30	U	SMPL
N4538-07 - 1	MBD-SB-1-2-0-5	5539	1.0333	-	7471B	PPB	08 Sep 2022 16:27:47	U	SMPL
N4538-08 - 1	MBD-MP-1-0-46	5535	1.0326	-	7471B	PPB	08 Sep 2022 16:30:07	U	SMPL
N4512-02LX5 - 1		571	0.1273	-	7471B	PPB	08 Sep 2022 16:32:24	U	SMPL
N4512-02A - 1		24821	4.5495	-	7471B	PPB	08 Sep 2022 16:39:27	U	SMPL
CCV72 - 1	CCV72	27183	4.9802	-	7471B	PPB	08 Sep 2022 16:41:43	U	SMPL
CCB72 - 1	CCB72	2	0.0236	-	7471B	PPB	08 Sep 2022 16:45:58	U	SMPL