


LB121944

7470A

INSTRUMENT ID: CV1

 μ Abs.:

39044



Linear

A= 0.0000e+000

B= 2.5003e-004

C= 3.1361e-002

Rho= 0.9990354

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.014	0.014	-68	0.000	-68					1
0.2	0.200	0.218	0.018	745	0.0 %	745					9
2.5	2.500	2.296	-0.204	9056	0.0 %	9056					-8
5.0	5.000	5.171	0.171	20555	0.0 %	20555					3
7.5	7.500	7.708	0.208	30705	0.0 %	30705					3
10.0	10.000	9.793	-0.207	39044	0.0 %	39044					-2

LB121944 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Sep 2022 10:28:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-68	-	0.0000	7470A	PPB	09 Sep 2022 10:41:27	S	Std
0.2 - 1	50.2	745	-	0.2000	7470A	PPB	09 Sep 2022 10:43:44	S	Std
2.5 - 1	52.5	9056	-	2.5000	7470A	PPB	09 Sep 2022 10:46:01	S	Std
5.0 - 1	55	20555	-	5.0000	7470A	PPB	09 Sep 2022 10:52:37	S	Std
7.5 - 1	57.5	30705	-	7.5000	7470A	PPB	09 Sep 2022 10:54:54	S	Std
10.0 - 1	510	39044	-	10.0000	7470A	PPB	09 Sep 2022 11:03:23	S	Std
ICV23 - 1	ICV23	15722	3.9623	-	7470A	PPB	09 Sep 2022 11:06:15	U	SMPL
ICB23 - 1	ICB23	-175	-0.0124	-	7470A	PPB	09 Sep 2022 11:08:29	U	SMPL
CCV73 - 1	CCV73	19908	5.0089	-	7470A	PPB	09 Sep 2022 11:10:45	U	SMPL
CCB73 - 1	CCB73	-165	-0.0099	-	7470A	PPB	09 Sep 2022 11:13:00	U	SMPL
CRA - 1	CRA	797	0.2306	-	7470A	PPB	09 Sep 2022 11:15:17	U	SMPL
HighStd - 1	HighStd	38026	9.5388	-	7470A	PPB	09 Sep 2022 11:17:32	U	SMPL
ChkStd - 1	ChkStd	27730	6.9646	-	7470A	PPB	09 Sep 2022 11:19:48	U	SMPL
PB147539BL - 1	PBW	-213	-0.0219	-	7470A	PPB	09 Sep 2022 11:22:05	U	SMPL
PB147539BS - 1	LCSW	14739	3.7165	-	7470A	PPB	09 Sep 2022 11:24:22	U	SMPL
N4501-02 - 1	SB-07-COMP	-131	-0.0014	-	7470A	PPB	09 Sep 2022 11:26:39	U	SMPL
N4501-04 - 1	SB-08-COMP	-70	0.0139	-	7470A	PPB	09 Sep 2022 11:28:56	U	SMPL
N4501-06 - 1	SB-10-COMP	156	0.0704	-	7470A	PPB	09 Sep 2022 11:31:13	U	SMPL
N4501-08 - 1	SB-11-COMP	42	0.0419	-	7470A	PPB	09 Sep 2022 11:33:30	U	SMPL
N4501-10 - 1	SB-12-COMP	-91	0.0086	-	7470A	PPB	09 Sep 2022 11:35:46	U	SMPL
CCV74 - 1	CCV74	20114	5.0604	-	7470A	PPB	09 Sep 2022 11:38:01	U	SMPL
CCB74 - 1	CCB74	-148	-0.0056	-	7470A	PPB	09 Sep 2022 11:40:16	U	SMPL
N4501-12 - 1	SB-13-COMP	-75	0.0126	-	7470A	PPB	09 Sep 2022 11:42:32	U	SMPL
N4501-14 - 1	SB-15-COMP	-64	0.0154	-	7470A	PPB	09 Sep 2022 11:44:47	U	SMPL
N4501-16 - 1	SB-16-COMP	-62	0.0159	-	7470A	PPB	09 Sep 2022 11:47:02	U	SMPL
N4509-09 - 1	S1-SO_112D-20220902	-61	0.0161	-	7470A	PPB	09 Sep 2022 11:49:19	U	SMPL
N4509-10 - 1	S1-SO_113D-20220902	-58	0.0169	-	7470A	PPB	09 Sep 2022 11:51:34	U	SMPL
N4509-11 - 1	S1-SO_114D-20220902	-51	0.0186	-	7470A	PPB	09 Sep 2022 11:53:51	U	SMPL
N4509-12 - 1	S1-SO_115D-20220902	-17	0.0271	-	7470A	PPB	09 Sep 2022 11:56:07	U	SMPL
N4509-13 - 1	S1-SO_116D-20220902	-69	0.0141	-	7470A	PPB	09 Sep 2022 11:58:24	U	SMPL
N4509-14 - 1	S1-SO_PW001-20220902	-41	0.0211	-	7470A	PPB	09 Sep 2022 12:00:41	U	SMPL
N4509-15 - 1	S1-SO_PW002-20220902	-47	0.0196	-	7470A	PPB	09 Sep 2022 12:02:57	U	SMPL
CCV75 - 1	CCV75	21767	5.4737	-	7470A	PPB	09 Sep 2022 12:05:12	U	SMPL
CCB75 - 1	CCB75	-156	-0.0076	-	7470A	PPB	09 Sep 2022 12:07:28	U	SMPL
N4509-16 - 1	S1-SO_UNK-20220902	-25	0.0251	-	7470A	PPB	09 Sep 2022 12:09:44	U	SMPL
N4509-16DUP - 1	S1-SO_UNK-20220902DUP	1	0.0316	-	7470A	PPB	09 Sep 2022 12:12:00	U	SMPL
N4509-16MS - 1	S1-SO_UNK-20220902MS	15161	3.8220	-	7470A	PPB	09 Sep 2022 12:14:16	U	SMPL
N4509-16MSD - 1	S1-SO_UNK-20220902MSD	15086	3.8033	-	7470A	PPB	09 Sep 2022 12:16:31	U	SMPL
PB147545BL - 1	PBW	-70	0.0139	-	7470A	PPB	09 Sep 2022 12:18:48	U	SMPL
PB147545BS - 1	LCSW	13968	3.5237	-	7470A	PPB	09 Sep 2022 12:21:05	U	SMPL
N4528-01 - 1	PIT-10	-80	0.0114	-	7470A	PPB	09 Sep 2022 12:23:21	U	SMPL
N4528-03 - 1	PIT-12	-8	0.0294	-	7470A	PPB	09 Sep 2022 12:25:40	U	SMPL
N4530-02 - 1	SB-17-COMP	1	0.0316	-	7470A	PPB	09 Sep 2022 12:27:57	U	SMPL
N4530-04 - 1	SB-18-COMP	-16	0.0274	-	7470A	PPB	09 Sep 2022 12:30:13	U	SMPL
CCV76 - 1	CCV76	20646	5.1934	-	7470A	PPB	09 Sep 2022 12:32:28	U	SMPL
CCB76 - 1	CCB76	-157	-0.0079	-	7470A	PPB	09 Sep 2022 12:34:44	U	SMPL
N4530-06 - 1	SB-19-COMP	-13	0.0281	-	7470A	PPB	09 Sep 2022 12:37:00	U	SMPL
N4530-08 - 1	SB-20-COMP	10	0.0339	-	7470A	PPB	09 Sep 2022 12:39:16	U	SMPL
N4530-10 - 1	SB-21-COMP	3	0.0321	-	7470A	PPB	09 Sep 2022 12:41:32	U	SMPL
N4530-12 - 1	SB-22-COMP	-28	0.0244	-	7470A	PPB	09 Sep 2022 12:43:47	U	SMPL
N4530-14 - 1	SB-23-COMP	26	0.0379	-	7470A	PPB	09 Sep 2022 12:46:03	U	SMPL
N4533-06 - 1	34106	34	0.0399	-	7470A	PPB	09 Sep 2022 12:48:19	U	SMPL
N4536-01 - 1	090722-A	114	0.0599	-	7470A	PPB	09 Sep 2022 12:50:35	U	SMPL
N4536-02 - 1	090722-B	4	0.0324	-	7470A	PPB	09 Sep 2022 12:52:52	U	SMPL
N4536-03 - 1	090722-C	21	0.0366	-	7470A	PPB	09 Sep 2022 12:55:09	U	SMPL
N4536-04 - 1	090722-D	-62	0.0159	-	7470A	PPB	09 Sep 2022 12:57:25	U	SMPL
CCV77 - 1	CCV77	18889	4.7541	-	7470A	PPB	09 Sep 2022 12:59:41	U	SMPL
CCB77 - 1	CCB77	-141	-0.0039	-	7470A	PPB	09 Sep 2022 13:01:57	U	SMPL
N4536-04DUP - 1	090722-DDUP	-35	0.0226	-	7470A	PPB	09 Sep 2022 13:04:13	U	SMPL
N4536-04MS - 1	090722-DMS	14818	3.7362	-	7470A	PPB	09 Sep 2022 13:06:29	U	SMPL
N4536-04MSD - 1	090722-DMSD	14388	3.6287	-	7470A	PPB	09 Sep 2022 13:08:45	U	SMPL
PB147490TB - 1		-109	0.0041	-	7470A	PPB	09 Sep 2022 13:11:01	U	SMPL
PB147520TB - 1		-30	0.0239	-	7470A	PPB	09 Sep 2022 13:13:17	U	SMPL
N4509-16LX5 - 1		-151	-0.0064	-	7470A	PPB	09 Sep 2022 13:15:33	U	SMPL
N4509-16A - 1		15433	3.8900	-	7470A	PPB	09 Sep 2022 13:17:49	U	SMPL
N4536-04LX5 - 1		-213	-0.0219	-	7470A	PPB	09 Sep 2022 13:20:06	U	SMPL
N4536-04A - 1		16193	4.0800	-	7470A	PPB	09 Sep 2022 13:22:24	U	SMPL
CCV78 - 1	CCV78	19380	4.8769	-	7470A	PPB	09 Sep 2022 13:24:40	U	SMPL
CCB78 - 1	CCB78	-161	-0.0089	-	7470A	PPB	09 Sep 2022 13:26:56	U	SMPL