



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121946

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	09/09/2022	11:35
2	CAL2	1	Water	NA	NA	20.8	7.01	09/09/2022	11:36
3	CAL3	1	Water	NA	NA	20.8	10.03	09/09/2022	11:37
4	ICV	1	Water	NA	NA	20.7	6.99	09/09/2022	11:40
5	CCV1	1	Water	NA	NA	20.6	2.01	09/09/2022	11:41
6	N4547-01	1	Water	NA	NA	20.7	4.99	09/09/2022	11:43
7	N4551-01	1	Water	NA	NA	20.2	8.07	09/09/2022	11:45
8	N4551-02	1	Water	NA	NA	20.1	8.05	09/09/2022	11:47
9	N4553-01	1	Water	NA	NA	20.4	7.12	09/09/2022	11:50
10	N4553-02	1	Water	NA	NA	20.5	7.54	09/09/2022	11:52
11	N4555-01	1	Water	NA	NA	20.4	7.82	09/09/2022	12:00
12	N4555-02	1	Water	NA	NA	20.5	7.46	09/09/2022	12:05
13	N4555-02DUP	1	Water	NA	NA	20.6	7.47	09/09/2022	12:06
14	CCV2	1	Water	NA	NA	20.7	12.03	09/09/2022	12:10

WORKLIST(Hardcopy Internal Chain)

13121946

WorkList Name : ph n4553

WorkList ID : 163059

Department : Wet-Chemistry

Date : 09-09-2022 11:27:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/07/2022	Water	N4551-01	pH	Cool 4 deg C	ENV170	M11	251032-1	09/07/2022	9040C
09/07/2022	Water	N4551-02	pH	Cool 4 deg C	ENV170	M11	251032-2	09/07/2022	9040C
09/07/2022	Water	N4553-01	pH	Cool 4 deg C	ENV170	N11	256074-1	09/07/2022	9040C
09/07/2022	Water	N4553-02	pH	Cool 4 deg C	ENV170	N11	256074-2	09/07/2022	9040C
09/07/2022	Water	N4555-01	pH	Cool 4 deg C	ENV170	M11	257248-3	09/07/2022	9040C
09/07/2022	Water	N4555-02	pH	Cool 4 deg C	ENV170	M11	257248-4	09/07/2022	9040C
09/08/2022	Water	N4547-01	pH	Cool 4 deg C	MAJO01	M11	DRAIN-WATER-TANK-1	09/08/2022	9040C

Date/Time 09/09/22 11:30

Raw Sample Received by: 77109

Raw Sample Relinquished by: 77109

Date/Time 09/09/22 13:00
Raw Sample Received by: CP SN
Raw Sample Relinquished by: 77109