

**PERCENT SOLID**

Supervisor: Sohil
Analyst: jignesh
Date: 9/12/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 09/09/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 09/10/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121953

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4550-01	D-CRTE-SLAB-1	1	0.91	8.60	9.51	8.91	93.0	
N4558-11	BP-VPB-RW5-GW-58-60	2	0.93	9.05	9.98	1.27	3.8	SLUDGE SAMPLE
N4561-01	72-11976	3	0.91	8.85	9.76	8.42	84.9	
N4561-02	72-11976-EPH-2	4	0.91	8.85	9.76	8.42	84.9	
N4562-01	CHRT-2046	5	0.91	8.81	9.72	9.22	94.3	
N4562-02	CHRT-2046-EPH-2	6	0.93	9.04	9.97	8.34	82.0	
N4562-03	CHRT-26524	7	0.93	8.84	9.77	8.35	83.9	
N4562-04	CHRT-26524-EPH-2	8	0.91	9.04	9.95	8.74	86.6	
N4570-01	BC216279-1-1	9	1.00	1.00	2.00	2.00	100.0	PILC REEL
N4570-02	BC216279-1-2	10	1.00	1.00	2.00	2.00	100.0	PILC REEL
N4571-01	D2682	11	1.00	1.00	2.00	2.00	100.0	debris
N4572-01	PSC-124037	12	0.88	8.72	9.6	6.67	66.4	GEL MATRIX
N4572-03	PSC-124019	13	0.92	9.07	9.99	7.14	68.6	GEL MATRIX
N4572-05	PSC-164718	14	0.88	8.68	9.56	6.85	68.8	GEL MATRIX
N4572-07	PSC-124035	15	0.91	9.05	9.96	6.57	62.5	GEL MATRIX
N4573-01	172	16	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N4573-02	72-11998	17	0.88	8.79	9.67	8.66	88.5	
N4573-03	72-11998-EPH-2	18	0.9	8.87	9.77	8.46	85.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$