



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Sohil

Run Number: LB121960

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	09/09/2022	16:40
2	CAL2	1	Water	NA	NA	20.8	7.00	09/09/2022	16:41
3	CAL3	1	Water	NA	NA	20.8	10.03	09/09/2022	16:42
4	ICV	1	Water	NA	NA	20.7	6.98	09/09/2022	16:45
5	CCV1	1	Water	NA	NA	20.7	2.01	09/09/2022	16:46
6	N4571-01	1	Solid	20.01	20	23.1	4.67	09/09/2022	16:50
7	N4573-01	1	Solid	20.02	20	23.6	6.05	09/09/2022	16:51
8	N4573-01DUP	1	Solid	20.03	20	23.7	6.06	09/09/2022	16:52
9	CCV2	1	Water	NA	NA	20.7	12.03	09/09/2022	16:55

VB121960

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N4572 WorkList ID : 163071 Department : Wet-Chemistry Date : 09-09-2022 16:03:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/09/2022	Solid	N4571-01	Corrosivity	Cool 4 deg C	PSEG03	N11	D2682	09/09/2022	9045D
09/09/2022	Solid	N4573-01	Corrosivity	Cool 4 deg C	PSEG03	N11	172	09/09/2022	9045D

Date/Time 09/09/22 16:31
Raw Sample Received by: J.C. (sm)
Raw Sample Relinquished by: J.C. (sm)

Date/Time 09/09/22 18:30
Raw Sample Received by: J.C. (sm)
Raw Sample Relinquished by: J.C. (sm)