



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121962

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	09/09/2022	16:40
2	CAL2	1	Water	NA	NA	20.8	7.00	09/09/2022	16:41
3	CAL3	1	Water	NA	NA	20.8	10.03	09/09/2022	16:42
4	ICV	1	Water	NA	NA	20.7	6.98	09/09/2022	16:45
5	CCV1	1	Water	NA	NA	20.7	2.01	09/09/2022	17:20
6	N4563-01	1	Water	NA	NA	23.4	7.39	09/09/2022	17:22
7	N4564-01	1	Water	NA	NA	23.2	7.60	09/09/2022	17:24
8	N4567-01	1	Water	NA	NA	20.2	4.47	09/09/2022	17:26
9	N4571-03	1	Water	NA	NA	23.3	4.98	09/09/2022	17:28
10	N4571-03DUP	1	Water	NA	NA	23.4	4.99	09/09/2022	17:29
11	CCV2	1	Water	NA	NA	20.7	12.03	09/09/2022	17:30

VB 121962

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N4564 WorkList ID : 163073 Department : Wet-Chemistry Date : 09-09-2022 16:04:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/09/2022	Water	N4563-01	pH	Cool 4 deg C	PSEG03	N11	FM786	09/09/2022	9040C
09/09/2022	Water	N4564-01	pH	Cool 4 deg C	PSEG03	N11	FMI-120	09/09/2022	9040C
09/09/2022	Water	N4567-01	pH	Cool 4 deg C	MAJ001	N11	DRAIN-WATER-TANK-1	09/09/2022	9040C
09/09/2022	Water	N4571-03	pH	Cool 4 deg C	PSEG03	N11	T2625	09/09/2022	9040C

Date/Time 09/09/22 16:37
Raw Sample Received by: J.P. (S.M.)
Raw Sample Relinquished by: J.P. (S.M.)

Date/Time 09/09/22 18:30
Raw Sample Received by: J.C. (S.M.)
Raw Sample Relinquished by: J.P. (S.M.)