



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121969

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	09/12/2022	10:25
2	N4563-01		1	14	100.00	>212.0	>212.0	09/12/2022	10:45
3	N4564-01		1	16	100.00	>212.0	>212.0	09/12/2022	11:15
4	N4564-01DUP		1	16	100.00	>212.0	>212.0	09/12/2022	11:45
5	N4571-03		1	12	100.00	>212.0	>212.0	09/12/2022	12:15
6	N4573-04		1	15	100.00	>212.0	>212.0	09/12/2022	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash Point-0912

WorkList ID : 163100

Department : Wet-Chemistry

Date : 09-12-2022 10:16:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/16/2022	Water	N4563-01	Flash Point	Cool 4 deg C	PSEG03	N11	FM786	09/09/2022	1010B
09/16/2022	Water	N4564-01	Flash Point	Cool 4 deg C	PSEG03	N11	FMI-120	09/09/2022	1010B
09/16/2022	Water	N4571-03	Flash Point	Cool 4 deg C	PSEG03	N11	T2625	09/09/2022	1010B
09/16/2022	Water	N4573-04	Flash Point	Cool 4 deg C	PSEG03	N11	173	09/09/2022	1010B

Date/Time 9/12/2022 10:20

Raw Sample Received by: SS (wetchem)

Raw Sample Relinquished by: SP (SM)

Date/Time 9/12/2022 14:00

Raw Sample Received by: CP (SM)

Raw Sample Relinquished by: SJ (wetchem)