


LB121921

7470A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.4838e-004

C= 2.4110e-003

Rho= 0.9998735

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.056  | 0.056  | 217   | 0.000      | 217   |       |       |       |       |    |
| 0.2    | 0.200  | 0.205  | 0.005  | 817   | 0.0 %      | 817   |       |       |       |       |    |
| 2.5    | 2.500  | 2.493  | -0.007 | 10026 | 0.0 %      | 10026 |       |       |       |       |    |
| 5.0    | 5.000  | 4.877  | -0.123 | 19627 | 0.0 %      | 19627 |       |       |       |       |    |
| 7.5    | 7.500  | 7.529  | 0.029  | 30304 | 0.0 %      | 30304 |       |       |       |       |    |
| 10.0   | 10.000 | 10.039 | 0.039  | 40408 | 0.0 %      | 40408 |       |       |       |       |    |

0.0001

## LB121971 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Sep 2022 13:12:47

| Sample ID       | Extended ID           | $\mu$ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-----------------------|------------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50                    | 217        | -       | 0.0000   | 7470A  | PPB   | 12 Sep 2022 13:26:32 | S    | Std  |
| 0.2 - 1         | 512                   | 817        | -       | 0.2000   | 7470A  | PPB   | 12 Sep 2022 13:28:51 | S    | Std  |
| 2.5 - 1         | 52.5                  | 10026      | -       | 2.5000   | 7470A  | PPB   | 12 Sep 2022 13:31:08 | S    | Std  |
| 5.0 - 1         | 53                    | 19627      | -       | 5.0000   | 7470A  | PPB   | 12 Sep 2022 13:33:25 | S    | Std  |
| 7.5 - 1         | 575                   | 30304      | -       | 7.5000   | 7470A  | PPB   | 12 Sep 2022 13:35:41 | S    | Std  |
| 10.0 - 1        | 510                   | 40408      | -       | 10.0000  | 7470A  | PPB   | 12 Sep 2022 13:37:57 | S    | Std  |
| ICV26 - 1       | ICV26                 | 15812      | 3.9298  | -        | 7470A  | PPB   | 12 Sep 2022 13:40:36 | U    | SMPL |
| ICB26 - 1       | ICB26                 | -52        | -0.0105 | -        | 7470A  | PPB   | 12 Sep 2022 13:42:51 | U    | SMPL |
| CCV82 - 1       | CCV82                 | 19051      | 4.7343  | -        | 7470A  | PPB   | 12 Sep 2022 13:45:06 | U    | SMPL |
| CCB82 - 1       | CCB82                 | -59        | -0.0122 | -        | 7470A  | PPB   | 12 Sep 2022 13:47:22 | U    | SMPL |
| CRA - 1         | CRA                   | 741        | 0.1865  | -        | 7470A  | PPB   | 12 Sep 2022 13:49:38 | U    | SMPL |
| HighStd - 1     | HighStd               | 38931      | 9.6721  | -        | 7470A  | PPB   | 12 Sep 2022 13:51:53 | U    | SMPL |
| ChkStd - 1      | ChkStd                | 27250      | 6.7708  | -        | 7470A  | PPB   | 12 Sep 2022 13:54:09 | U    | SMPL |
| PB147571BL - 1  | PBW                   | -79        | -0.0172 | -        | 7470A  | PPB   | 12 Sep 2022 13:56:26 | U    | SMPL |
| PB147571BS - 1  | LCSW                  | 14683      | 3.6494  | -        | 7470A  | PPB   | 12 Sep 2022 13:58:43 | U    | SMPL |
| N4528-08-1 01   | PIT-10                | -20        | -0.0026 | -        | 7470A  | PPB   | 12 Sep 2022 14:00:59 | U    | SMPL |
| N4528-09-1 03   | PIT-12                | 51         | 0.0151  | -        | 7470A  | PPB   | 12 Sep 2022 14:03:16 | U    | SMPL |
| N4539-03 - 1    | MBD-WC-S-SB-1-2-0-5   | 22         | 0.0079  | -        | 7470A  | PPB   | 12 Sep 2022 14:05:33 | U    | SMPL |
| N4539-04 - 1    | MBD-WC-S-MP-1-0-46    | 63         | 0.0181  | -        | 7470A  | PPB   | 12 Sep 2022 14:07:51 | U    | SMPL |
| N4539-04DUP - 1 | MBD-WC-S-MP-1-0-46DUP | 60         | 0.0173  | -        | 7470A  | PPB   | 12 Sep 2022 14:10:06 | U    | SMPL |
| CCV83 - 1       | CCV83                 | 18763      | 4.6628  | -        | 7470A  | PPB   | 12 Sep 2022 14:12:22 | U    | SMPL |
| CCB83 - 1       | CCB83                 | -83        | -0.0182 | -        | 7470A  | PPB   | 12 Sep 2022 14:14:37 | U    | SMPL |
| N4539-04MS - 1  | MBD-WC-S-MP-1-0-46MS  | 13832      | 3.4380  | -        | 7470A  | PPB   | 12 Sep 2022 14:16:52 | U    | SMPL |
| N4539-04MSD - 1 | MBD-WC-S-MP-1-0-46MSD | 13521      | 3.3608  | -        | 7470A  | PPB   | 12 Sep 2022 14:19:07 | U    | SMPL |
| PB147546TB - 1  | PB147546TB            | -30        | -0.0050 | -        | 7470A  | PPB   | 12 Sep 2022 14:21:23 | U    | SMPL |
| PB147573TB - 1  | PB147573TB            | 25         | 0.0086  | -        | 7470A  | PPB   | 12 Sep 2022 14:23:38 | U    | SMPL |
| N4539-04LX5 - 1 |                       | -103       | -0.0232 | -        | 7470A  | PPB   | 12 Sep 2022 14:25:55 | U    | SMPL |
| N4539-04A - 1   |                       | 14708      | 3.6556  | -        | 7470A  | PPB   | 12 Sep 2022 14:30:54 | U    | SMPL |
| CCV84 - 1       | CCV84                 | 20777      | 5.1630  | -        | 7470A  | PPB   | 12 Sep 2022 14:51:56 | U    | SMPL |
| CCB84 - 1       | CCB84                 | 37         | 0.0116  | -        | 7470A  | PPB   | 12 Sep 2022 14:54:16 | U    | SMPL |