



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB121974
Analysis Date: 09/12/2022
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: Jignesh
REVIEWED BY: Sohil
Extraction Date: 09/12/2022
Extraction IN Time: 08:40
Extraction OUT Time: 09:40
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB121974BL	LB121974BL	WATER	1.3	1000	100	2.9651	2.9651	3.02	2.9652	2.9652	0.0001	0.1
2	LB121974BS	LB121974BS	WATER	1.3	1000	100	2.6308	2.6308	3.01	2.6478	2.6478	0.0170	17
3	N4567-01	DRAIN-WATER-TANK-1	WATER	1.6	1000	100	2.9810	2.9810	3.03	2.9842	2.9842	0.0032	3.2
4	N4567-02	N4567-01MS	WATER	1.6	1000	100	3.0478	3.0478	3.04	3.0710	3.0710	0.0232	23.2
5	N4567-03	N4567-01MSD	WATER	1.6	1000	100	3.0267	3.0267	3.05	3.0497	3.0497	0.0230	23



QC Batch# LB121974

Test: TPH

Analysis Date: 09/12/2022

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2938
pH Paper 0-14	M4909
Sodium Sulfate	EP2254
1:1 HCL	WP98461
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP98457
LCSWD	NA	NA
MS/MSD	5.00 ML	WP98458

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 11:01

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 10:15

Bal Check Time: 09:15 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 11:30

Out Time1: 11:00

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 12:31

1.0000 gram Balance: 1.0003 (0.9950-1.0050) In Time2: 12:00

Bal Check Time: 13:20 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 13:10

Out Time2: 12:30

127974

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph n4567 WorkList ID : 163081 Department : Wet-Chemistry Date : 09-12-2022 08:15:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/16/2022	Water	N4567-01	TPH	Conc H2SO4	tr MAJO01	N11	DRAIN-WATER-TANK-1	09/09/2022	1664A
09/16/2022	Water	N4567-02	TPH	Conc H2SO4	tr MAJO01	N11	N4597-01MS	09/09/2022	1664A
09/16/2022	Water	N4567-03	TPH	Conc H2SO4	tr MAJO01	N11	N4597-01MSD	09/09/2022	1664A

Date/Time 09.12.22 08:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 09.12.22 11:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]