



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121976

Analyst By : Jignesh

Supervisor Review By : Sohil

Slope : 99.2

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/12/2022	17:10
2	CAL2	1	Water	NA	NA	20.7	7.00	09/12/2022	17:11
3	CAL3	1	Water	NA	NA	20.8	10.03	09/12/2022	17:12
4	ICV	1	Water	NA	NA	20.7	6.99	09/12/2022	17:15
5	CCV1	1	Water	NA	NA	20.7	2.01	09/12/2022	17:20
6	N4580-01	1	Water	NA	NA	24.6	7.87	09/12/2022	17:22
7	N4586-01	1	Water	NA	NA	21.7	6.33	09/12/2022	17:25
8	N4588-04	1	Water	NA	NA	24.9	6.86	09/12/2022	17:26
9	N4588-05	1	Water	NA	NA	24.9	6.68	09/12/2022	17:30
10	N4588-07	1	Water	NA	NA	25.2	4.66	09/12/2022	17:31
11	N4590-01	1	Water	NA	NA	24.7	7.19	09/12/2022	17:33
12	N4590-01DUP	1	Water	NA	NA	24.6	7.20	09/12/2022	17:34
13	CCV2	1	Water	NA	NA	20.7	12.03	09/12/2022	17:35

127976

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N4586

WorkList ID : 163106

Department : Wet-Chemistry

Date : 09-12-2022 16:51:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/12/2022	Water	N4580-01	pH C	Cool 4 deg C	PSEG03	N11	CO-1	09/12/2022	9040C
09/12/2022	Water	N4586-01	pH B	Cool 4 deg C	MAJO01	N11	DRAIN-WATER-TANK-1	09/12/2022	9040C
09/12/2022	Water	N4588-04	pH C	Cool 4 deg C	PSEG03	N11	AUD-1666	09/12/2022	9040C
09/12/2022	Water	N4588-05	pH C	Cool 4 deg C	PSEG03	N11	AUD-1667	09/12/2022	9040C
09/12/2022	Water	N4588-07	pH C	Cool 4 deg C	PSEG03	N11	AUD-1665	09/12/2022	9040C
09/12/2022	Water	N4590-01	pH A	Cool 4 deg C	PSEG03	N11	FRAC-TANK-N48909	09/12/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

09.12.22 17:00
J C (5m)
J C (5m)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

09.12.22 18:00
J C (5m)
J C (5m)