


7470A

LB121978
INSTRUMENT ID: CV2

Linear

 μ Abs.:

3749E



A= 0.0000e+000

B= 2.6664e-004

C= 3.6667e-002

Rho= 0.9983901

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0
0.0	0.000	-0.011	-0.011	-178	0.000	-178					-
0.2	0.200	0.222	0.022	695	0.0 %	695					11
2.5	2.500	2.358	-0.142	8706	0.0 %	8706					-6
5.0	5.000	5.392	0.392	20084	0.0 %	20084					8
7.5	7.500	7.204	-0.296	26882	0.0 %	26882					-4
10.0	10.000	10.035	0.035	37496	0.0 %	37496					0

LB121978 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Sep 2022 09:51:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-178	-	0.0000	7470A	PPB	13 Sep 2022 09:55:49	S	Std
0.2 - 1	50.2	695	-	0.2000	7470A	PPB	13 Sep 2022 09:58:06	S	Std
2.5 - 1	52.5	8706	-	2.5000	7470A	PPB	13 Sep 2022 10:00:23	S	Std
5.0 - 1	55	20084	-	5.0000	7470A	PPB	13 Sep 2022 10:05:32	S	Std
7.5 - 1	57.5	26882	-	7.5000	7470A	PPB	13 Sep 2022 10:07:48	S	Std
10.0 - 1	510	37496	-	10.0000	7470A	PPB	13 Sep 2022 10:12:50	S	Std
ICV27 - 1	ICV27	14753	3.9704	-7470A	PPB	13 Sep 2022 10:16:46	U	SMPL	
ICB27 - 1	ICB27	-126	0.0031	-7470A	PPB	13 Sep 2022 10:19:00	U	SMPL	
CCV85 - 1	CCV85	17784	4.7786	-7470A	PPB	13 Sep 2022 10:21:15	U	SMPL	
CCB85 - 1	CCB85	-156	-0.0049	-7470A	PPB	13 Sep 2022 10:23:30	U	SMPL	
CRA - 1	CRA	558	0.1855	-7470A	PPB	13 Sep 2022 10:25:46	U	SMPL	
HighStd - 1	HighStd	37168	9.9471	-7470A	PPB	13 Sep 2022 10:28:02	U	SMPL	
ChkStd - 1	ChkStd	24705	6.6240	-7470A	PPB	13 Sep 2022 10:30:17	U	SMPL	
PB147599BL - 1	PBW	-209	-0.0191	-7470A	PPB	13 Sep 2022 10:32:34	U	SMPL	
PB147599BS - 1	LCSW	14200	3.8229	-7470A	PPB	13 Sep 2022 10:37:21	U	SMPL	
N4561-03 - 1	72-11976	-129	0.0023	-7470A	PPB	13 Sep 2022 10:39:38	U	SMPL	
N4571-01 - 1	D2682	-78	0.0159	-7470A	PPB	13 Sep 2022 10:41:54	U	SMPL	
N4571-01DUP - 1	D2682DUP	-25	0.0300	-7470A	PPB	13 Sep 2022 10:44:11	U	SMPL	
N4571-01MS - 1	D2682MS	11957	3.2249	-7470A	PPB	13 Sep 2022 10:46:28	U	SMPL	
N4571-01MSD - 1	D2682MSD	12391	3.3406	-7470A	PPB	13 Sep 2022 10:48:43	U	SMPL	
CCV86 - 1	CCV86	18928	5.0836	-7470A	PPB	13 Sep 2022 10:53:44	U	SMPL	
CCB86 - 1	CCB86	-168	-0.0081	-7470A	PPB	13 Sep 2022 10:55:59	U	SMPL	
N4572-02 - 1	PSC-124037	-68	0.0185	-7470A	PPB	13 Sep 2022 10:58:15	U	SMPL	
N4572-04 - 1	PSC-124019	-48	0.0239	-7470A	PPB	13 Sep 2022 11:00:31	U	SMPL	
N4572-06 - 1	PSC-164718	-71	0.0177	-7470A	PPB	13 Sep 2022 11:02:46	U	SMPL	
N4572-08 - 1	PSC-124035	-37	0.0268	-7470A	PPB	13 Sep 2022 11:05:02	U	SMPL	
N4573-04 - 1	173	206	0.0916	-7470A	PPB	13 Sep 2022 11:07:17	U	SMPL	
N4577-01 - 1	2672	-104	0.0089	-7470A	PPB	13 Sep 2022 11:09:34	U	SMPL	
PB147600BL - 1	PBW	-111	0.0071	-7470A	PPB	13 Sep 2022 11:11:50	U	SMPL	
PB147600BS - 1	LCSW	13794	3.7147	-7470A	PPB	13 Sep 2022 11:16:49	U	SMPL	
N4571-02 - 1	D2675	-69	0.0183	-7470A	PPB	13 Sep 2022 11:19:06	U	SMPL	
N4571-02DUP - 1	D2675DUP	40	0.0473	-7470A	PPB	13 Sep 2022 11:21:21	U	SMPL	
CCV87 - 1	CCV87	19059	5.1185	-7470A	PPB	13 Sep 2022 11:23:37	U	SMPL	
CCB87 - 1	CCB87	-133	0.0012	-7470A	PPB	13 Sep 2022 11:25:53	U	SMPL	
N4571-02MS - 1	D2675MS	12305	3.3176	-7470A	PPB	13 Sep 2022 11:28:12	U	SMPL	
N4571-02MSD - 1	D2675MSD	12308	3.3184	-7470A	PPB	13 Sep 2022 11:30:28	U	SMPL	
N4573-01 - 1	172	-142	-0.0012	-7470A	PPB	13 Sep 2022 11:32:44	U	SMPL	
PB147601BL - 1	PBW	-9	0.0343	-7470A	PPB	13 Sep 2022 11:35:00	U	SMPL	
PB147601BS - 1	LCSW	12655	3.4110	-7470A	PPB	13 Sep 2022 11:37:16	U	SMPL	
N4521-02 - 1	COMP	284	0.1124	-7470A	PPB	13 Sep 2022 11:39:33	U	SMPL	
N4521-02DUP - 1	COMPDUP	440	0.1540	-7470A	PPB	13 Sep 2022 11:41:49	U	SMPL	
N4521-02MS - 1	COMPMS	13857	3.7315	-7470A	PPB	13 Sep 2022 11:46:57	U	SMPL	
N4521-02MSD - 1	COMPMSD	12715	3.4270	-7470A	PPB	13 Sep 2022 11:49:13	U	SMPL	
N4551-01 - 1	251032-1	-209	-0.0191	-7470A	PPB	13 Sep 2022 11:51:30	U	SMPL	
CCV88 - 1	CCV88	17417	4.6807	-7470A	PPB	13 Sep 2022 11:53:46	U	SMPL	
CCB88 - 1	CCB88	-150	-0.0033	-7470A	PPB	13 Sep 2022 11:56:01	U	SMPL	
N4551-02 - 1	251032-2	-164	-0.0071	-7470A	PPB	13 Sep 2022 11:58:18	U	SMPL	
N4553-01 - 1	256074-1	-143	-0.0015	-7470A	PPB	13 Sep 2022 12:00:34	U	SMPL	
N4553-02 - 1	256074-2	-138	-0.0001	-7470A	PPB	13 Sep 2022 12:02:49	U	SMPL	
N4555-01 - 1	257248-3	-156	-0.0049	-7470A	PPB	13 Sep 2022 12:06:30	U	SMPL	
N4555-02 - 1	257248-4	-172	-0.0092	-7470A	PPB	13 Sep 2022 12:08:46	U	SMPL	
N4563-01 - 1	FM786	-174	-0.0097	-7470A	PPB	13 Sep 2022 12:11:02	U	SMPL	
N4564-01 - 1	FMI-120	-183	-0.0121	-7470A	PPB	13 Sep 2022 12:13:18	U	SMPL	
N4571-03 - 1	T2625	1588	0.4601	-7470A	PPB	13 Sep 2022 12:15:35	U	SMPL	
PB147575TB - 1	PB147575TB	-116	0.0057	-7470A	PPB	13 Sep 2022 12:17:52	U	SMPL	
PB147576TB - 1	PB147576TB	-96	0.0111	-7470A	PPB	13 Sep 2022 12:20:08	U	SMPL	
CCV89 - 1	CCV89	18106	4.8644	-7470A	PPB	13 Sep 2022 12:27:33	U	SMPL	
CCB89 - 1	CCB89	-169	-0.0084	-7470A	PPB	13 Sep 2022 12:29:49	U	SMPL	
N4571-01LX5 - 1		-174	-0.0097	-7470A	PPB	13 Sep 2022 12:32:04	U	SMPL	
N4571-01A - 1		16042	4.3141	-7470A	PPB	13 Sep 2022 12:34:20	U	SMPL	
N4571-02LX5 - 1		-199	-0.0164	-7470A	PPB	13 Sep 2022 12:36:35	U	SMPL	
N4571-02A - 1		17505	4.7042	-7470A	PPB	13 Sep 2022 12:38:51	U	SMPL	
N4521-02LX5 - 1		-193	-0.0148	-7470A	PPB	13 Sep 2022 12:41:07	U	SMPL	
N4521-02A - 1		16930	4.5509	-7470A	PPB	13 Sep 2022 12:47:58	U	SMPL	
CCV90 - 1	CCV90	17908	4.8116	-7470A	PPB	13 Sep 2022 12:52:54	U	SMPL	
CCB90 - 1	CCB90	-173	-0.0095	-7470A	PPB	13 Sep 2022 12:55:11	U	SMPL	