

LA121984


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

53181



A= 0.0000e+000

B= 1.8519e-004

C= 5.7132e-003

Rho= 0.9995375

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.004	-0.004	-54	0.000	-54					1
0.2	0.200	0.178	-0.022	933	0.0 %	933					-11
2.5	2.500	2.512	0.012	13532	0.0 %	13532					0
5.0	5.000	4.937	-0.063	26627	0.0 %	26627					-1
7.5	7.500	7.723	0.223	41672	0.0 %	41672					3
10.0	10.000	9.854	-0.146	53181	0.0 %	53181					-1

LB121984 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Sep 2022 13:54:40

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-54	-	0.0000	7471B	PPB	13 Sep 2022 13:57:24	S	Std
0.2 - 1	50.2	933	-	0.2000	7471B	PPB	13 Sep 2022 13:59:42	S	Std
2.5 - 1	52.5	13532	-	2.5000	7471B	PPB	13 Sep 2022 14:01:59	S	Std
5.0 - 1	55	26627	-	5.0000	7471B	PPB	13 Sep 2022 14:04:16	S	Std
7.5 - 1	57.5	41672	-	7.5000	7471B	PPB	13 Sep 2022 14:06:33	S	Std
10.0 - 1	510	53181	-	10.0000	7471B	PPB	13 Sep 2022 14:11:27	S	Std
ICV28 - 1	ICV28	22133	4.1045	-	7471B	PPB	13 Sep 2022 14:14:16	U	SMPL
ICB28 - 1	ICB28	-166	-0.0250	-	7471B	PPB	13 Sep 2022 14:16:30	U	SMPL
CCV91 - 1	CCV91	27711	5.1375	-	7471B	PPB	13 Sep 2022 14:18:47	U	SMPL
CCB91 - 1	CCB91	-171	-0.0260	-	7471B	PPB	13 Sep 2022 14:21:02	U	SMPL
CRA - 1	CRA	905	0.1733	-	7471B	PPB	13 Sep 2022 14:23:18	U	SMPL
HighStd - 1	HighStd	52898	9.8019	-	7471B	PPB	13 Sep 2022 14:25:33	U	SMPL
ChkStd - 1	ChkStd	38207	7.0813	-	7471B	PPB	13 Sep 2022 14:27:48	U	SMPL
PB147617BL - 1	PBS	-234	-0.0376	-	7471B	PPB	13 Sep 2022 14:30:07	U	SMPL
PB147617BS - 1	LCSS	18796	3.4866	-	7471B	PPB	13 Sep 2022 14:32:25	U	SMPL
N4545-01 - 1	HR-SOIL-0908	117533	21.7717	-	7471B	PPB	13 Sep 2022 14:34:41	U	SMPL
N4550-01 - 1	D-CRTE-SLAB-1	12211	2.2671	-	7471B	PPB	13 Sep 2022 14:36:59	U	SMPL
N4561-01 - 1	72-11976	3489	0.6518	-	7471B	PPB	13 Sep 2022 14:39:19	U	SMPL
N4562-01 - 1	CHRT-2046	6395	1.1900	-	7471B	PPB	13 Sep 2022 14:41:36	U	SMPL
N4562-03 - 1	CHRT-26524	4614	0.8602	-	7471B	PPB	13 Sep 2022 14:43:51	U	SMPL
CCV92 - 1	CCV92	27089	5.0223	-	7471B	PPB	13 Sep 2022 14:46:06	U	SMPL
CCB92 - 1	CCB92	-179	-0.0274	-	7471B	PPB	13 Sep 2022 14:48:22	U	SMPL
N4572-01 - 1	PSC-124037	2455	0.4604	-	7471B	PPB	13 Sep 2022 14:50:39	U	SMPL
N4572-03 - 1	PSC-124019	2142	0.4024	-	7471B	PPB	13 Sep 2022 14:52:54	U	SMPL
N4572-05 - 1	PSC-164718	1544	0.2916	-	7471B	PPB	13 Sep 2022 14:55:10	U	SMPL
N4572-07 - 1	PSC-124035	2055	0.3863	-	7471B	PPB	13 Sep 2022 14:57:26	U	SMPL
N4573-02 - 1	72-11998	5727	1.0663	-	7471B	PPB	13 Sep 2022 15:04:37	U	SMPL
N4573-02DUP - 1	72-11998DUP	5699	1.0611	-	7471B	PPB	13 Sep 2022 15:06:52	U	SMPL
N4573-02MS - 1	72-11998MS	23916	4.4347	-	7471B	PPB	13 Sep 2022 15:12:04	U	SMPL
N4573-02MSD - 1	72-11998MSD	25236	4.6792	-	7471B	PPB	13 Sep 2022 15:14:20	U	SMPL
N4584-01 - 1	33811	466	0.0920	-	7471B	PPB	13 Sep 2022 15:16:38	U	SMPL
N4588-01 - 1	AUD-1647	147	0.0329	-	7471B	PPB	13 Sep 2022 15:18:56	U	SMPL
CCV93 - 1	CCV93	26467	4.9071	-	7471B	PPB	13 Sep 2022 15:21:11	U	SMPL
CCB93 - 1	CCB93	-196	-0.0306	-	7471B	PPB	13 Sep 2022 15:23:27	U	SMPL
N4588-02 - 1	40152	1769	0.3333	-	7471B	PPB	13 Sep 2022 15:25:44	U	SMPL
N4591-01 - 1	OR-02-091222	1839	0.3463	-	7471B	PPB	13 Sep 2022 15:28:00	U	SMPL
N4592-01 - 1	EO-02-091222	1903	0.3581	-	7471B	PPB	13 Sep 2022 15:30:15	U	SMPL
PB147618BL - 1	PBS	-5	0.0048	-	7471B	PPB	13 Sep 2022 15:32:31	U	SMPL
PB147618BS - 1	LCSS	19592	3.6340	-	7471B	PPB	13 Sep 2022 15:37:23	U	SMPL
N4569-01 - 1	SASP2-T-6-(0-5)	12445	2.3104	-	7471B	PPB	13 Sep 2022 15:39:39	U	SMPL
N4569-01DUP - 1	SASP2-T-6-(0-5)DUP	13924	2.5843	-	7471B	PPB	13 Sep 2022 15:41:56	U	SMPL
N4569-01MS - 1	SASP2-T-6-(0-5)MS	32221	5.9727	-	7471B	PPB	13 Sep 2022 15:47:11	U	SMPL
N4569-01MSD - 1	SASP2-T-6-(0-5)MSD	34058	6.3129	-	7471B	PPB	13 Sep 2022 15:54:59	U	SMPL
N4569-02 - 1	SASP2-T-6-(5-10)	13633	2.5304	-	7471B	PPB	13 Sep 2022 15:57:15	U	SMPL
CCV94 - 1	CCV94	24589	4.5594	-	7471B	PPB	13 Sep 2022 15:59:33	U	SMPL
CCB94 - 1	CCB94	-66	-0.0065	-	7471B	PPB	13 Sep 2022 16:01:49	U	SMPL
N4578-03 - 1	CHERRY-HILL-COMP-1	1083	0.2063	-	7471B	PPB	13 Sep 2022 16:04:06	U	SMPL
N4578-07 - 1	CHERRY-HILL-COMP-2	580	0.1131	-	7471B	PPB	13 Sep 2022 16:06:21	U	SMPL
N4578-11 - 1	CHERRY-HILL-COMP-3	384	0.0768	-	7471B	PPB	13 Sep 2022 16:08:37	U	SMPL
N4578-15 - 1	CHERRY-HILL-COMP-4	1724	0.3250	-	7471B	PPB	13 Sep 2022 16:10:53	U	SMPL
N4578-19 - 1	CHERRY-HILL-COMP-5	2371	0.4448	-	7471B	PPB	13 Sep 2022 16:13:08	U	SMPL
N4578-23 - 1	CHERRY-HILL-COMP-6	2836	0.5309	-	7471B	PPB	13 Sep 2022 16:15:25	U	SMPL
N4578-27 - 1	CHERRY-HILL-COMP-7	737	0.1422	-	7471B	PPB	13 Sep 2022 16:17:42	U	SMPL
N4578-31 - 1	CHERRY-HILL-COMP-8	399	0.0796	-	7471B	PPB	13 Sep 2022 16:19:58	U	SMPL
N4578-36 - 1	CHERRY-HILL-COMP-9	532	0.1042	-	7471B	PPB	13 Sep 2022 16:22:16	U	SMPL
N4578-41 - 1	CHERRY-HILL-COMP-10	406	0.0809	-	7471B	PPB	13 Sep 2022 16:24:32	U	SMPL
CCV95 - 1	CCV95	24366	4.5181	-	7471B	PPB	13 Sep 2022 16:26:47	U	SMPL
CCB95 - 1	CCB95	-43	-0.0022	-	7471B	PPB	13 Sep 2022 16:29:03	U	SMPL
N4579-03 - 1	CHERRY-HILL-COMP-11	834	0.1602	-	7471B	PPB	13 Sep 2022 16:31:20	U	SMPL
N4579-07 - 1	CHERRY-HILL-COMP-12	32	0.0116	-	7471B	PPB	13 Sep 2022 16:33:35	U	SMPL
N4579-11 - 1	CHERRY-HILL-COMP-13	500	0.0983	-	7471B	PPB	13 Sep 2022 16:35:51	U	SMPL
N4579-15 - 1	CHERRY-HILL-COMP-14	470	0.0928	-	7471B	PPB	13 Sep 2022 16:38:06	U	SMPL
N4579-19 - 1	CHERRY-HILL-COMP-15	209	0.0444	-	7471B	PPB	13 Sep 2022 16:40:22	U	SMPL
N4579-23 - 1	CHERRY-HILL-COMP-16	586	0.1142	-	7471B	PPB	13 Sep 2022 16:42:38	U	SMPL
N4581-01 - 1	CRUSHED-MASONRY-PILE-6	321	0.0652	-	7471B	PPB	13 Sep 2022 16:44:54	U	SMPL
N4582-01 - 1	CRUSHED-MASONRY-PILE-7	420	0.0835	-	7471B	PPB	13 Sep 2022 16:47:15	U	SMPL
CCV96 - 1	CCV96	28456	5.2755	-	7471B	PPB	13 Sep 2022 16:55:03	U	SMPL
CCB96 - 1	CCB96	-154	-0.0228	-	7471B	PPB	13 Sep 2022 16:57:18	U	SMPL
N4573-02LX5 - 1		1277	0.2422	-	7471B	PPB	13 Sep 2022 16:59:36	U	SMPL
N4573-02A - 1		27588	5.1147	-	7471B	PPB	13 Sep 2022 17:01:52	U	SMPL
N4569-01LX5 - 1		2693	0.5044	-	7471B	PPB	13 Sep 2022 17:04:08	U	SMPL
N4569-01A - 1		38880	7.2059	-	7471B	PPB	13 Sep 2022 17:06:25	U	SMPL
N4545-01DLX5 - 1		25909	4.8038	-	7471B	PPB	13 Sep 2022 17:08:40	U	SMPL
CCV97 - 1	CCV97	28360	5.2577	-	7471B	PPB	13 Sep 2022 17:10:59	U	SMPL
CCB97 - 1	CCB97	-59	-0.0052	-	7471B	PPB	13 Sep 2022 17:15:34	U	SMPL