

Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB121985

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP98915
calibration std. phosphate 0.5 ppm	WP98914
calibration std. phosphate 0.3 ppm	WP98913
calibration std. phosphate 0.1 ppm	WP98912
calibration std. phosphate 0.05 ppm	WP98911
calibration std. 0 ppm	WP98910
phosphate CCV std.	WP98916
Combined reagent	WP98921
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP98644
Phosphate ICV-LCS Std	WP98917

Intercept: -0.0014

Slope: 0.6547

Regression: 0.999902

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.002		09/13/2022	13:30
2	CAL2	0.05	1	50	50	0.033	0.053	-6	09/13/2022	13:30
3	CAL3	0.10	1	50	50	0.066	0.103	-3	09/13/2022	13:31
4	CAL4	0.30	1	50	50	0.188	0.289	3.7	09/13/2022	13:31
5	CAL5	0.50	1	50	50	0.326	0.5	0	09/13/2022	13:32
6	CAL6	1.00	1	50	50	0.655	1.003	-0.3	09/13/2022	13:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB121985

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.497	09/13/2022	13:33
2	ICB		1	50	50	0.001	0.004	09/13/2022	13:33
3	CCV1	0.50	1	50	50	0.329	0.505	09/13/2022	13:34
4	CCB1		1	50	50	0.003	0.007	09/13/2022	13:34
5	PB147615BL		1	50	50	0.002	0.005	09/13/2022	13:35
6	PB147615BS	0.50	1	50	50	0.320	0.491	09/13/2022	13:35
7	N4460-01		1	50	50	0.638	0.977	09/13/2022	13:36
8	N4460-01DUP		1	50	50	0.639	0.978	09/13/2022	13:36
9	N4460-01MS	0.50	1	50	50	0.925	1.415	09/13/2022	13:37
10	N4460-01MSD	0.50	1	50	50	0.922	1.410	09/13/2022	13:37
11	N4462-01		1	50	50	1.125	1.720	09/13/2022	13:38
12	N4460-01MS	0.50	2	50	50	0.454	0.696	09/13/2022	13:38
13	N4460-01MSD	0.50	2	50	50	0.452	0.693	09/13/2022	13:39
14	N4462-01		10	50	50	0.267	0.410	09/13/2022	13:39
15	CCV2	0.50	1	50	50	0.323	0.495	09/13/2022	13:40
16	CCB2		1	50	50	0.002	0.005	09/13/2022	13:40