

Analytical Summary ReportAnalysis Method: 365.3ANALYST: IwonaParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: jigneshRun Number: LB121986BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP98915
calibration std. phosphate 0.5 ppm	WP98914
calibration std. phosphate 0.3 ppm	WP98913
calibration std. phosphate 0.1 ppm	WP98912
calibration std. phosphate 0.05 ppm	WP98911
calibration std. 0 ppm	WP98910
phosphate CCV std.	WP98916
Combined reagent	WP98921
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP98644
Phosphate ICV-LCS Std	WP98917
Phosphate LOD-MDL Std 0.025ppm	WP98919

Intercept: -0.0014Slope: 0.6547Regression: 0.999902

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000			09/13/2022	13:30
2	CAL2	0.05	1	50	50	0.033			09/13/2022	13:30
3	CAL3	0.10	1	50	50	0.066			09/13/2022	13:31
4	CAL4	0.30	1	50	50	0.188			09/13/2022	13:31
5	CAL5	0.50	1	50	50	0.326			09/13/2022	13:32
6	CAL6	1.00	1	50	50	0.655			09/13/2022	13:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB121986

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.497	09/13/2022	13:33
2	ICB		1	50	50	0.001	0.004	09/13/2022	13:33
3	CCV1	0.50	1	50	50	0.326	0.500	09/13/2022	13:45
4	CCB1		1	50	50	0.002	0.005	09/13/2022	13:45
5	PB147616BL		1	1.00	50	0.003	0.007	09/13/2022	13:46
6	PB147616BS	0.50	1	1.00	50	0.322	0.494	09/13/2022	13:46
7	N3980-03		1	1.00	50	0.016	0.027	09/13/2022	13:47
8	CCV2	0.50	1	50	50	0.326	0.500	09/13/2022	13:47
9	CCB2		1	50	50	0.003	0.007	09/13/2022	13:48