



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121987

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	09/13/2022	10:15
2	N4580-01		1	13	100.00	>212.0	>212.0	09/13/2022	11:00
3	N4588-06		1	16	100.00	>212.0	>212.0	09/13/2022	11:30
4	N4588-07		1	15	100.00	>212.0	>212.0	09/13/2022	12:00
5	N4590-01		1	13	100.00	>212.0	>212.0	09/13/2022	12:30
6	N4590-01DUP		1	12	100.00	>212.0	>212.0	09/13/2022	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB121987

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash point-0913

WorkList ID : 163129

Department : Wet-Chemistry

Date : 09-13-2022 10:11:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/15/2022	Water	N4580-01	Flash Point	Cool 4 deg C	PSEG03	N11	CO-1	09/12/2022	1010B
09/15/2022	Water	N4588-06	Flash Point	Cool 4 deg C	PSEG03	N11	AUD-66-67	09/12/2022	1010B
09/15/2022	Water	N4588-07	Flash Point	Cool 4 deg C	PSEG03	N11	AUD-1665	09/12/2022	1010B
09/15/2022	Water	N4590-01	Flash Point	Cool 4 deg C	PSEG03	N11	FRAC-TANK-N48909	09/12/2022	1010B

Date/Time 9/13/2022 10:15
 Raw Sample Received by: SJ (wet-chem)
 Raw Sample Relinquished by: CP (SM)

Date/Time 9/13/2022 15:00
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: SJ (wet-chem)