



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/14/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 09/13/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 09/14/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121990

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4604-01	A5748	1	0.91	8.64	9.55	8.76	90.9	
N4604-02	A5748MS	2	0.91	8.64	9.55	8.76	90.9	
N4604-03	A5748MSD	3	0.91	8.64	9.55	8.76	90.9	
N4604-04	A5763	4	0.93	9.03	9.96	8.83	87.5	
N4604-05	A5764	5	0.92	9.02	9.94	9.46	94.7	
N4604-06	A5765	6	0.91	8.86	9.77	8.79	88.9	
N4604-07	A5766	7	0.92	8.96	9.88	9.48	95.5	
N4604-08	A5767	8	0.91	9.04	9.95	9.5	95.0	
N4604-09	A5769	9	0.92	9.03	9.95	8.72	86.4	
N4604-10	A5770	10	0.93	9.03	9.96	9.06	90.0	
N4604-11	A5771	11	0.91	8.97	9.88	9.67	97.7	
N4604-12	A5772	12	0.95	8.68	9.63	9.07	93.5	
N4604-13	A5773	13	0.91	8.73	9.64	9.09	93.7	
N4604-14	A5774	14	0.93	8.88	9.81	8.8	88.6	
N4604-15	A5775	15	0.91	9.06	9.97	9.08	90.2	
N4604-16	A5776	16	0.89	8.63	9.52	8.45	87.6	
N4604-17	A5777	17	0.91	9.06	9.97	8.97	89.0	
N4604-18	A5778	18	0.93	9.03	9.96	8.92	88.5	
N4604-19	A5779	19	0.92	8.71	9.63	9.1	93.9	
N4604-20	A5780	20	0.91	8.84	9.75	8.69	88.0	
N4604-21	A5781	21	0.91	8.75	9.66	8.51	86.9	
N4604-22	A5782	22	0.93	8.90	9.83	9.34	94.5	
N4604-23	A5784	23	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$