



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/15/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:05
In Date: 09/14/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 09/15/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121994

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4598-01	MH0AA0	1	0.91	8.98	9.89	9.1	91.2	
N4598-02	MH0AA1	2	0.92	8.85	9.77	8.97	91.0	
N4598-03	MH0AA2	3	0.9	8.82	9.72	9.03	92.2	
N4598-04	MH0AA3	4	0.92	8.70	9.62	8.74	89.9	
N4598-05	MH0AA4	5	0.92	8.83	9.75	8.96	91.1	
N4598-06	MH0AA5	6	0.92	8.58	9.5	9.00	94.2	
N4598-07	MH0AA6	7	0.96	8.81	9.77	8.7	87.9	
N4598-08	MH0AA7	8	0.93	8.70	9.63	9.17	94.7	
N4598-09	MH0AA8	9	0.91	8.91	9.82	9.17	92.7	
N4598-10	MH0AA9	10	0.91	8.75	9.66	9.05	93.0	
N4598-11	MH0AB0	11	0.93	8.87	9.8	9.12	92.3	
N4598-12	MH0AB1	12	0.95	9.01	9.96	9.04	89.8	
N4598-13	MH0AB2	13	0.94	8.81	9.75	8.58	86.7	
N4598-14	MH0AB3	14	0.91	8.97	9.88	8.84	88.4	
N4598-15	MH0AB4	15	0.93	8.75	9.68	9.3	95.7	
N4598-16	MH0AB5	16	0.91	8.99	9.9	9.47	95.2	
N4598-17	MH0AB6	17	0.91	8.94	9.85	8.73	87.5	
N4598-18	MH0AB6D	18	0.91	8.94	9.85	8.73	87.5	
N4598-19	MH0AB6S	19	0.91	8.94	9.85	8.73	87.5	
N4598-20	MH0AB7	20	0.96	8.87	9.83	9.08	91.5	
N4598-21	MH0AB8	21	0.91	8.93	9.84	9.42	95.3	
N4598-22	MH0AB9	22	0.91	9.01	9.92	9.41	94.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$