



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/20/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 09:10
In Date: 09/15/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 2.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 09/16/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122007

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4560-01	MF5A31	1	0.9	8.90	9.8	7.32	72.1	
N4560-02	MF5A32	2	0.91	9.04	9.95	7.45	72.3	
N4560-03	MF5A33	3	0.92	9.03	9.95	7.19	69.4	
N4560-04	MF5A34	4	0.92	8.91	9.83	6.82	66.2	
N4560-05	MF5A43	5	0.91	8.91	9.82	7.03	68.7	
N4560-06	MF5A43D	6	0.91	8.91	9.82	7.03	68.7	
N4560-07	MF5A43S	7	0.91	8.91	9.82	7.03	68.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

122007

WorkList Name : %1-n4560

WorkList ID : 163191

Department : Wet-Chemistry

Date : 09-15-2022 09:11:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/18/2022	Solid	N4560-01	Percent Solids	Cool 4 deg C	USEP01	R21	MF5A31	09/08/2022	Chemtech -SO
09/18/2022	Solid	N4560-02	Percent Solids	Cool 4 deg C	USEP01	R21	MF5A32	09/08/2022	Chemtech -SO
09/18/2022	Solid	N4560-03	Percent Solids	Cool 4 deg C	USEP01	R21	MF5A33	09/08/2022	Chemtech -SO
09/18/2022	Solid	N4560-04	Percent Solids	Cool 4 deg C	USEP01	R21	MF5A34	09/08/2022	Chemtech -SO
09/18/2022	Solid	N4560-05	Percent Solids	Cool 4 deg C	USEP01	R21	MF5A43	09/08/2022	Chemtech -SO
09/18/2022	Solid	N4560-06	Percent Solids	Cool 4 deg C	USEP01	R21	MF5A43D	09/08/2022	Chemtech -SO
09/18/2022	Solid	N4560-07	Percent Solids	Cool 4 deg C	USEP01	R21	MF5A43S	09/08/2022	Chemtech -SO

Date/Time 09.15.22 12:30
 Raw Sample Received by: JDL/CAC
 Raw Sample Relinquished by: CP 5m

Date/Time 09.15.22 15:00
 Raw Sample Received by: CP 5m
 Raw Sample Relinquished by: JDL/CAC