


LB122012

SFAM01.1 INSTRUMENT ID: CY1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.00	0.000	-0.025	-0.025	65	0.000	65					1
0.05	0.050					185	0				-14
0.20	0.200	0.172	-0.028	992	0.0 %	992					0
2.50	2.500	2.503	0.003	11921	0.0 %	11921					1
5.00	5.000	5.055	0.055	23889	0.0 %	23889					1
7.50	7.500	7.603	0.103	35840	0.0 %	35840					1
10.0	10.000	9.892	-0.108	46575	0.0 %	46575					1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	65	-		0 SFAM01.1	PPB	9/15/2022 11:10	Std
	0.2 S01	992	-		0.2 SFAM01.1	PPB	9/15/2022 11:12	Std
	2.5 S02	11921	-		2.5 SFAM01.1	PPB	9/15/2022 11:14	Std
	5 S03	23889	-		5 SFAM01.1	PPB	9/15/2022 11:17	Std
	7.5 S04	35840	-		7.5 SFAM01.1	PPB	9/15/2022 11:19	Std
	10 S05	46575	-		10 SFAM01.1	PPB	9/15/2022 11:21	Std
ICV030	ICV030	19082	4.0298	-	SFAM01.1	PPB	9/15/2022 11:24	SMPL
ICB030	ICB030	10	-0.0371	-	SFAM01.1	PPB	9/15/2022 11:27	SMPL
CCV006	CCV006	23868	5.0503	-	SFAM01.1	PPB	9/15/2022 11:29	SMPL
CCB006	CCB006	3	-0.0386	-	SFAM01.1	PPB	9/15/2022 11:31	SMPL
PB147662BL	PBS662	63	-0.0258	-	SFAM01.1	PPB	9/15/2022 11:40	SMPL
N4542-01	MF5A03	1477	0.2757	-	SFAM01.1	PPB	9/15/2022 11:42	SMPL
N4542-02	MF5A03D	1613	0.3047	-	SFAM01.1	PPB	9/15/2022 11:45	SMPL
N4542-03	MF5A03S	11748	2.4659	-	SFAM01.1	PPB	9/15/2022 11:47	SMPL
N4542-04	MF5A04	1345	0.2475	-	SFAM01.1	PPB	9/15/2022 11:49	SMPL
N4542-05	MF5A05	1561	0.2936	-	SFAM01.1	PPB	9/15/2022 11:51	SMPL
N4542-06	MF5A09	1347	0.248	-	SFAM01.1	PPB	9/15/2022 11:54	SMPL
N4542-07	MF5A10	1693	0.3217	-	SFAM01.1	PPB	9/15/2022 11:56	SMPL
N4542-08	MF5A11	1625	0.3072	-	SFAM01.1	PPB	9/15/2022 11:58	SMPL
N4542-09	MF5A12	1550	0.2912	-	SFAM01.1	PPB	9/15/2022 12:00	SMPL
N4542-10	MF5A18	629	0.0948	-	SFAM01.1	PPB	9/15/2022 12:03	SMPL
N4542-11	MF5A21	643	0.0978	-	SFAM01.1	PPB	9/15/2022 12:05	SMPL
N4542-12	MF5A22	618	0.0925	-	SFAM01.1	PPB	9/15/2022 12:07	SMPL
N4542-13	MF5A42	882	0.1488	-	SFAM01.1	PPB	9/15/2022 12:09	SMPL
N4542-14	MF5A15	1397	0.2586	-	SFAM01.1	PPB	9/15/2022 12:12	SMPL
N4542-15	MF5A23	3515	0.7103	-	SFAM01.1	PPB	9/15/2022 12:14	SMPL
N4542-16	MF5A24	64438	13.7015	-	SFAM01.1	PPB	9/15/2022 12:16	SMPL
N4542-17	MF5A25	14264	3.0024	-	SFAM01.1	PPB	9/15/2022 12:19	SMPL
CCV007	CCV007	23318	4.933	-	SFAM01.1	PPB	9/15/2022 12:21	SMPL
CCB007	CCB007	-15	-0.0425	-	SFAM01.1	PPB	9/15/2022 12:23	SMPL
N4542-18	MF5A26	33815	7.1714	-	SFAM01.1	PPB	9/15/2022 12:25	SMPL
N4542-19	MF5A27	8605	1.7956	-	SFAM01.1	PPB	9/15/2022 12:28	SMPL
N4542-20	MF5A28	37621	7.983	-	SFAM01.1	PPB	9/15/2022 12:30	SMPL
N4542-21	MF5A29	865237	HIGH	-	SFAM01.1	PPB	9/15/2022 12:32	
N4542-22	MF5A30	11971	2.5134	-	SFAM01.1	PPB	9/15/2022 12:34	SMPL
PB147663BL	PBS663	-72	-0.0546	-	SFAM01.1	PPB	9/15/2022 12:36	SMPL
N4598-01	MH0AA0	3099	0.6216	-	SFAM01.1	PPB	9/15/2022 12:39	SMPL
N4598-02	MH0AA1	2541	0.5026	-	SFAM01.1	PPB	9/15/2022 12:41	SMPL
N4598-03	MH0AA2	3701	0.7499	-	SFAM01.1	PPB	9/15/2022 12:43	SMPL
N4598-04	MH0AA3	15231	3.2086	-	SFAM01.1	PPB	9/15/2022 12:45	SMPL
N4598-05	MH0AA4	2647	0.5252	-	SFAM01.1	PPB	9/15/2022 12:48	SMPL
N4598-06	MH0AA5	2355	0.4629	-	SFAM01.1	PPB	9/15/2022 12:50	SMPL
N4598-07	MH0AA6	18990	4.0101	-	SFAM01.1	PPB	9/15/2022 12:52	SMPL
N4598-08	MH0AA7	22153	4.6846	-	SFAM01.1	PPB	9/15/2022 12:55	SMPL
N4598-09	MH0AA8	56399	11.9872	-	SFAM01.1	PPB	9/15/2022 12:57	SMPL
N4598-10	MH0AA9	14905	3.1391	-	SFAM01.1	PPB	9/15/2022 12:59	SMPL

LB122012

INSTRUMENT ID : CV1

N4598-11	MH0AB0	52519	11.1599 -	SFAM01.1	PPB	9/15/2022 13:02	SMPL
N4598-12	MH0AB1	7875	1.64 -	SFAM01.1	PPB	9/15/2022 13:05	SMPL
N4598-13	MH0AB2	48504	10.3037 -	SFAM01.1	PPB	9/15/2022 13:07	SMPL
N4598-14	MH0AB3	2673	0.5307 -	SFAM01.1	PPB	9/15/2022 13:09	SMPL
CCV008	CCV008	23383	4.9469 -	SFAM01.1	PPB	9/15/2022 13:12	SMPL
CCB008	CCB008	-7	-0.0408 -	SFAM01.1	PPB	9/15/2022 13:14	SMPL
N4598-15	MH0AB4	24768	5.2422 -	SFAM01.1	PPB	9/15/2022 13:16	SMPL
N4598-16	MH0AB5	1398	0.2588 -	SFAM01.1	PPB	9/15/2022 13:18	SMPL
N4598-17	MH0AB6	1515	0.2838 -	SFAM01.1	PPB	9/15/2022 13:21	SMPL
N4598-18	MH0AB6D	1518	0.2844 -	SFAM01.1	PPB	9/15/2022 13:26	SMPL
N4598-19	MH0AB6S	12050	2.5303 -	SFAM01.1	PPB	9/15/2022 13:28	SMPL
N4598-20	MH0AB7	19443	4.1067 -	SFAM01.1	PPB	9/15/2022 13:30	SMPL
N4598-21	MH0AB8	29013	6.1474 -	SFAM01.1	PPB	9/15/2022 13:32	SMPL
N4598-22	MH0AB9	2768	0.551 -	SFAM01.1	PPB	9/15/2022 13:35	SMPL
PB147661BL	PBS661	-54	-0.0508 -	SFAM01.1	PPB	9/15/2022 13:37	SMPL
N4015-03	MDL-SOIL	381	0.042 -	SFAM01.1	PPB	9/15/2022 13:42	SMPL
N4542-16DLX5	MF5A24	13082	2.7503 -	SFAM01.1	PPB	9/15/2022 13:44	SMPL
N4542-21DLX10	MF5A29	192775	41.068 -	SFAM01.1	PPB	9/15/2022 13:46	SMPL
N4598-09DLX2	MH0AA8	28514	6.041 -	SFAM01.1	PPB	9/15/2022 13:49	SMPL
N4598-11DLX2	MH0AB0	26627	5.6387 -	SFAM01.1	PPB	9/15/2022 13:51	SMPL
N4598-13DLX2	MH0AB2	23823	5.0407 -	SFAM01.1	PPB	9/15/2022 13:53	SMPL
N4542-21DL2X100	MF5A29	21250	4.4921 -	SFAM01.1	PPB	9/15/2022 13:55	SMPL
CCV009	CCV009	23415	4.9537 -	SFAM01.1	PPB	9/15/2022 13:58	SMPL
CCB009	CCB009	-13	-0.0421 -	SFAM01.1	PPB	9/15/2022 14:00	SMPL