



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB122016

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/15/2022	14:10
2	CAL2	1	Water	NA	NA	20.7	7.00	09/15/2022	14:11
3	CAL3	1	Water	NA	NA	20.7	10.03	09/15/2022	14:12
4	ICV	1	Water	NA	NA	20.8	6.99	09/15/2022	14:15
5	CCV1	1	Water	NA	NA	20.8	2.01	09/15/2022	14:20
6	N4628-02	1	Water	NA	NA	23.1	6.74	09/15/2022	14:22
7	N4628-04	1	Water	NA	NA	23.0	7.00	09/15/2022	14:25
8	N4628-06	1	Water	NA	NA	22.6	7.30	09/15/2022	14:26
9	N4663-01	1	Water	NA	NA	20.2	7.16	09/15/2022	14:33
10	N4663-03	1	Water	NA	NA	20.2	6.76	09/15/2022	14:40
11	N4663-05	1	Water	NA	NA	20.3	6.74	09/15/2022	14:42
12	N4663-05DUP	1	Water	NA	NA	20.2	6.75	09/15/2022	14:43
13	CCV2	1	Water	NA	NA	20.8	12.03	09/15/2022	14:45

WORKLIST(Hardcopy Internal Chain)

13122016

WorkList Name : ph n4663

WorkList ID : 163208

Department : Wet-Chemistry

Date : 09-15-2022 14:53:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/13/2022	Water	N4628-02	pH	H	Cool 4 deg C	DVIR01	N11	HW-54	09/13/2022 9040C
09/13/2022	Water	N4628-04	pH	H	Cool 4 deg C	DVIR01	N11	MW-7	09/13/2022 9040C
09/13/2022	Water	N4628-06	pH	H	Cool 4 deg C	DVIR01	N11	HW-8	09/13/2022 9040C
09/14/2022	Water	N4663-01	pH	H	Cool 4 deg C	DVIR01	N11	HW-43	09/14/2022 9040C
09/14/2022	Water	N4663-03	pH	H	Cool 4 deg C	DVIR01	N11	SVE-03	09/14/2022 9040C
09/14/2022	Water	N4663-05	pH	H	Cool 4 deg C	DVIR01	N11	HW-9	09/14/2022 9040C

Date/Time 09.15.22 14:00

Raw Sample Received by: 101 WCC

Raw Sample Relinquished by: CP SM

Date/Time

09.15.22

17:30

Raw Sample Received by:

CP SM

Raw Sample Relinquished by:

Page 1 of 1

10 WCC