

LB122623


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53883



A= 0.0000e+000

B= 1.8486e-004

C= -2.2292e-002

Rho= 0.9998978

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.0	0.000	-0.016	-0.016	34	0.000	34					—
0.2	0.200	0.208	0.008	1246	0.0 %	1246					4
2.5	2.500	2.442	-0.058	13328	0.0 %	13328					-2
5.0	5.000	5.087	0.087	27637	0.0 %	27637					2
7.5	7.500	7.541	0.041	40913	0.0 %	40913					1
10.0	10.000	9.939	-0.061	53883	0.0 %	53883					-1

LB122023 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Sep 2022 15:03:07

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7471B	PPB	15 Sep 2022 15:20:29	S	Std
0.2 - 1	50.2	1246	-	0.2000	7471B	PPB	15 Sep 2022 15:22:47	S	Std
2.5 - 1	52.5	13328	-	2.5000	7471B	PPB	15 Sep 2022 15:25:04	S	Std
5.0 - 1	55	27637	-	5.0000	7471B	PPB	15 Sep 2022 15:27:21	S	Std
7.5 - 1	57.5	40913	-	7.5000	7471B	PPB	15 Sep 2022 15:29:39	S	Std
10.0 - 1	510	53883	-	10.0000	7471B	PPB	15 Sep 2022 15:31:57	S	Std
ICV31 - 1	ICV31	21442	3.9415	-	7471B	PPB	15 Sep 2022 15:34:45	U	SMPL
ICB31 - 1	ICB31	-17	-0.0254	-	7471B	PPB	15 Sep 2022 15:36:59	U	SMPL
CCV10 - 1	CCV10	27554	5.0714	-	7471B	PPB	15 Sep 2022 15:39:16	U	SMPL
CCB10 - 1	CCB10	-46	-0.0308	-	7471B	PPB	15 Sep 2022 15:41:31	U	SMPL
CRA - 1	CRA	1209	0.2012	-	7471B	PPB	15 Sep 2022 15:43:48	U	SMPL
HighStd - 1	HighStd	53989	9.9583	-	7471B	PPB	15 Sep 2022 15:46:03	U	SMPL
ChkStd - 1	ChkStd	38913	7.1713	-	7471B	PPB	15 Sep 2022 15:48:19	U	SMPL
PB147664BL - 1	PBS	-86	-0.0382	-	7471B	PPB	15 Sep 2022 15:50:37	U	SMPL
PB147664BS - 1	LCSS	21533	3.9584	-	7471B	PPB	15 Sep 2022 15:52:26	U	SMPL
N4600-01 - 1	OH-SOIL-0913	5509	0.9961	-	7471B	PPB	15 Sep 2022 15:57:42	U	SMPL
N4612-01 - 1	CL-01-091322	3816	0.6831	-	7471B	PPB	15 Sep 2022 15:59:59	U	SMPL
N4613-01 - 1	SASP2-T-6-(10-18)	4155	0.7458	-	7471B	PPB	15 Sep 2022 16:02:16	U	SMPL
N4613-02 - 1	SASP2-T-6-(18-23)	120	-0.0001	-	7471B	PPB	15 Sep 2022 16:04:34	U	SMPL
N4613-03 - 1	SASP2-T-6-(25-37)	169	0.0089	-	7471B	PPB	15 Sep 2022 16:06:52	U	SMPL
CCV11 - 1	CCV11	27595	5.0790	-	7471B	PPB	15 Sep 2022 16:09:07	U	SMPL
CCB11 - 1	CCB11	-34	-0.0286	-	7471B	PPB	15 Sep 2022 16:11:22	U	SMPL
N4627-01 - 1	EO-3-091422	11985	2.1933	-	7471B	PPB	15 Sep 2022 16:13:39	U	SMPL
N4630-01 - 1	WC-17-4-1	321302	59.3744	-	7471B	PPB	15 Sep 2022 16:15:54	U	SMPL
N4630-06 - 1	WC-17-4-2	371670	68.6856	-	7471B	PPB	15 Sep 2022 16:18:10	U	SMPL
N4630-11 - 1	WC-17-4-3	242616	44.8283	-	7471B	PPB	15 Sep 2022 16:20:40	U	SMPL
N4630-16 - 1	WC-17-4-4	258020	47.6760	-	7471B	PPB	15 Sep 2022 16:23:09	U	SMPL
N4630-21 - 1	WC-17-4-5	200913	37.1190	-	7471B	PPB	15 Sep 2022 16:25:35	U	SMPL
N4634-01 - 1	WC-1	1091	0.1794	-	7471B	PPB	15 Sep 2022 16:28:04	U	SMPL
N4634-04 - 1	WC-2	3396	0.6055	-	7471B	PPB	15 Sep 2022 16:30:26	U	SMPL
N4640-01 - 1	VNJ-240	1831	0.3162	-	7471B	PPB	15 Sep 2022 16:32:44	U	SMPL
N4640-03 - 1	72-11930	1510	0.2569	-	7471B	PPB	15 Sep 2022 16:34:59	U	SMPL
CCV12 - 1	CCV12	27890	5.1335	-	7471B	PPB	15 Sep 2022 16:37:15	U	SMPL
CCB12 - 1	CCB12	-27	-0.0273	-	7471B	PPB	15 Sep 2022 16:39:30	U	SMPL
N4644-01 - 1	TR-6-091422	4512	0.8118	-	7471B	PPB	15 Sep 2022 16:46:37	U	SMPL
N4644-01DUP - 1	TR-6-091422DUP	4701	0.8467	-	7471B	PPB	15 Sep 2022 16:48:53	U	SMPL
N4644-01MS - 1	TR-6-091422MS	27701	5.0986	-	7471B	PPB	15 Sep 2022 16:53:40	U	SMPL
N4644-01MSD - 1	TR-6-091422MSD	28009	5.1555	-	7471B	PPB	15 Sep 2022 16:55:55	U	SMPL
N4644-01LX5 - 1		1803	0.3110	-	7471B	PPB	15 Sep 2022 16:58:13	U	SMPL
N4644-01A - 1		27495	5.0605	-	7471B	PPB	15 Sep 2022 17:00:30	U	SMPL
CCV13 - 1	CCV13	27478	5.0574	-	7471B	PPB	15 Sep 2022 17:02:47	U	SMPL
CCB13 - 1	CCB13	-60	-0.0334	-	7471B	PPB	15 Sep 2022 17:05:05	U	SMPL