



PERCENT SOLID

Supervisor: Sohil
Analyst: Nilesh
Date: 9/19/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 09/16/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 09/17/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122027

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4656-01	SP-1	1	0.91	8.83	9.74	9.1	92.8	
N4656-03	SP-1A	2	0.91	9.00	9.91	9.39	94.2	
N4656-04	SP-1B	3	0.92	8.88	9.8	9.33	94.7	
N4656-05	SP-1C	4	0.91	9.04	9.95	9.44	94.4	
N4656-06	SP-1D	5	0.92	8.86	9.78	9.28	94.4	
N4656-07	SP-1E	6	0.92	8.80	9.72	9.02	92.0	
N4656-08	SP-1F	7	0.91	8.95	9.86	9.11	91.6	
N4656-09	SP-1G	8	0.91	9.05	9.96	9.3	92.7	
N4656-10	SP-1H	9	0.91	9.06	9.97	9.26	92.2	
N4656-11	SP-2	10	0.91	9.01	9.92	9.35	93.7	
N4656-13	SP-2A	11	0.91	9.06	9.97	9.43	94.0	
N4656-14	SP-2B	12	0.9	9.03	9.93	9.34	93.5	
N4656-15	SP-2C	13	0.9	8.94	9.84	9.33	94.3	
N4656-16	SP-2D	14	0.91	8.91	9.82	9.13	92.3	
N4656-17	SP-2E	15	0.91	8.98	9.89	9.25	92.9	
N4656-18	SP-2F	16	0.92	8.85	9.77	9.21	93.7	
N4656-19	SP-2G	17	0.9	8.93	9.83	9.2	92.9	
N4656-20	SP-2H	18	0.91	8.87	9.78	9.26	94.1	
N4656-21	SP-3	19	0.9	8.85	9.75	9.08	92.4	
N4656-23	SP-3A	20	0.91	8.89	9.8	9.17	92.9	
N4656-24	SP-3B	21	0.9	8.94	9.84	9.26	93.5	
N4656-25	SP-3C	22	0.9	8.92	9.82	9.02	91.0	
N4656-26	SP-3D	23	0.91	8.77	9.68	9.12	93.6	
N4656-27	SP-3E	24	0.92	9.01	9.93	9.28	92.8	
N4656-28	SP-3F	25	0.91	9.06	9.97	9.1	90.4	
N4656-29	SP-3G	26	0.91	9.08	9.99	9.29	92.3	
N4656-30	SP-3H	27	0.91	8.83	9.74	9.03	92.0	
N4656-31	SP-4	28	0.92	8.83	9.75	9.16	93.3	



PERCENT SOLID

Supervisor: Sohil
Analyst: Nilesh
Date: 9/19/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 09/16/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 09/17/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122027

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4656-33	SP-4A	29	0.91	9.02	9.93	9.4	94.1	
N4656-34	SP-4B	30	0.91	9.02	9.93	9.37	93.8	
N4656-35	SP-4C	31	0.9	9.03	9.93	9.39	94.0	
N4656-36	SP-4D	32	0.9	8.89	9.79	9.03	91.5	
N4656-37	SP-4E	33	0.91	8.93	9.84	8.87	89.1	
N4656-38	SP-4F	34	0.91	8.75	9.66	8.73	89.4	
N4656-39	SP-4G	35	0.92	8.76	9.68	9.3	95.7	
N4656-40	SP-4H	36	0.91	8.89	9.8	9.27	94.0	
N4668-01	D22007-8.5	47	0.92	8.76	9.68	8.23	83.4	
N4668-02	D22007-17.5	48	0.91	8.91	9.82	7.26	71.3	
N4668-03	D22006-5.5	49	0.9	9.09	9.99	8.84	87.3	
N4668-04	D22006-18	50	0.91	8.86	9.77	7.79	77.7	
N4668-05	A2-9-18	51	0.91	8.74	9.65	7.55	76.0	
N4668-06	D22003-8	52	0.91	8.97	9.88	7.07	68.7	
N4668-07	D22003-18	53	0.91	9.08	9.99	7.72	75.0	
N4668-08	A5-6-18	54	0.92	9.05	9.97	8.27	81.2	
N4668-09	A1-6-18	55	0.9	9.08	9.98	8.04	78.6	
N4668-10	D22016-13.5	56	0.91	9.04	9.95	8.11	79.6	
N4668-11	D22016-18	57	0.91	8.92	9.83	7.71	76.2	
N4668-12	A3-8-18	58	0.91	9.08	9.99	8.1	79.2	
N4673-01	JC-520	61	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4673-02	JC-521	62	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4673-03	JC-522	63	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4673-04	JC-523	64	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4673-05	JC-524	65	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4673-06	JC-525	66	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4673-07	JC-526	67	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4673-08	JC527	68	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE



PERCENT SOLID

Supervisor: Sohil
Analyst: Nilesh
Date: 9/19/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 09/16/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 09/17/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122027

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4690-01	40112	69	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4690-02	40113	70	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4690-03	40114	71	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4690-04	40115	72	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4690-05	40116	73	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4691-05	SVOC-GPC-BLANK	37	1.00	1.00	2.00	2.00	100.0	
N4691-06	PEST-GPC-BLANK	38	1.00	1.00	2.00	2.00	100.0	
N4691-07	PEST-GPC-BLANK-SPIKE	39	1.00	1.00	2.00	2.00	100.0	
N4691-08	PCB-GPC-BLANK	40	1.00	1.00	2.00	2.00	100.0	
N4691-09	PCB-GPC-BLANK-SPIKE	41	1.00	1.00	2.00	2.00	100.0	
N4691-10	SVOC-GPC2-BLANK	42	1.00	1.00	2.00	2.00	100.0	
N4691-11	PEST-GPC2-BLANK	43	1.00	1.00	2.00	2.00	100.0	
N4691-12	PEST-GPC2-BLANK-SPIKE	44	1.00	1.00	2.00	2.00	100.0	
N4691-13	PCB-GCP2-BLANK	45	1.00	1.00	2.00	2.00	100.0	
N4691-14	PCB-GCP2-BLANK-SPIKE	46	1.00	1.00	2.00	2.00	100.0	
N4695-01	INCINERATOR-ASH	59	0.89	8.68	9.57	8.02	82.1	
N4695-02	INCINERATOR-ASH	60	0.89	8.68	9.57	8.02	82.1	
N4697-01	SB-22-7.5-8.0	76	0.91	8.97	9.88	8.98	90.0	
N4697-02	SB-22-COMP	77	0.91	9.06	9.97	9.16	91.1	
N4697-03	SB-23-6.5-7.0	78	0.93	8.78	9.71	8.85	90.2	
N4697-04	SB-23-COMP	79	0.93	9.05	9.98	8.84	87.4	
N4697-05	SB-24-6.5-7.0	80	0.91	8.77	9.68	8.27	83.9	
N4697-06	SB-24-COMP	81	0.93	9.02	9.95	8.74	86.6	
N4697-07	SB-31-6.5-7.0	82	0.92	9.04	9.96	9.16	91.2	
N4697-08	SB-31-COMP	83	0.92	8.98	9.9	8.42	83.5	
N4697-09	SB-21-6.5-7.0	84	0.91	9.05	9.96	7.57	73.6	
N4697-10	SB-21-COMP	85	0.91	9.02	9.93	8.23	81.2	
N4697-11	SB-20-6.5-7.0	86	0.93	8.63	9.56	8.09	83.0	



PERCENT SOLID

Supervisor: Sohil
Analyst: Nilesch
Date: 9/19/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 09/16/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 09/17/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122027

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4697-12	SB-20-COMP	87	0.96	8.83	9.79	8.66	87.2	
N4697-13	SB-19-6.5-7.0	88	0.91	9.06	9.97	8.42	82.9	
N4697-14	SB-19-COMP	89	0.91	8.97	9.88	8.55	85.2	
N4697-15	SB-18-6.5-7.0	90	0.91	9.05	9.96	8.64	85.4	
N4697-16	SB-18-COMP	91	0.98	8.85	9.83	8.34	83.2	
N4697-17	SB-25-6.5-7.0	92	0.93	8.90	9.83	7.84	77.6	
N4697-18	SB-25-COMP	93	0.95	9.01	9.96	8.3	81.6	
N4697-19	SB-17-6.5-7.0	94	0.92	8.92	9.84	8.38	83.6	
N4697-20	SB-17-COMP	95	0.93	8.84	9.77	8.48	85.4	
N4701-01	OK-02-091622	74	0.93	8.70	9.63	9.19	94.9	
N4701-02	OK-02-091622-EPH-2	75	0.91	9.02	9.93	8.62	85.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$