

LB122029


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

49169



A= 0.0000e+000

B= 2.0979e-004

C= -8.8022e-002

Rho= 0.9980096

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4/5/1
0.0	0.000	-0.048	-0.048	191	0.000	191					
0.2	0.200	0.181	-0.019	1283	0.0 %	1283					
2.5	2.500	2.836	0.336	13938	0.0 %	13938					
5.0	5.000	4.884	-0.116	23700	0.0 %	23700					
7.5	7.500	7.120	-0.380	34359	0.0 %	34359					
10.0	10.000	10.227	0.227	49169	0.0 %	49169					

4/5/1
 1
 10
 13
 1
 1
 5
 2

LB122029 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Sep 2022 10:25:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	191	-	0.0000	7470A	PPB	16 Sep 2022 10:34:04	S	Std
0.2 - 1	50.2	1283	-	0.2000	7470A	PPB	16 Sep 2022 10:39:08	S	Std
2.5 - 1	52.5	13938	-	2.5000	7470A	PPB	16 Sep 2022 10:41:24	S	Std
5.0 - 1	55	23700	-	5.0000	7470A	PPB	16 Sep 2022 10:43:41	S	Std
7.5 - 1	59.5	34359	-	7.5000	7470A	PPB	16 Sep 2022 10:45:59	S	Std
10.0 - 1	51.8	49169	-	10.0000	7470A	PPB	16 Sep 2022 10:50:51	S	Std
ICV32 - 1	ICV32	19410	3.9839	-	7470A	PPB	16 Sep 2022 10:57:37	U	SMPL
ICB32 - 1	ICB32	187	-0.0488	-	7470A	PPB	16 Sep 2022 10:59:52	U	SMPL
CCV14 - 1	CCV14	23031	4.7436	-	7470A	PPB	16 Sep 2022 11:02:08	U	SMPL
CCB14 - 1	CCB14	166	-0.0532	-	7470A	PPB	16 Sep 2022 11:04:23	U	SMPL
CRA - 1	CRA	1199	0.1635	-	7470A	PPB	16 Sep 2022 11:06:39	U	SMPL
HighStd - 1	HighStd	49052	10.2024	-	7470A	PPB	16 Sep 2022 11:08:53	U	SMPL
ChkStd - 1	ChkStd	33581	6.9568	-	7470A	PPB	16 Sep 2022 11:11:09	U	SMPL
PB147666BL - 1	PBW	90	-0.0691	-	7470A	PPB	16 Sep 2022 11:15:56	U	SMPL
PB147666BS - 1	LCSW	19440	3.9902	-	7470A	PPB	16 Sep 2022 11:18:11	U	SMPL
N4597-02 - 1	SB-01-COMP	-63	-0.1012	-	7470A	PPB	16 Sep 2022 11:20:28	U	SMPL
N4597-04 - 1	SB-02-COMP	-24	-0.0931	-	7470A	PPB	16 Sep 2022 11:22:46	U	SMPL
N4597-06 - 1	SB-03-COMP	-18	-0.0918	-	7470A	PPB	16 Sep 2022 11:25:03	U	SMPL
N4600-02 - 1	OH-SOIL-0913	68	-0.0738	-	7470A	PPB	16 Sep 2022 11:27:20	U	SMPL
N4602-01 - 1	COMP1-0913	101	-0.0668	-	7470A	PPB	16 Sep 2022 11:29:35	U	SMPL
CCV15 - 1	CCV15	23445	4.8304	-	7470A	PPB	16 Sep 2022 11:31:50	U	SMPL
CCB15 - 1	CCB15	137	-0.0593	-	7470A	PPB	16 Sep 2022 11:34:05	U	SMPL
N4602-02 - 1	COMP2-0913	55	-0.0765	-	7470A	PPB	16 Sep 2022 11:36:21	U	SMPL
N4602-03 - 1	17192	105	-0.0660	-	7470A	PPB	16 Sep 2022 11:38:37	U	SMPL
N4630-02 - 1	WC-17-4-1	155	-0.0555	-	7470A	PPB	16 Sep 2022 11:40:53	U	SMPL
N4630-07 - 1	WC-17-4-2	144	-0.0578	-	7470A	PPB	16 Sep 2022 11:43:09	U	SMPL
N4630-12 - 1	WC-17-4-3	99	-0.0673	-	7470A	PPB	16 Sep 2022 11:45:25	U	SMPL
N4630-17 - 1	WC-17-4-4	69	-0.0735	-	7470A	PPB	16 Sep 2022 11:47:41	U	SMPL
N4630-22 - 1	WC-17-4-5	47	-0.0782	-	7470A	PPB	16 Sep 2022 11:49:58	U	SMPL
N4634-03 - 1	WC-1	63	-0.0748	-	7470A	PPB	16 Sep 2022 11:52:15	U	SMPL
N4634-06 - 1	WC-2	45	-0.0786	-	7470A	PPB	16 Sep 2022 11:54:33	U	SMPL
N4634-06DUP - 1	WC-2DUP	38	-0.0800	-	7470A	PPB	16 Sep 2022 11:56:48	U	SMPL
CCV16 - 1	CCV16	23353	4.8111	-	7470A	PPB	16 Sep 2022 11:59:04	U	SMPL
CCB16 - 1	CCB16	-31	-0.0945	-	7470A	PPB	16 Sep 2022 12:01:19	U	SMPL
N4634-06MS - 1	WC-2MS	18424	3.7771	-	7470A	PPB	16 Sep 2022 12:03:36	U	SMPL
N4634-06MSD - 1	WC-2MSD	18936	3.8845	-	7470A	PPB	16 Sep 2022 12:05:52	U	SMPL
PB147625TB - 1	PB147625TB	-15	-0.0912	-	7470A	PPB	16 Sep 2022 12:08:07	U	SMPL
N4634-06LX5 - 1		-181	-0.1260	-	7470A	PPB	16 Sep 2022 12:10:23	U	SMPL
N4634-06A - 1		23174	4.7736	-	7470A	PPB	16 Sep 2022 12:12:39	U	SMPL
CCV17 - 1	CCV17	24589	5.0704	-	7470A	PPB	16 Sep 2022 12:14:55	U	SMPL
CCB17 - 1	CCB17	61	-0.0752	-	7470A	PPB	16 Sep 2022 12:17:12	U	SMPL