



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB122034
Analysis Date: 09/16/2022
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Sohil
Extraction Date: 09/16/2022
Extraction IN Time: 08:15
Extraction OUT Time: 09:00
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB122034BL	LB122034BL	WATER	1.3	1000	100	2.9687	2.9687	3.02	2.9688	2.9688	0.0001	0.1
2	LB122034BS	LB122034BS	WATER	1.3	1000	100	3.0569	3.0569	3.01	3.0735	3.0735	0.0166	16.6
3	LB122034BSD	LB122034BSD	WATER	1.3	1000	100	3.1474	3.1474	3.03	3.1644	3.1644	0.0170	17
4	N4628-02	HW-54	WATER	1.6	1000	100	2.9557	2.9557	3.04	2.9591	2.9591	0.0034	3.4
5	N4628-04	MW-7	WATER	1.6	1000	100	3.0113	3.0113	3.03	3.0430	3.0430	0.0317	31.7
6	N4628-06	HW-8	WATER	1.6	1000	100	3.0497	3.0497	3.04	3.0549	3.0549	0.0052	5.2
7	N4643-01	GRAB	WATER	1.6	1000	100	3.0904	3.0904	3.05	3.2111	3.2111	0.1207	120.7
8	N4663-01	HW-43	WATER	1.6	1000	100	2.9436	2.9436	3.02	2.9450	2.9450	0.0014	1.4
9	N4663-03	SVE-03	WATER	1.6	1000	100	3.0747	3.0747	3.03	3.0756	3.0756	0.0009	0.9
10	N4663-05	HW-9	WATER	1.6	1000	100	2.9692	2.9692	3.04	2.9693	2.9693	0.0001	0.1



QC Batch# LB122034

Test: TPH

Analysis Date: 09/16/2022

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2938
pH Paper 0-14	M4909
Sodium Sulfate	EP2254
1:1 HCL	WP98461
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP98457
LCSWD	5.00 ML	WP98458
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 10:01

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time1: 09:15

Bal Check Time: 08:30 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 10:30

Out Time1: 10:00

After Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 11:31

1.0000 gram Balance: 1.0003 (0.9950-1.0050) In Time2: 11:00

Bal Check Time: 12:10 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 12:00

Out Time2: 11:30

18/22034

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph n4563 WorkList ID : 163212 Department : Wet-Chemistry Date : 09-16-2022 08:06:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/19/2022	Water	N4628-02	TPH	3	Conc H2SO4 to	DVIR01	N11	HW-54	09/13/2022 1664A
09/19/2022	Water	N4628-04	TPH	3	Conc H2SO4 to	DVIR01	N11	MW-7	09/13/2022 1664A
09/19/2022	Water	N4628-06	TPH	3	Conc H2SO4 to	DVIR01	N11	HW-8	09/13/2022 1664A
09/27/2022	Water	N4643-01	TPH		Conc H2SO4 to	ARAM01	N11	GRAB	09/14/2022 1664A
09/20/2022	Water	N4663-01	TPH	K	Conc H2SO4 to	DVIR01	N11	HW-43	09/14/2022 1664A
09/20/2022	Water	N4663-03	TPH	3	Conc H2SO4 to	DVIR01	N11	SVE-03	09/14/2022 1664A
09/20/2022	Water	N4663-05	TPH	3	Conc H2SO4 to	DVIR01	N11	HW-9	09/14/2022 1664A

Date/Time 09-16-22 08:10
Raw Sample Received by: JN CWC
Raw Sample Relinquished by: JN CWC

Date/Time 09-16-22 10:30
Raw Sample Received by: JN CWC
Raw Sample Relinquished by: JN CWC