

LB122041


SFAM01.1

INSTRUMENT ID: CVL

Linear

 μ Abs.:

46928



A= 0.0000e+000

B= 2.1111e-004 slope

C= -2.9418e-002 intercept

Rho= 0.9997736

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% of
0.00	0.000	-0.032	-0.032	-13	0.000	-13					=
0.05	0.050					185	0				-7
0.20	0.200	0.187	-0.013	1024	0.0 %	1024					0
2.50	2.500	2.502	0.002	11989	0.0 %	11989					0
5.00	5.000	5.024	0.024	23939	0.0 %	23939					2
7.50	7.500	7.642	0.142	36339	0.0 %	36339					-1
10.0	10.000	9.877	-0.123	46928	0.0 %	46928					

LB122041

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	-13	-		0 SFAM01.1	PPB	9/16/2022 13:50	Std
	0.2	1024	-		0.2 SFAM01.1	PPB	9/16/2022 13:52	Std
	2.5	11989	-		2.5 SFAM01.1	PPB	9/16/2022 13:56	Std
	5	23939	-		5 SFAM01.1	PPB	9/16/2022 13:58	Std
	7.5	36339	-		7.5 SFAM01.1	PPB	9/16/2022 14:00	Std
	10	46928	-		10 SFAM01.1	PPB	9/16/2022 14:03	Std
ICV033	ICV033	19088	4.0002 -		SFAM01.1	PPB	9/16/2022 14:05	SMPL
ICB033	ICB033	-119	-0.0545 -		SFAM01.1	PPB	9/16/2022 14:08	SMPL
CCV018	CCV018	24078	5.0537 -		SFAM01.1	PPB	9/16/2022 14:10	SMPL
CCB018	CCB018	-129	-0.0567 -		SFAM01.1	PPB	9/16/2022 14:12	SMPL
PB147702BL	PBS702	-57	-0.0415 -		SFAM01.1	PPB	9/16/2022 14:14	SMPL
N4015-04	MDL-SOIL	364	0.0474 -		SFAM01.1	PPB	9/16/2022 14:19	SMPL
PB147703BL	PBS703	-38	-0.0374 -		SFAM01.1	PPB	9/16/2022 14:22	SMPL
N4522-01	MF5A00	806	0.1407 -		SFAM01.1	PPB	9/16/2022 14:24	SMPL
N4522-02	MF5A00D	773	0.1338 -		SFAM01.1	PPB	9/16/2022 14:26	SMPL
N4522-03	MF5A00S	12838	2.6808 -		SFAM01.1	PPB	9/16/2022 14:28	SMPL
N4522-04	MF5A01	1105	0.2039 -		SFAM01.1	PPB	9/16/2022 14:31	SMPL
N4522-05	MF5A02	1970	0.3865 -		SFAM01.1	PPB	9/16/2022 14:33	SMPL
N4522-06	MF5A06	2194	0.4338 -		SFAM01.1	PPB	9/16/2022 14:35	SMPL
N4522-07	MF5A07	927	0.1663 -		SFAM01.1	PPB	9/16/2022 14:37	SMPL
N4522-08	MF5A08	769	0.1329 -		SFAM01.1	PPB	9/16/2022 14:40	SMPL
N4522-09	MF5A13	1332	0.2518 -		SFAM01.1	PPB	9/16/2022 14:42	SMPL
N4522-10	MF5A14	3514	0.7124 -		SFAM01.1	PPB	9/16/2022 14:44	SMPL
N4522-11	MF5A16	852	0.1504 -		SFAM01.1	PPB	9/16/2022 14:47	SMPL
N4522-12	MF5A17	902	0.161 -		SFAM01.1	PPB	9/16/2022 14:49	SMPL
N4522-13	MF5A19	749	0.1287 -		SFAM01.1	PPB	9/16/2022 14:51	SMPL
N4522-14	MF5A20	589	0.0949 -		SFAM01.1	PPB	9/16/2022 14:53	SMPL
N4522-15	MF5A40	648	0.1074 -		SFAM01.1	PPB	9/16/2022 14:56	SMPL
N4522-16	MF5A41	739	0.1266 -		SFAM01.1	PPB	9/16/2022 14:58	SMPL
PB147704BL	PBS704	-62	-0.0425 -		SFAM01.1	PPB	9/16/2022 15:00	SMPL
CCV019	CCV019	24181	5.0754 -		SFAM01.1	PPB	9/16/2022 15:02	SMPL
CCB019	CCB019	-111	-0.0529 -		SFAM01.1	PPB	9/16/2022 15:05	SMPL
N4560-01	MF5A31	62969	13.2639 -		SFAM01.1	PPB	9/16/2022 15:07	SMPL
N4560-02	MF5A32	857345	HIGH -		SFAM01.1	PPB	9/16/2022 15:09	
N4560-03	MF5A33	633691	HIGH -		SFAM01.1	PPB	9/16/2022 15:11	
N4560-04	MF5A34	568558	HIGH -		SFAM01.1	PPB	9/16/2022 15:13	
N4560-05	MF5A43	740039	HIGH -		SFAM01.1	PPB	9/16/2022 15:15	
N4560-06	MF5A43D	731989	HIGH -		SFAM01.1	PPB	9/16/2022 15:17	
N4560-07	MF5A43S	738609	HIGH -		SFAM01.1	PPB	9/16/2022 15:19	
N4560-01DLX5		12762	2.6647 -		SFAM01.1	PPB	9/16/2022 15:26	SMPL
N4560-02DLX10		183172	38.6397 -		SFAM01.1	PPB	9/16/2022 15:28	SMPL
N4560-03DLX10		109288	23.0422 -		SFAM01.1	PPB	9/16/2022 15:31	SMPL
N4560-04DLX10		103595	21.8404 -		SFAM01.1	PPB	9/16/2022 15:33	SMPL
N4560-05DLX10		142909	30.1399 -		SFAM01.1	PPB	9/16/2022 15:35	SMPL
N4560-06DLX10		140738	29.6816 -		SFAM01.1	PPB	9/16/2022 15:38	SMPL
N4560-07DLX10		144377	30.4498 -		SFAM01.1	PPB	9/16/2022 15:40	SMPL

LB122041

INSTRUMENT ID : CV1

CCV020	CCV020	22197	4.6566 -	SFAM01.1	PPB	9/16/2022 15:42	SMPL
CCB020	CCB020	-189	-0.0693 -	SFAM01.1	PPB	9/16/2022 15:45	SMPL
N4560-02DL2X100		19764	4.1429 -	SFAM01.1	PPB	9/16/2022 15:47	SMPL
N4560-03DL2X50		22310	4.6804 -	SFAM01.1	PPB	9/16/2022 15:49	SMPL
N4560-04DL2X50		21500	4.5094 -	SFAM01.1	PPB	9/16/2022 15:51	SMPL
N4560-05DL2X100		15343	3.2096 -	SFAM01.1	PPB	9/16/2022 15:54	SMPL
N4560-06DL2X100		15009	3.1391 -	SFAM01.1	PPB	9/16/2022 15:56	SMPL
N4560-07DL2X100		15053	3.1484 -	SFAM01.1	PPB	9/16/2022 15:58	SMPL
CCV021	CCV021	24035	5.0446 -	SFAM01.1	PPB	9/16/2022 16:00	SMPL
CCB021	CCB021	-129	-0.0567 -	SFAM01.1	PPB	9/16/2022 16:03	SMPL