



Analytical Summary Report

Analysis Method: 9041A

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB122045

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/16/2022	15:05
2	CAL2	1	Water	NA	NA	20.7	7.01	09/16/2022	15:06
3	CAL3	1	Water	NA	NA	20.7	10.03	09/16/2022	15:07
4	ICV	1	Water	NA	NA	20.7	6.98	09/16/2022	15:10
5	CCV1	1	Water	NA	NA	20.6	2.01	09/16/2022	15:11
6	N4695-01	1	Solid	20.03	20	23.6	6.51	09/16/2022	15:15
7	N4695-01DUP	1	Solid	20.04	20	23.7	6.52	09/16/2022	15:16
8	CCV2	1	Water	NA	NA	20.6	12.03	09/16/2022	15:20