



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122051

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	09/16/2022	16:15
2	CAL2	1	Water	NA	NA	20.8	7.00	09/16/2022	16:16
3	CAL3	1	Water	NA	NA	20.8	10.03	09/16/2022	16:17
4	ICV	1	Water	NA	NA	20.7	6.99	09/16/2022	16:19
5	CCV1	1	Water	NA	NA	20.7	2.01	09/16/2022	16:20
6	N4648-02	1	Water	NA	NA	22.7	7.59	09/16/2022	16:22
7	N4673-09	1	Water	NA	NA	23.3	8.86	09/16/2022	16:25
8	N4675-01	1	Water	NA	NA	20.1	8.84	09/16/2022	16:30
9	N4678-01	1	Water	NA	NA	20.1	8.01	09/16/2022	16:32
10	N4680-01	1	Water	NA	NA	20.2	8.02	09/16/2022	16:35
11	N4682-01	1	Water	NA	NA	20.3	8.92	09/16/2022	16:39
12	N4684-01	1	Water	NA	NA	20.7	8.19	09/16/2022	16:40
13	N4686-01	1	Water	NA	NA	20.4	8.10	09/16/2022	16:42
14	N4690-06	1	Water	NA	NA	22.9	6.95	09/16/2022	16:45
15	N4690-08	1	Water	NA	NA	23.4	6.11	09/16/2022	16:47
16	CCV2	1	Water	NA	NA	20.8	12.03	09/16/2022	16:50
17	N4692-01	1	Water	NA	NA	21.7	7.54	09/16/2022	16:51
18	N4692-02	1	Water	NA	NA	22.1	8.00	09/16/2022	17:00
19	N4692-05	1	Water	NA	NA	22.0	4.41	09/16/2022	17:02
20	N4692-05DUP	1	Water	NA	NA	22.1	4.42	09/16/2022	17:03

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	20.7	2.01	09/16/2022	17:05

1322051

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N4690 WorkList ID : 163260 Department : Wet-Chemistry Date : 09-16-2022 17:10:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/15/2022	Water	N4648-02	pH	A	Cool 4 deg C	PSEG03	N11	N48854-1040	09/15/2022 9040C
09/16/2022	Water	N4673-09	pH	A	Cool 4 deg C	PSEG03	N11	33633	09/16/2022 9040C
09/15/2022	Water	N4675-01	pH	C	Cool 4 deg C	ENVI70	N11	006-1	09/15/2022 9040C
09/15/2022	Water	N4678-01	pH	C	Cool 4 deg C	ENVI70	N11	5869-1	09/15/2022 9040C
09/15/2022	Water	N4680-01	pH	C	Cool 4 deg C	ENVI70	N11	5869-2	09/15/2022 9040C
09/15/2022	Water	N4682-01	pH	D	Cool 4 deg C	ENVI70	N11	006-2	09/15/2022 9040C
09/15/2022	Water	N4684-01	pH	D	Cool 4 deg C	ENVI70	N11	5486-2	09/15/2022 9040C
09/15/2022	Water	N4686-01	pH	E	Cool 4 deg C	ENVI70	N11	5486-1	09/15/2022 9040C
09/16/2022	Water	N4690-06	pH	A	Cool 4 deg C	PSEG03	N11	33829	09/16/2022 9040C
09/16/2022	Water	N4690-08	pH	A	Cool 4 deg C	PSEG03	N11	33808	09/16/2022 9040C
09/16/2022	Water	N4692-01	pH		Cool 4 deg C	PSEG03	N11	34202	09/16/2022 9040C
09/16/2022	Water	N4692-02	pH		Cool 4 deg C	PSEG03	N11	34203	09/16/2022 9040C
09/16/2022	Water	N4692-05	pH		Cool 4 deg C	PSEG03	N11	34175	09/16/2022 9040C

Date/Time 09/16/22 16:10
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC

Date/Time 09-16-22 18:00
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC