

LB122058


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

45932



A= 0.0000e+000

B= 2.2005e-004

C= -4.8777e-002

Rho= 0.9995223

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.017	-0.017	143	0.000	143					-
0.2	0.200	0.164	-0.036	969	0.0 %	969					-12
2.5	2.500	2.540	0.040	11766	0.0 %	11766					2
5.0	5.000	5.165	0.165	23693	0.0 %	23693					3
7.5	7.500	7.290	-0.210	33349	0.0 %	33349					-3
10.0	10.000	10.058	0.058	45932	0.0 %	45932					1

LB122058 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Sep 2022 09:58:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	143	-	0.0000	7471B	PPB	19 Sep 2022 10:19:25	S	Std
0.2 - 1	50.2	969	-	0.2000	7471B	PPB	19 Sep 2022 10:21:43	S	Std
2.5 - 1	52.5	11766	-	2.5000	7471B	PPB	19 Sep 2022 10:24:00	S	Std
5.0 - 1	55	23693	-	5.0000	7471B	PPB	19 Sep 2022 10:26:18	S	Std
7.5 - 1	57.5	33349	-	7.5000	7471B	PPB	19 Sep 2022 10:36:13	S	Std
10.0 - 1	59	45932	-	10.0000	7471B	PPB	19 Sep 2022 10:46:10	S	Std
ICV34 - 1	ICV34	19003	4.1327	-	7471B	PPB	19 Sep 2022 10:51:29	U	SMPL
ICB34 - 1	ICB34	29	-0.0424	-	7471B	PPB	19 Sep 2022 10:53:46	U	SMPL
CCV22 - 1	CCV22	25037	5.4605	-	7471B	PPB	19 Sep 2022 11:05:17	U	SMPL
CCB22 - 1	CCB22	-103	-0.0714	-	7471B	PPB	19 Sep 2022 11:07:33	U	SMPL
CRA - 1	CRA	1146	0.2034	-	7471B	PPB	19 Sep 2022 11:09:51	U	SMPL
HighStd - 1	HighStd	44465	9.7355	-	7471B	PPB	19 Sep 2022 11:12:06	U	SMPL
ChkStd - 1	ChkStd	31666	6.9192	-	7471B	PPB	19 Sep 2022 11:14:22	U	SMPL
PB147707BL - 1	PB147707BL	-136	-0.0787	-	7471B	PPB	19 Sep 2022 11:16:41	U	SMPL
PB147707BS - 1	PB147707BS	18244	3.9657	-	7471B	PPB	19 Sep 2022 11:21:44	U	SMPL
N4653-01 - 1	CAR-1948	6707	1.4271	-	7471B	PPB	19 Sep 2022 11:24:00	U	SMPL
N4655-01 - 1	SU-3-091522	2244	0.4450	-	7471B	PPB	19 Sep 2022 11:26:18	U	SMPL
N4656-01 - 1	SP-1	2275	0.4518	-	7471B	PPB	19 Sep 2022 11:28:35	U	SMPL
N4656-11 - 1	SP-2	1768	0.3403	-	7471B	PPB	19 Sep 2022 11:30:52	U	SMPL
N4656-21 - 1	SP-3	764	0.1193	-	7471B	PPB	19 Sep 2022 11:33:07	U	SMPL
CCV23 - 1	CCV23	24072	5.2481	-	7471B	PPB	19 Sep 2022 11:35:22	U	SMPL
CCB23 - 1	CCB23	-29	-0.0552	-	7471B	PPB	19 Sep 2022 11:37:38	U	SMPL
N4656-31 - 1	SP-4	665	0.0976	-	7471B	PPB	19 Sep 2022 11:39:55	U	SMPL
N4656-31DUP - 1	SP-4DUP	585	0.0799	-	7471B	PPB	19 Sep 2022 11:42:10	U	SMPL
N4656-31MS - 1	SP-4MS	18545	4.0320	-	7471B	PPB	19 Sep 2022 11:44:25	U	SMPL
N4656-31MSD - 1	SP-4MSD	17185	3.7327	-	7471B	PPB	19 Sep 2022 11:46:41	U	SMPL
N4664-01 - 1	HR-02-091522	3215	0.6587	-	7471B	PPB	19 Sep 2022 11:48:59	U	SMPL
N4665-01 - 1	COMP-1	2704	0.5462	-	7471B	PPB	19 Sep 2022 11:51:16	U	SMPL
N4665-02 - 1	COMP-2	1280	0.2329	-	7471B	PPB	19 Sep 2022 11:53:33	U	SMPL
N4656-31LX5 - 1		36	-0.0409	-	7471B	PPB	19 Sep 2022 11:55:50	U	SMPL
N4656-31A - 1		18986	4.1290	-	7471B	PPB	19 Sep 2022 11:58:07	U	SMPL
CCV24 - 1	CCV24	22049	4.8030	-	7471B	PPB	19 Sep 2022 12:00:22	U	SMPL
CCB24 - 1	CCB24	32	-0.0417	-	7471B	PPB	19 Sep 2022 12:02:39	U	SMPL
N4630-01DLX10 - 1		36810	8.0511	-	7471B	PPB	19 Sep 2022 12:04:56	U	SMPL
N4630-06DLX10 - 1		41859	9.1621	-	7471B	PPB	19 Sep 2022 12:07:11	U	SMPL
N4630-11DLX10 - 1		26229	5.7228	-	7471B	PPB	19 Sep 2022 12:12:27	U	SMPL
N4630-16DLX10 - 1		28103	6.1351	-	7471B	PPB	19 Sep 2022 12:14:43	U	SMPL
N4630-21DLX10 - 1		22164	4.8283	-	7471B	PPB	19 Sep 2022 12:16:59	U	SMPL
CCB25 - 1	CCB25	21963	4.7841	-	7471B	PPB	19 Sep 2022 12:19:18	U	SMPL
CCB25 - 1	CCB25	-156	-0.0831	-	7471B	PPB	19 Sep 2022 12:21:35	U	SMPL