

LB122069


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

45500



A= 0.0000e+000

B= 2.1880e-004

C= 2.9564e-004

Rho= 0.9999442

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.025	0.025	114	0.000	114					0/11
0.2	0.200	0.186	-0.014	850	0.0 %	850					-7
2.5	2.500	2.460	-0.040	11243	0.0 %	11243					-2
5.0	5.000	5.003	0.003	22865	0.0 %	22865					0
7.5	7.500	7.569	0.069	34591	0.0 %	34591					0
10.0	10.000	9.956	-0.044	45500	0.0 %	45500					0

LB122063 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Sep 2022 13:36:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	114	-	0.0000	7470A	PPB	19 Sep 2022 13:54:05	S	Std
0.2 - 1	50.2	850	-	0.2000	7470A	PPB	19 Sep 2022 13:56:22	S	Std
2.5 - 1	52.5	11243	-	2.5000	7470A	PPB	19 Sep 2022 13:58:39	S	Std
5.0 - 1	55	22865	-	5.0000	7470A	PPB	19 Sep 2022 14:00:56	S	Std
7.5 - 1	57.5	34591	-	7.5000	7470A	PPB	19 Sep 2022 14:05:47	S	Std
10.0 - 1	511	45500	-	10.0000	7470A	PPB	19 Sep 2022 14:08:03	S	Std
ICV35 - 1	ICV35	18015	3.9421	-	7470A	PPB	19 Sep 2022 14:10:46	U	SMPL
ICB35 - 1	ICB35	187	0.0412	-	7470A	PPB	19 Sep 2022 14:13:00	U	SMPL
CCV26 - 1	CCV26	22713	4.9700	-	7470A	PPB	19 Sep 2022 14:15:16	U	SMPL
CCB26 - 1	CCB26	169	0.0373	-	7470A	PPB	19 Sep 2022 14:17:31	U	SMPL
CRA - 1	CRA	834	0.1828	-	7470A	PPB	19 Sep 2022 14:19:46	U	SMPL
HighStd - 1	HighStd	46082	10.0833	-	7470A	PPB	19 Sep 2022 14:22:01	U	SMPL
ChkStd - 1	ChkStd	32926	7.2047	-	7470A	PPB	19 Sep 2022 14:24:17	U	SMPL
PB147745BL - 1	PBW	-60	-0.0128	-	7470A	PPB	19 Sep 2022 14:26:35	U	SMPL
PB147745BS - 1	LCSW	17454	3.8193	-	7470A	PPB	19 Sep 2022 14:28:52	U	SMPL
N4628-02 - 1	HW-54	142	0.0314	-	7470A	PPB	19 Sep 2022 14:31:08	U	SMPL
N4628-03 - 1	HW-54	45	0.0101	-	7470A	PPB	19 Sep 2022 14:33:26	U	SMPL
N4628-04 - 1	MW-7	1088	0.2384	-	7470A	PPB	19 Sep 2022 14:35:43	U	SMPL
N4628-05 - 1	MW-7	22	0.0051	-	7470A	PPB	19 Sep 2022 14:38:00	U	SMPL
N4628-06 - 1	HW-8	2105	0.4609	-	7470A	PPB	19 Sep 2022 14:40:16	U	SMPL
CCV27 - 1	CCV27	20960	4.5864	-	7470A	PPB	19 Sep 2022 14:49:51	U	SMPL
CCB27 - 1	CCB27	59	0.0132	-	7470A	PPB	19 Sep 2022 14:52:05	U	SMPL
N4628-07 - 1	HW-8	-28	-0.0058	-	7470A	PPB	19 Sep 2022 14:54:21	U	SMPL
N4648-01 - 1	N48854-1040	201	0.0443	-	7470A	PPB	19 Sep 2022 14:56:37	U	SMPL
N4663-01 - 1	HW-43	150	0.0331	-	7470A	PPB	19 Sep 2022 14:58:52	U	SMPL
N4663-01DUP - 1	HW-43DUP	97	0.0215	-	7470A	PPB	19 Sep 2022 15:01:08	U	SMPL
N4663-02 - 1	HW-43	34	0.0077	-	7470A	PPB	19 Sep 2022 15:03:24	U	SMPL
N4663-03 - 1	SVE-03	185	0.0408	-	7470A	PPB	19 Sep 2022 15:05:40	U	SMPL
N4663-04 - 1	SVE-03	49	0.0110	-	7470A	PPB	19 Sep 2022 15:07:57	U	SMPL
N4663-05 - 1	HW-9	241	0.0530	-	7470A	PPB	19 Sep 2022 15:10:14	U	SMPL
N4663-06 - 1	HW-9	39	0.0088	-	7470A	PPB	19 Sep 2022 15:12:31	U	SMPL
N4663-07 - 1	N4663-01MS	15421	3.3745	-	7470A	PPB	19 Sep 2022 15:14:47	U	SMPL
CCV28 - 1	CCV28	21165	4.6313	-	7470A	PPB	19 Sep 2022 15:17:02	U	SMPL
CCB28 - 1	CCB28	158	0.0349	-	7470A	PPB	19 Sep 2022 15:19:18	U	SMPL
N4663-08 - 1	N4663-01MSD	14763	3.2305	-	7470A	PPB	19 Sep 2022 15:21:34	U	SMPL
N4675-01 - 1	006-1	29	0.0066	-	7470A	PPB	19 Sep 2022 15:23:50	U	SMPL
N4678-01 - 1	5869-1	162	0.0357	-	7470A	PPB	19 Sep 2022 15:26:06	U	SMPL
N4680-01 - 1	5869-2	150	0.0331	-	7470A	PPB	19 Sep 2022 15:28:21	U	SMPL
N4682-01 - 1	006-2	47	0.0106	-	7470A	PPB	19 Sep 2022 15:30:37	U	SMPL
N4684-01 - 1	5486-2	40	0.0090	-	7470A	PPB	19 Sep 2022 15:32:54	U	SMPL
N4686-01 - 1	5486-1	37	0.0084	-	7470A	PPB	19 Sep 2022 15:35:11	U	SMPL
N4663-01LX5 - 1		-83	-0.0179	-	7470A	PPB	19 Sep 2022 15:37:28	U	SMPL
N4663-01A - 1		16303	3.5675	-	7470A	PPB	19 Sep 2022 15:39:45	U	SMPL
CCV29 - 1	CCV29	21954	4.8039	-	7470A	PPB	19 Sep 2022 15:46:55	U	SMPL
CCB29 - 1	CCB29	65	0.0145	-	7470A	PPB	19 Sep 2022 15:49:11	U	SMPL