



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122067

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	09/19/2022	09:50
2	N4648-02		1	13	100.00	>212.0	>212.0	09/19/2022	10:15
3	N4648-02DUP		1	14	100.00	>212.0	>212.0	09/19/2022	10:45
4	N4673-09		1	15	100.00	>212.0	>212.0	09/19/2022	11:15
5	N4690-06		1	12	100.00	>212.0	>212.0	09/19/2022	11:45
6	N4690-08		1	16	100.00	>212.0	>212.0	09/19/2022	12:15
7	N4692-03		1	13	100.00	>212.0	>212.0	09/19/2022	13:00
8	N4692-05		1	16	100.00	>212.0	>212.0	09/19/2022	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash Point-091922 WorkList ID : 163285 Department : Wet-Chemistry Date : 09-19-2022 09:09:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/20/2022	Water	N4648-02	Flash Point	Cool 4 deg C	PSEG03	N11	N48854-1040	09/15/2022	1010B
09/23/2022	Water	N4673-09	Flash Point	Cool 4 deg C	PSEG03	N11	33633	09/16/2022	1010B
09/23/2022	Water	N4690-06	Flash Point	Cool 4 deg C	PSEG03	N11	33829	09/16/2022	1010B
09/23/2022	Water	N4690-08	Flash Point	Cool 4 deg C	PSEG03	N11	33808	09/16/2022	1010B
09/23/2022	Water	N4692-03	Flash Point	Cool 4 deg C	PSEG03	N11	34202-03	09/16/2022	1010B
09/23/2022	Water	N4692-05	Flash Point	Cool 4 deg C	PSEG03	N11	34175	09/16/2022	1010B

Date/Time 9/19/2022 9:45
 Raw Sample Received by: SJ (wet chem)
 Raw Sample Relinquished by: QP(SM)

Date/Time 9/19/2022 14:00
 Raw Sample Received by: CP(SM)
 Raw Sample Relinquished by: SJ(WC)