

LB 122081


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42374



A= 0.0000e+000

B= 2.3290e-004

C= 2.1232e-002

Rho= 0.9996535

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%DI
0.0	0.000	0.009	0.009	-53	0.000	-53					5
0.2	0.200	0.232	0.032	904	0.0 %	904					16
2.5	2.500	2.360	-0.140	10041	0.0 %	10041					-6
5.0	5.000	5.074	0.074	21694	0.0 %	21694					1
7.5	7.500	7.635	0.135	32693	0.0 %	32693					2
10.0	10.000	9.890	-0.110	42374	0.0 %	42374					-1

LB122081 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Sep 2022 11:27:53

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-53	-	0.0000	7470A	PPB	20 Sep 2022 11:32:50	S	Std
0.2 - 1	51.2	904	-	0.2000	7470A	PPB	20 Sep 2022 11:38:53	S	Std
2.5 - 1	52.1	10041	-	2.5000	7470A	PPB	20 Sep 2022 11:41:10	S	Std
5.0 - 1	55	21694	-	5.0000	7470A	PPB	20 Sep 2022 11:43:27	S	Std
7.5 - 1	57.5	32693	-	7.5000	7470A	PPB	20 Sep 2022 11:45:43	S	Std
10.0 - 1	510	42374	-	10.0000	7470A	PPB	20 Sep 2022 11:50:38	S	Std
ICV36 - 1	ICV36	16289	3.8150	-	7470A	PPB	20 Sep 2022 11:53:22	U	SMPL
ICB36 - 1	ICB36	17	0.0252	-	7470A	PPB	20 Sep 2022 11:55:37	U	SMPL
CCV30 - 1	CCV30	20071	4.6958	-	7470A	PPB	20 Sep 2022 11:57:53	U	SMPL
CCB30 - 1	CCB30	34	0.0292	-	7470A	PPB	20 Sep 2022 12:00:08	U	SMPL
CRA - 1	CRA	732	0.1917	-	7470A	PPB	20 Sep 2022 12:02:25	U	SMPL
HighStd - 1	HighStd	43359	10.1196	-	7470A	PPB	20 Sep 2022 12:04:40	U	SMPL
ChkStd - 1	ChkStd	28313	6.6154	-	7470A	PPB	20 Sep 2022 12:09:43	U	SMPL
PB147738BL - 1	PBW	-26	0.0152	-	7470A	PPB	20 Sep 2022 12:11:59	U	SMPL
PB147738BS - 1	LCSW	17349	4.0619	-	7470A	PPB	20 Sep 2022 12:14:16	U	SMPL
N4653-02 - 1	CAR-1948	42	0.0310	-	7470A	PPB	20 Sep 2022 12:16:33	U	SMPL
N4656-02 - 1	SP-1	199	0.0676	-	7470A	PPB	20 Sep 2022 12:18:50	U	SMPL
N4656-12 - 1	SP-2	225	0.0736	-	7470A	PPB	20 Sep 2022 12:21:07	U	SMPL
N4656-22 - 1	SP-3	221	0.0727	-	7470A	PPB	20 Sep 2022 12:23:24	U	SMPL
N4656-32 - 1	SP-4	262	0.0823	-	7470A	PPB	20 Sep 2022 12:25:40	U	SMPL
CCV31 - 1	CCV31	21584	5.0482	-	7470A	PPB	20 Sep 2022 12:39:34	U	SMPL
CCB31 - 1	CCB31	55	0.0340	-	7470A	PPB	20 Sep 2022 12:41:49	U	SMPL
N4695-01 - 1	INCINERATOR-ASH	269	0.0839	-	7470A	PPB	20 Sep 2022 12:44:05	U	SMPL
N4697-02 - 1	SB-22-COMP	245	0.0783	-	7470A	PPB	20 Sep 2022 12:46:20	U	SMPL
N4697-04 - 1	SB-23-COMP	170	0.0608	-	7470A	PPB	20 Sep 2022 12:48:36	U	SMPL
N4697-06 - 1	SB-24-COMP	146	0.0552	-	7470A	PPB	20 Sep 2022 12:50:52	U	SMPL
N4697-08 - 1	SB-31-COMP	205	0.0690	-	7470A	PPB	20 Sep 2022 12:53:07	U	SMPL
N4697-10 - 1	SB-21-COMP	225	0.0736	-	7470A	PPB	20 Sep 2022 12:55:24	U	SMPL
N4697-12 - 1	SB-20-COMP	257	0.0811	-	7470A	PPB	20 Sep 2022 12:57:40	U	SMPL
N4697-14 - 1	SB-19-COMP	261	0.0820	-	7470A	PPB	20 Sep 2022 12:59:57	U	SMPL
N4697-16 - 1	SB-18-COMP	297	0.0904	-	7470A	PPB	20 Sep 2022 13:02:14	U	SMPL
N4697-18 - 1	SB-25-COMP	254	0.0804	-	7470A	PPB	20 Sep 2022 13:04:29	U	SMPL
CCV32 - 1	CCV32	21035	4.9203	-	7470A	PPB	20 Sep 2022 13:06:45	U	SMPL
CCB32 - 1	CCB32	100	0.0445	-	7470A	PPB	20 Sep 2022 13:09:01	U	SMPL
N4697-20 - 1	SB-17-COMP	194	0.0664	-	7470A	PPB	20 Sep 2022 13:11:17	U	SMPL
N4697-20DUP - 1	SB-17-COMP DUP	216	0.0715	-	7470A	PPB	20 Sep 2022 13:13:32	U	SMPL
PB147756BL - 1	PBW	161	0.0587	-	7470A	PPB	20 Sep 2022 13:20:21	U	SMPL
PB147756BS - 1	LCSW	15946	3.7351	-	7470A	PPB	20 Sep 2022 13:22:39	U	SMPL
N4648-02 - 1	N48854-1040	115	0.0480	-	7470A	PPB	20 Sep 2022 13:24:55	U	SMPL
N4648-02DUP - 1	N48854-1040DUP	177	0.0625	-	7470A	PPB	20 Sep 2022 13:27:13	U	SMPL
N4648-02MS - 1	N48854-1040MS	16282	3.8133	-	7470A	PPB	20 Sep 2022 13:29:30	U	SMPL
N4648-02MSD - 1	N48854-1040MSD	16204	3.7952	-	7470A	PPB	20 Sep 2022 13:34:05	U	SMPL
N4697-20MS - 1	SB-17-COMPMS	16808	3.9359	-	7470A	PPB	20 Sep 2022 13:36:37	U	SMPL
N4697-20MSD - 1	SB-17-COMPMSD	16508	3.8660	-	7470A	PPB	20 Sep 2022 13:38:53	U	SMPL
CCV33 - 1	CCV33	21546	5.0393	-	7470A	PPB	20 Sep 2022 13:41:30	U	SMPL
CCB33 - 1	CCB33	64	0.0361	-	7470A	PPB	20 Sep 2022 13:43:46	U	SMPL
N4700-03 - 1	TOTE-5563	241	0.0774	-	7470A	PPB	20 Sep 2022 13:46:03	U	SMPL
PB147694TB - 1	PB147694TB	205	0.0690	-	7470A	PPB	20 Sep 2022 13:48:18	U	SMPL
PB147696TB - 1	PB147696TB	207	0.0694	-	7470A	PPB	20 Sep 2022 13:50:34	U	SMPL
N4697-20LX5 - 1		21	0.0261	-	7470A	PPB	20 Sep 2022 13:52:50	U	SMPL
N4697-20A - 1		17173	4.0209	-	7470A	PPB	20 Sep 2022 13:55:05	U	SMPL
N4648-02LX5 - 1		-69	0.0052	-	7470A	PPB	20 Sep 2022 13:57:22	U	SMPL
N4648-02A - 1		17705	4.1448	-	7470A	PPB	20 Sep 2022 13:59:39	U	SMPL
CCV34 - 1	CCV34	22168	5.1842	-	7470A	PPB	20 Sep 2022 14:01:56	U	SMPL
CCB34 - 1	CCB34	-82	0.0021	-	7470A	PPB	20 Sep 2022 14:04:14	U	SMPL