



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122085

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	81.2	09/20/2022	14:18
2	N4700-03		1	15	100.00	>212.0	>212.0	09/20/2022	15:10
3	N4700-03DUP		1	13	100.00	>212.0	>212.0	09/20/2022	15:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flash point-0920

WorkList ID : 163336

Department : Wet-Chemistry

Date : 09-20-2022 14:13:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/21/2022	Water	N4700-03	Flash Point	Cool 4 deg C	PSEG03	N11	TOTE-5563	09/16/2022	1010B

Date/Time 9/20/22 14:15
Raw Sample Received by: SI (wc)
Raw Sample Relinquished by: CP (SM)

Date/Time 9/20/22 16:00
Raw Sample Received by: CP (SM)
Raw Sample Relinquished by: SI (wc)